

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/23/17
 Analyst : AK
 Data File : lb85444.csv

Approved By : 
 Approved Date : 1/24/17
 Worksheet # : 1

m9012 A-B Total Amenable & Reactive Cyanide -16

Lab Sample ID	Client ID		Raw Amt PPB	Dil	Matrix	A. Date	Prep Method	Analysis Method			Line 1	
Parameter						%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Cyanide												
0.0PPBCN	0.0PPBCN				W	1/20/17						
Cyanide		PASS	0.476		0.000						mg/L	
5.0PPBCN	5.0PPBCN				W	1/20/17						
Cyanide		PASS	5.649		0.006						mg/L	
10PPBCN	10PPBCN				W	1/20/17						
Cyanide		PASS	9.898		0.010						mg/L	
50PPBCN	50PPBCN				W	1/20/17						
Cyanide		PASS	52.492		0.052						mg/L	
100PPBCN	100PPBCN				W	1/20/17						
Cyanide		PASS	97.185		0.097						mg/L	
250PPBCN	250PPBCN				W	1/20/17						
Cyanide		PASS	247.984		0.248						mg/L	
500PPBCN	500PPBCN				W	1/20/17						
Cyanide		PASS	501.318		0.501						mg/L	
ICV1	ICV1				W	1/20/17						
Cyanide		PASS	92.173		0.09	90.0	90	110			mg/L	
ICB1	ICB1				W	1/20/17						
Cyanide		PASS	1.136		0.001			+/-0.0030			mg/L	
CCV1	CCV1				W	1/20/17						
Cyanide		PASS	235.760		0.24	96.0	90	110			mg/L	
CCB1	CCB1				W	1/20/17						
Cyanide		PASS	1.468		0.001			+/-0.0030			mg/L	
PB96189BLS	PB96189BLS				S	1/20/17						
Cyanide		PASS	1.183		0.067			+/-0.0330			mg/Kg	
PB96189BSS	PB96189BSS				S	1/20/17						
Cyanide		PASS	94.394		5.33	107.0	85.00	115.00			mg/Kg	
I1304-01	ROLL-OFF-15-41-COMP		1		S	1/20/17						
Cyanide		PASS	1.461		0.080						mg/Kg	
I1304-01D	ROLL-OFF-15-41-COMPD		1		S	1/20/17						
Cyanide		PASS	1.487		0.082				2.5	20	mg/Kg	
I1304-01S	ROLL-OFF-15-41-COMPS		1		S	1/20/17						
Cyanide		FAIL	31.079		1.730	74.0	75	125			mg/Kg	
I1304-01SD	ROLL-OFF-15-41-COMPSD		1		S	1/20/17						
Cyanide		FAIL	30.695		1.690	73.0	75	125	2.3	20	mg/Kg	
I1304-04	1634-NOTTINGHAM-COMP		1		S	1/20/17						
Cyanide		PASS	1.364		0.075						mg/Kg	
I1304-07	ROLL-OFF-15939-COMP		1		S	1/20/17						
Cyanide		PASS	1.732		0.096						mg/Kg	
I1308-01	MH-2-0-12FT-COMP		1		S	1/20/17						
Cyanide		PASS	1.163		0.061						mg/Kg	

Chemtech Consulting Group


Analytical Review Report


Date Printed : 1/23/17

Analyst : AK

Data File : lb85444.csv

Approved By : 

Approved Date : 

Worksheet # : 

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis		Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method		Line 2
Cyanide										
CCV2	CCV2				W	1/20/17				
Cyanide		PASS	250.518		0.25	100.0	90	110		mg/L
CCB2	CCB2				W	1/20/17				
Cyanide		PASS	1.440		0.001			+/-0.0030		mg/L
CCV3	CCV3				W	1/20/17				
Cyanide		PASS	238.530		0.24	96.0	90	110		mg/L
CCB3	CCB3				W	1/20/17				
Cyanide		PASS	1.493		0.001			+/-0.0030		mg/L
LOWPB96189	LOWPB96189				W	1/20/17				
Cyanide		PASS	10.781		0.011					mg/L
HIGHPB96189	HIGHPB96189				W	1/20/17				
Cyanide		PASS	510.006		0.510					mg/L
CCV4	CCV4				W	1/20/17				
Cyanide		PASS	245.167		0.25	100.0	90	110		mg/L
CCB4	CCB4				W	1/20/17				
Cyanide		PASS	1.989		0.002			+/-0.0030		mg/L