

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/24/17

Analyst : AK

Data File : lb85469.csv

Approved By : 

Approved Date : 1/27/17

Worksheet # :

msm4500 - NO2B - Nitrite - 8

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Nitrite											
0.0061PPM	0.0061PPM			W	1/24/17						
Nitrite	PASS	0.004	0.004							mg/L	
0.0122PPM	0.0122PPM			W	1/24/17						
Nitrite	PASS	0.010	0.010							mg/L	
0.0243PPM	0.0243PPM			W	1/24/17						
Nitrite	PASS	0.024	0.024							mg/L	
0.0608PPM	0.0608PPM			W	1/24/17						
Nitrite	PASS	0.061	0.061							mg/L	
0.1216PPM	0.1216PPM			W	1/24/17						
Nitrite	PASS	0.125	0.125							mg/L	
0.304PPM	0.304PPM			W	1/24/17						
Nitrite	PASS	0.307	0.307							mg/L	
0.608PPM	0.608PPM			W	1/24/17						
Nitrite	PASS	0.606	0.606							mg/L	
ICV1	ICV1			W	1/24/17						
Nitrite	PASS	0.314	0.31		103.0	90	110			mg/L	
ICB1	ICB1			W	1/24/17						
Nitrite	PASS	-0.002	-0.002				+/-0.0060			mg/L	
CCV1	CCV1			W	1/24/17						
Nitrite	PASS	0.307	0.31		103.0	90	110			mg/L	
CCB1	CCB1			W	1/24/17						
Nitrite	PASS	-0.002	-0.002				+/-0.0060			mg/L	
LB85469BLW	LB85469BLW			W	1/24/17						
Nitrite	PASS	-0.002	-0.002				+/-0.0060			mg/L	
LB85469BSW	LB85469BSW			W	1/24/17						
Nitrite	PASS	0.315	0.315		104.0	80	120			mg/L	
I1272-33	0041		1	W	1/24/17						
Nitrite	OR	0.733	0.733							mg/L	
I1272-33DL	0041DL		10	W	1/24/17						
Nitrite	PASS	0.307	3.07							mg/L	
CCV2	CCV2			W	1/24/17						
Nitrite	PASS	0.317	0.32		107.0	90	110			mg/L	
CCB2	CCB2			W	1/24/17						
Nitrite	PASS	-0.002	-0.002				+/-0.0060			mg/L	