

LB85519

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

1/25/2017 15:25

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.622	0.0	0.099	
ICB1	1.270	0.0	0.006	
CCV1	241.891	0.0	0.247	
CCB1	0.908	0.0	0.005	
PB96318BLW	0.656	0.0	0.005	
PB96318BSW	96.549	0.0	0.101	
MIDPB96318	243.365	0.0	0.248	
I1272-30	296.561	0.0	0.301	
CCV2	244.404	0.0	0.249	
CCB2	1.145	0.0	0.006	

N	10
Mean	122.137
SD	122.2583
CV%	100.10

LB 85519

Aquakem v. 7.2AQ1

AK

Results from time period:

Wed Jan 25 15:06:07 2017

Wed Jan 25 15:10:00 2017

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	0.8271	µg/l	1/25/2017 13:23:26	
5.0PPBCN	A	Total CN	P	5.4307	µg/l	1/25/2017 13:23:27	
10PPBCN	A	Total CN	P	9.9163	µg/l	1/25/2017 13:23:28	
50PPBCN	A	Total CN	P	50.861	µg/l	1/25/2017 13:23:29	
100PPBCN	A	Total CN	P	99.1021	µg/l	1/25/2017 13:23:30	
250PPBCN	A	Total CN	P	247.5437	µg/l	1/25/2017 13:23:31	
500PPBCN	A	Total CN	P	501.319	µg/l	1/25/2017 13:23:32	
ICV1	S	Total CN	P	94.6224	µg/l	1/25/2017 15:06:07	
ICB1	S	Total CN	P	1.2703	µg/l	1/25/2017 15:06:08	
CCV1	S	Total CN	P	241.8908	µg/l	1/25/2017 15:06:09	
CCB1	S	Total CN	P	0.9081	µg/l	1/25/2017 15:06:10	
PB96318B1	S	Total CN	P	0.6561	µg/l	1/25/2017 15:06:11	
PB96318B5	S	Total CN	P	96.5494	µg/l	1/25/2017 15:06:12	
MIDPB963	S	Total CN	P	243.3653	µg/l	1/25/2017 15:06:13	
I1272-30	S	Total CN	P	296.5614	µg/l	1/25/2017 15:06:15	
CCV2	S	Total CN	P	244.4035	µg/l	1/25/2017 15:06:17	
CCB2	S	Total CN	P	1.1449	µg/l	1/25/2017 15:09:58	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : ELK

1/25/2017 13:32


Test Total CN

Accepted 1/25/2017 13:32

Factor 998.9
Bias 0.005

Coeff. of det. 0.999950

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.005	0.8271	0.0000	
2	5.0PPBCN	0.010	5.4307	5.0000	
3	10PPBCN	0.014	9.9163	10.0000	
4	50PPBCN	0.055	50.8610	50.0000	
5	100PPBCN	0.104	99.1021	100.0000	
6	250PPBCN	0.252	247.5437	250.0000	
7	500PPBCN	0.506	501.3190	500.0000	