

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/27/17

Analyst : *JS*

Data File : lb85519.csv

Approved By : *m*

Approved Date : *1/27/17*

Worksheet # :

ms m 4500 - CN, CIE - Cyanide - 8

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method				Line 2
							LCL	UCL	RPD	Max RPD	Units	
Cyanide												
0.0PPBCN	0.0PPBCN				W	1/25/17						
Cyanide		PASS	0.827		0.001						mg/L	
5.0PPBCN	5.0PPBCN				W	1/25/17						
Cyanide		PASS	5.431		0.005						mg/L	
10PPBCN	10PPBCN				W	1/25/17						
Cyanide		PASS	9.916		0.010						mg/L	
50PPBCN	50PPBCN				W	1/25/17						
Cyanide		PASS	50.861		0.051						mg/L	
100PPBCN	100PPBCN				W	1/25/17						
Cyanide		PASS	99.102		0.099						mg/L	
250PPBCN	250PPBCN				W	1/25/17						
Cyanide		PASS	247.544		0.248						mg/L	
500PPBCN	500PPBCN				W	1/25/17						
Cyanide		PASS	501.319		0.501						mg/L	
ICV1	ICV1				W	1/25/17						
Cyanide		PASS	94.622		0.09	90.0	85	115			mg/L	
ICB1	ICB1				W	1/25/17						
Cyanide		PASS	1.270		0.001			+/-0.0010			mg/L	
CCV1	CCV1				W	1/25/17						
Cyanide		PASS	241.891		0.24	96.0	90	110			mg/L	
CCB1	CCB1				W	1/25/17						
Cyanide		PASS	0.908		0.001			+/-0.0010			mg/L	
PB96318BLW	PB96318BLW				W	1/25/17						
Cyanide		PASS	0.656		0.001			+/-0.0010			mg/L	
PB96318BSW	PB96318BSW				W	1/25/17						
Cyanide		PASS	96.549		0.097	97.0	85	115			mg/L	
MIDPB96318	MIDPB96318				W	1/25/17						
Cyanide		PASS	243.365		0.243						mg/L	
I1272-30	0012			1	W	1/25/17						
Cyanide		PASS	296.561		0.297						mg/L	
CCV2	CCV2				W	1/25/17						
Cyanide		PASS	244.404		0.24	96.0	90	110			mg/L	
CCB2	CCB2				W	1/25/17						
Cyanide		PASS	1.145		0.001			+/-0.0010			mg/L	