

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/27/17

Analyst : AK

Data File : lb85554.csv

Approved By : 

Approved Date : 1/27/17

Worksheet # :


m9012 A-B Total Amenable & Reactive cyanide -16

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Cyanide												
0.0PPBCN	0.0PPBCN				W	1/27/17						
Cyanide		PASS	-0.094		-0.000						mg/L	
5.0PPBCN	5.0PPBCN				W	1/27/17						
Cyanide		PASS	4.531		0.005						mg/L	
10PPBCN	10PPBCN				W	1/27/17						
Cyanide		PASS	9.157		0.009						mg/L	
50PPBCN	50PPBCN				W	1/27/17						
Cyanide		PASS	50.840		0.051						mg/L	
100PPBCN	100PPBCN				W	1/27/17						
Cyanide		PASS	99.716		0.100						mg/L	
250PPBCN	250PPBCN				W	1/27/17						
Cyanide		PASS	251.713		0.252						mg/L	
500PPBCN	500PPBCN				W	1/27/17						
Cyanide		PASS	499.138		0.499						mg/L	
ICV1	ICV1				W	1/27/17						
Cyanide		PASS	93.134		0.09	90.0	90	110			mg/L	
ICB1	ICB1				W	1/27/17						
Cyanide		PASS	2.186		0.002			+/-0.0030			mg/L	
CCV1	CCV1				W	1/27/17						
Cyanide		PASS	238.286		0.24	96.0	90	110			mg/L	
CCB1	CCB1				W	1/27/17						
Cyanide		PASS	0.384		0.000			+/-0.0030			mg/L	
PB96321BLS	PB96321BLS				S	1/27/17						
Cyanide		PASS	0.356		0.018			+/-0.0330			mg/Kg	
PB96321BSS	PB96321BSS				S	1/27/17						
Cyanide		PASS	97.277		4.86	97.0	85.00	115.00			mg/Kg	
LOWPB96321	LOWPB96321				W	1/27/17						
Cyanide		PASS	9.330		0.009						mg/L	
HIGHPB96321	HIGHPB96321				W	1/27/17						
Cyanide		PASS	490.286		0.490						mg/L	
I1312-01	CAS-SB-01-0-2			I	S	1/27/17						
Cyanide		PASS	0.510		0.028						mg/Kg	
I1312-01D	CAS-SB-01-0-2D			I	S	1/27/17						
Cyanide		PASS	0.242		0.013				0	20	mg/Kg	
I1312-02S	CAS-SB-01-0-2S			I	S	1/27/17						
Cyanide		PASS	33.891		1.870	83.0	75	125			mg/Kg	
I1312-03SD	CAS-SB-01-0-2SD			I	S	1/27/17						
Cyanide		PASS	33.709		1.840	83.0	75	125	1.6	20	mg/Kg	
I1312-04	CAS-SB-101-0-2			I	S	1/27/17						
Cyanide		PASS	0.649		0.035						mg/Kg	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/27/17
 Analyst : AK
 Data File : lb85554.csv

Approved By : 
 Approved Date :
 Worksheet # :

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis					Line 1
Parameter		PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units		Line 2
Cyanide												
I1312-05	CAS-SB-02-0-2		1	S	1/27/17							
Cyanide	PASS	0.607		0.032						mg/Kg		
CCV2	CCV2			W	1/27/17							
Cyanide	PASS	246.795		0.25	100.0	90	110			mg/L		
CCB2	CCB2			W	1/27/17							
Cyanide	PASS	-0.003		0.000			+/-0.0030			mg/L		
I1312-06	CAS-SB-03-0-2		1	S	1/27/17							
Cyanide	PASS	-0.032		-0.002						mg/Kg		
I1312-07	CAS-BALLARD01-0-2		1	S	1/27/17							
Cyanide	PASS	-0.177		-0.009						mg/Kg		
I1312-08	CAS-SB-02-5-7		1	S	1/27/17							
Cyanide	PASS	1.864		0.110						mg/Kg		
I1312-11	CAS-SB-03-5-7		1	S	1/27/17							
Cyanide	PASS	0.886		0.050						mg/Kg		
CCV3	CCV3			W	1/27/17							
Cyanide	PASS	252.189		0.25	100.0	90	110			mg/L		
CCB3	CCB3			W	1/27/17							
Cyanide	PASS	0.474		0.000			+/-0.0030			mg/L		