

LB85577



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908-789-8900 Fax: 908-789-892

Analysis Method: SM5310B TOC
Parameter: TOC
Run Number: LB85577
Instrument: Appolo-9000

msm5310 B - TOC - 10



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908-789-8900 Fax: 908-789-892

Analytical Summary Report

Analysis Method: SM5310B TOC
Parameter: TOC
Run Number: LB85577
Instrument: Appolo-9000

Seq	Lab ID	Client ID	Sample Type	Inst Conc. ppm/ppb *	Dil	Analytical Date
1	0.0PPM	CAL	CAL	7751.000	1	1/30/17
2	0.5PPM	CAL	CAL	21494.000	1	1/30/17
3	1.0PPM	CAL	CAL	34457.000	1	1/30/17
4	5.0PPM	CAL	CAL	142741.000	1	1/30/17
5	10.0PPM	CAL	CAL	292332.000	1	1/30/17
6	20.0PPM	CAL	CAL	545609.000	1	1/30/17
7	2.0PPM	CAL	CAL	63973.000	1	1/30/17
8	ICV	ICV	ICV	9.671	1	1/30/17
9	ICB	ICB	ICB	0.000	1	1/30/17
10	CCV	CCV	CCV	10.324	1	1/30/17
11	CCB	CCB	CCB	0.343	1	1/30/17
12	LB85577BLW	LB85577BLW	MB	0.210	1	1/30/17
13	LB85577BSW	LB85577BSW	LCS	9.482	1	1/30/17
14	I1272-18	0001	SAM	25.182	1	1/30/17
15	I1272-18	0001	SAM	11.597	2	1/30/17
16	CCV	CCV	CCV	9.774	1	1/30/17
17	CCB	CCB	CCB	0.000	1	1/30/17

* Cal results are in raw ammount.

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/2/17
 Analyst : AK
 Data File : LB85577.MDB

Approved By : 
 Approved Date : 
 Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1	
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
TOC												
0.0PPM	CAL				W	1/30/17						
TOC		PASS	7751.000		7750						mg/L	
0.5PPM	CAL				W	1/30/17						
TOC		PASS	21494.000		21500						mg/L	
1.0PPM	CAL				W	1/30/17						
TOC		PASS	34457.000		34500						mg/L	
5.0PPM	CAL				W	1/30/17						
TOC		PASS	142741.000		143000						mg/L	
10.0PPM	CAL				W	1/30/17						
TOC		PASS	292332.000		292000						mg/L	
20.0PPM	CAL				W	1/30/17						
TOC		PASS	545609.000		546000						mg/L	
2.0PPM	CAL				W	1/30/17						
TOC		PASS	63973.000		64000						mg/L	
ICV1	ICV1				W	1/30/17						
TOC		PASS	9.671		9.67	97.0	90	110			mg/L	
ICB1	ICB1				W	1/30/17						
TOC		PASS	0.000		0.000			+/-0.0800			mg/L	
CCV1	CCV1				W	1/30/17						
TOC		PASS	10.324		10.30	103.0	90	110			mg/L	
CCB1	CCB1				W	1/30/17						
TOC		PASS	0.343		0.343			+/-0.0800			mg/L	
LB85577BLW	LB85577BLW				W	1/30/17						
TOC		PASS	0.210		0.210			+/-0.0800			mg/L	
LB85577BSW	LB85577BSW				W	1/30/17						
TOC		PASS	9.482		9.48	94.8	90	110			mg/L	
I1272-18	0001			1	W	1/30/17						
TOC		OR	25.182		25.2						mg/L	
I1272-18DL	0001DL			2	W	1/30/17						
TOC		PASS	11.597		23.2						mg/L	
CCV2	CCV2				W	1/30/17						
TOC		PASS	9.774		9.77	98.0	90	110			mg/L	
CCB2	CCB2				W	1/30/17						
TOC		PASS	0.000		0.000			+/-0.0800			mg/L	

Cal results
are in raw
amount.