

LB85578



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908-789-8900 Fax: 908-789-892

Analysis Method: 9060 TOC
Parameter: TOC
Run Number: LB85578
Instrument: Appolo-9000

M9060-A-TOC-18

Analytical Summary Report

Analysis Method: 9060 TOC
Parameter: TOC
Run Number: LB85578
Instrument: Appolo-9000

Seq	Lab ID	Client ID	Sample Type	Inst Conc. ppm/ppb*	Dil	Analytical Date
1	0.0PPM	0.0PPM	CAL	7751.000	1	1/30/17
2	0.5PPM	0.5PPM	CAL	21494.000	1	1/30/17
3	1.0PPM	1.0PPM	CAL	34457.000	1	1/30/17
4	5.0PPM	5.0PPM	CAL	142741.000	1	1/30/17
5	10.0PPM	10.0PPM	CAL	292332.000	1	1/30/17
6	20.0PPM	20.0PPM	CAL	545609.000	1	1/30/17
7	2.0PPM	2.0PPM	CAL	63973.000	1	1/30/17
8	ICV	ICV	ICV	9.671	1	1/30/17
9	ICB	ICB	ICB	0.000	1	1/30/17
10	CCV	CCV	CCV	10.713	1	1/31/17
11	CCB	CCB	CCB	0.000	1	1/31/17
12	LB85578BLW	LB85578BLW	MB	0.159	1	1/31/17
13	LB85578BSW	LB85578BSW	LCS	10.477	1	1/31/17
14	I1272-19	0001	SAM	22.551	1	1/31/17
15	I1272-19	0001	SAM	11.966	2	1/31/17
16	I1535-01	S49-0414	SAM	1.060	1	1/31/17
17	I1535-02	S49-0413	SAM	4.667	1	1/31/17
18	I1535-03	S49-EB-2017-1	SAM	0.707	1	1/31/17
19	I1535-01S	S49-0414S	MS	12.124	1	1/31/17
20	I1535-01SD	S49-0414SD	MSD	12.137	1	1/31/17
21	CCV	CCV	CCV	10.868	1	1/31/17
22	CCB	CCB	CCB	0.000	1	1/31/17

* Cal results are in raw amount.

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/2/17
 Analyst : AP
 Data File : LB85578.MDB

Approved By :
 Approved Date : 2/2/17
 Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units
TOC											
0.0PPM	0.0PPM				W	1/30/17					
TOC		PASS	7751.000		7750						mg/L
0.5PPM	0.5PPM				W	1/30/17					
TOC		PASS	21494.000		21500						mg/L
1.0PPM	1.0PPM				W	1/30/17					
TOC		PASS	34457.000		34500						mg/L
5.0PPM	5.0PPM				W	1/30/17					
TOC		PASS	142741.000		143000						mg/L
10.0PPM	10.0PPM				W	1/30/17					
TOC		PASS	292332.000		292000						mg/L
20.0PPM	20.0PPM				W	1/30/17					
TOC		PASS	545609.000		546000						mg/L
2.0PPM	2.0PPM				W	1/30/17					
TOC		PASS	63973.000		64000						mg/L
ICV1	ICV1				W	1/30/17					
TOC		PASS	9.671		9.67	97.0	90	110			mg/L
ICB1	ICB1				W	1/30/17					
TOC		PASS	0.000		0.000			+/-0.0970			mg/L
CCV1	CCV1				W	1/31/17					
TOC		PASS	10.713		10.70	107.0	90	110			mg/L
CCB1	CCB1				W	1/31/17					
TOC		PASS	0.000		0.000			+/-0.0970			mg/L
LB85578BLW	LB85578BLW				W	1/31/17					
TOC		PASS	0.159		0.159			+/-0.0970			mg/L
LB85578BSW	LB85578BSW				W	1/31/17					
TOC		PASS	10.477		10.50	105.0	90	110			mg/L
I1272-19	0001			1	W	1/31/17					
TOC		OR	22.551		22.6						mg/L
I1272-19DL	0001DL			2	W	1/31/17					
TOC		PASS	11.966		23.9						mg/L
I1535-01	S49-0414			1	W	1/31/17					
TOC		PASS	1.060		1.06						mg/L
I1535-02	S49-0413			1	W	1/31/17					
TOC		PASS	4.667		4.67						mg/L
I1535-03	S49-EB-2017-1			1	W	1/31/17					
TOC		PASS	0.707		0.707						mg/L
I1535-01S	S49-0414S			1	W	1/31/17					
TOC		PASS	12.124		12.10	110.0	75	125			mg/L
I1535-01SD	S49-0414SD			1	W	1/31/17					
TOC		PASS	12.137		12.10	110.0	75	125	0	20	mg/L

Cal results are in raw amount

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/2/17

Analyst : AP

Data File : LB85578.MDB

Approved By : 

Approved Date : 

Worksheet # : 

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			
Parameter			PPB		Final Conc	%Rec	Method	Method			Line 1
TOC											Line 2
CCV2	CCV2				W	1/31/17					
TOC		PASS	10.868		10.90	109.0	90	110			mg/L
CCB2	CCB2				W	1/31/17					
TOC		PASS	0.000		0.000			+/-0.0970			mg/L