

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/31/17

Analyst : AK

Data File : lb85584.csv

Approved By : 

Approved Date : 1/31/17

Worksheet # :

m905 A-13 Total Amenable & Reactive Cyanide-16

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Cyanide												
100PPBCN	100PPBCN				W	1/30/17						
Cyanide		PASS	100.685		0.101						mg/L	
250PPBCN	250PPBCN				W	1/30/17						
Cyanide		PASS	252.342		0.252						mg/L	
500PPBCN	500PPBCN				W	1/30/17						
Cyanide		PASS	498.744		0.499						mg/L	
0.0PPBCN	0.0PPBCN				W	1/30/17						
Cyanide		PASS	-0.240		-0.000						mg/L	
5.0PPBCN	5.0PPBCN				W	1/30/17						
Cyanide		PASS	4.306		0.004						mg/L	
10PPBCN	10PPBCN				W	1/30/17						
Cyanide		PASS	9.520		0.010						mg/L	
50PPBCN	50PPBCN				W	1/30/17						
Cyanide		PASS	49.642		0.050						mg/L	
ICV1	ICV1				W	1/30/17						
Cyanide		PASS	94.651		0.09	90.0	90	110			mg/L	
ICB1	ICB1				W	1/30/17						
Cyanide		PASS	-0.344		0.000			+/-0.0030			mg/L	
CCV1	CCV1				W	1/30/17						
Cyanide		PASS	243.352		0.24	96.0	90	110			mg/L	
CCB1	CCB1				W	1/30/17						
Cyanide		PASS	-0.152		0.000			+/-0.0030			mg/L	
PB96417BLS	PB96417BLS				S	1/30/17						
Cyanide		PASS	-0.560		-0.028			+/-0.0330			mg/Kg	
PB96417BSS	PB96417BSS				S	1/30/17						
Cyanide		PASS	99.091		4.95	99.0	85.00	115.00			mg/Kg	
LOWPB96417	LOWPB96417				W	1/30/17						
Cyanide		PASS	9.529		0.010						mg/L	
HIGHPB96417	HIGHPB96417				W	1/30/17						
Cyanide		PASS	501.105		0.501						mg/L	
I1448-06	CHA-SB-02-2-4			I	S	1/30/17						
Cyanide		PASS	0.132		0.007						mg/Kg	
I1448-06D	CHA-SB-02-2-4D			I	S	1/30/17						
Cyanide		PASS	-0.436		-0.025				0	20	mg/Kg	
I1448-06S	CHA-SB-02-2-4S			I	S	1/30/17						
Cyanide		PASS	38.402		2.130	96.0	75	125			mg/Kg	
I1448-06SD	CHA-SB-02-2-4SD			I	S	1/30/17						
Cyanide		PASS	36.941		2.050	92.0	75	125	3.8	20	mg/Kg	
I1448-07	CHA-SB-03-2-4			I	S	1/30/17						
Cyanide		PASS	-0.344		-0.019						mg/Kg	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/31/17
Analyst : AK
Data File : lb85584.csv

Approved By : 
Approved Date : 1/31/17
Worksheet # :

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep Method	Analysis Method	Line 1
Parameter		PPB	Final Conc	%Rec		LCL	UCL	Line 2
Cyanide								
I1448-08	CHA-BL-01-0-2		1	S	1/30/17			
Cyanide	PASS	-0.777	-0.042					mg/Kg
CCV2	CCV2		W	1/30/17				
Cyanide	PASS	246.054	0.25	100.0	90	110		mg/L
CCB2	CCB2		W	1/30/17				
Cyanide	PASS	-0.301	0.000			+/-0.0030		mg/L
I1529-01	ELT-SB-02-13-15		1	S	1/30/17			
Cyanide	PASS	-0.525	-0.028					mg/Kg
I1529-03	ELT-BL-0-2		1	S	1/30/17			
Cyanide	PASS	1.136	0.066					mg/Kg
I1552-01	FK-EF-01-003-ABCD		1	S	1/30/17			
Cyanide	PASS	-0.793	-0.040					mg/Kg
I1553-01	FK-BPM-01-A1B-A2B-A3B-A4		1	S	1/30/17			
Cyanide	PASS	-0.546	-0.027					mg/Kg
I1554-01	FK-BPM-01-AA1-AA2-AA3-A		1	S	1/30/17			
Cyanide	PASS	0.067	0.003					mg/Kg
I1555-01	FK-BPM-01-AA5-AA6-AA7-A		1	S	1/30/17			
Cyanide	PASS	-0.497	-0.026					mg/Kg
PB96418BLW	PB96418BLW		W	1/30/17				
Cyanide	PASS	-0.834	-0.001			+/-0.0030		mg/L
PB96418BSW	PB96418BSW		W	1/30/17				
Cyanide	PASS	96.533	0.10	100.0	85	115		mg/L
I1448-05	CHA-FB012417		1	W	1/30/17			
Cyanide	PASS	-0.930	-0.001					mg/L
CCV3	CCV3		W	1/30/17				
Cyanide	PASS	264.355	0.26	104.0	90	110		mg/L
CCB3	CCB3		W	1/30/17				
Cyanide	PASS	-0.327	0.000			+/-0.0030		mg/L
I1448-05D	CHA-FB012417D		1	W	1/30/17			
Cyanide	PASS	-0.489	0.000				0	20 mg/L
I1448-05S	CHA-FB012417S		1	W	1/30/17			
Cyanide	PASS	41.325	0.041	103.0	75	125		mg/L
I1448-05SD	CHA-FB012417SD		1	W	1/30/17			
Cyanide	PASS	41.964	0.042	105.0	75	125	2.4	20 mg/L
I1529-06	ELT-FB12617		1	W	1/30/17			
Cyanide	PASS	-0.340	-0.000					mg/L
I1529-02	ELT-SB-01-58-60		1	S	1/30/17			
Cyanide	PASS	-0.662	-0.039					mg/Kg
CCV4	CCV4		W	1/30/17				
Cyanide	PASS	266.262	0.27	108.0	90	110		mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/31/17
Analyst : AK
Data File : lb85584.csv

Approved By : 
Approved Date : 1/31/17
Worksheet # : 1/31/17

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis					Line 1
Parameter		PPB		Final Conc	%Rec	Method	Method					Line 2
Cyanide												
CCB4	CCB4			W	1/30/17							
Cyanide		PASS	-0.445	0.000			+/-0.0030				mg/L	