

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/31/17

Analyst : *AK*

Data File : lb85608.csv

Approved By : *[Signature]*

Approved Date : *1/31/17*

Worksheet # :

msm4500 - 12 org Bar C - TKN - 7

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
TKN												
0.5PPM	0.5PPM				W	1/31/17						
TKN		PASS	0.564		0.564						mg/L	
1.0PPM	1.0PPM				W	1/31/17						
TKN		PASS	1.018		1.02						mg/L	
2.5PPM	2.5PPM				W	1/31/17						
TKN		PASS	2.460		2.46						mg/L	
5.0PPM	5.0PPM				W	1/31/17						
TKN		PASS	4.887		4.89						mg/L	
6.7PPM	6.7PPM				W	1/31/17						
TKN		PASS	6.694		6.69						mg/L	
10.0PPM	10.0PPM				W	1/31/17						
TKN		PASS	10.043		10.0						mg/L	
ICV1	ICV1				W	1/31/17						
TKN		PASS	5.127		5.13	103.0	80	120			mg/L	
ICB1	ICB1				W	1/31/17						
TKN		PASS	0.111		0.111			+/-0.0960			mg/L	
CCV1	CCV1				W	1/31/17						
TKN		PASS	5.152		5.15	103.0	80	120			mg/L	
CCB1	CCB1				W	1/31/17						
TKN		PASS	0.111		0.111			+/-0.0960			mg/L	
PB96508BLW	PB96508BLW				W	1/31/17						
TKN		PASS	0.115		0.115			+/-0.0960			mg/L	
PB96508BSW	PB96508BSW				W	1/31/17						
TKN		PASS	5.310		5.31	106.0	80	120			mg/L	
I1272-21	0004			1	W	1/31/17						
TKN		OR	46.066		46.1						mg/L	
CCV2	CCV2				W	1/31/17						
TKN		PASS	5.410		5.41	108.0	80	120			mg/L	
CCB2	CCB2				W	1/31/17						
TKN		PASS	0.129		0.129			+/-0.0960			mg/L	
I1272-21DL	0004DL			10	W	1/31/17						
TKN		PASS	2.616		26.2						mg/L	
CCV3	CCV3				W	1/31/17						
TKN		PASS	5.316		5.32	106.0	80	120			mg/L	
CCB3	CCB3				W	1/31/17						
TKN		PASS	0.113		0.113			+/-0.0960			mg/L	