

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/31/17

Analyst : AK

Data File : lb85609.csv

Approved By : 

Approved Date : 1/31/17

Worksheet # : 1

m.s.m.4500 - NH3 B.G.H - Ammonia - 131

Lab Sample ID	Client ID	Raw Amt PPB	Dil	Matrix	A. Date	Prep Method	Analysis Method	Line 1
Parameter			Final Conc	%Rec		LCL	UCL	Line 2
Ammonia								
0.1PPM	0.1PPM			W	1/31/17			
Ammonia as N	PASS	0.108	0.108					mg/L
0.2PPM	0.2PPM			W	1/31/17			
Ammonia as N	PASS	0.206	0.206					mg/L
0.4PPM	0.4PPM			W	1/31/17			
Ammonia as N	PASS	0.398	0.398					mg/L
1.0PPM	1.0PPM			W	1/31/17			
Ammonia as N	PASS	0.980	0.980					mg/L
1.3PPM	1.3PPM			W	1/31/17			
Ammonia as N	PASS	1.329	1.33					mg/L
2.0PPM	2.0PPM			W	1/31/17			
Ammonia as N	PASS	2.012	2.01					mg/L
ICV1	ICV1			W	1/31/17			
Ammonia as N	PASS	0.948	0.95	95.0	90	110		mg/L
ICB1	ICB1			W	1/31/17			
Ammonia as N	PASS	0.019	0.019			+/-0.0340		mg/L
CCV1	CCV1			W	1/31/17			
Ammonia as N	PASS	0.910	0.91	91.0	90	110		mg/L
CCB1	CCB1			W	1/31/17			
Ammonia as N	PASS	0.019	0.019			+/-0.0340		mg/L
PB96522BLW	PB96522BLW			W	1/31/17			
Ammonia as N	PASS	0.018	0.018			+/-0.0340		mg/L
PB96522BSW	PB96522BSW			W	1/31/17			
Ammonia as N	PASS	0.957	0.96	96.0	80	120		mg/L
I1272-23	0007	1		W	1/31/17			
Ammonia as N	OR	4.503	4.50					mg/L
CCV2	CCV2			W	1/31/17			
Ammonia as N	PASS	0.904	0.90	90.0	90	110		mg/L
CCB2	CCB2			W	1/31/17			
Ammonia as N	PASS	0.022	0.022			+/-0.0340		mg/L
I1272-23DL	0007DL	10		W	1/31/17			
Ammonia as N	PASS	0.480	4.80					mg/L
CCV3	CCV3			W	1/31/17			
Ammonia as N	PASS	0.970	0.97	97.0	90	110		mg/L
CCB3	CCB3			W	1/31/17			
Ammonia as N	PASS	0.014	0.014			+/-0.0340		mg/L