

===== CA85619 =====
Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

2/1/2017 10:31

Reactive
Test: ~~Total~~ CN *WIL*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.140	0.0	0.092	
ICB1	1.225	0.0	0.005	
CCV1	245.336	0.0	0.230	
CCB1	1.290	0.0	0.006	
PB96571BLS	0.797	0.0	0.005	
I1572-01	1.423	0.0	0.006	
I1572-01D	1.117	0.0	0.005	
I1588-04	0.829	0.0	0.005	
I1597-21	3.924	0.0	0.008	
I1597-22	0.359	0.0	0.005	
I1597-23	0.336	0.0	0.005	
I1597-24	0.344	0.0	0.005	
I1597-25	0.617	0.0	0.005	
I1597-26	1.314	0.0	0.006	
CCV2	242.747	0.0	0.228	
CCB2	1.125	0.0	0.005	
I1597-27	0.321	0.0	0.005	
I1597-28	0.517	0.0	0.005	
I1597-29	0.824	0.0	0.005	
I1597-30	0.844	0.0	0.005	
CCV3	240.276	0.0	0.226	
CCB3	1.035	0.0	0.005	

N	22
Mean	38.261
SD	85.5581
CV%	223.62

LB 85619

AK

Aquakem v. 7.2AQ1

Results from time period:

Wed Feb 01 09:33:08 2017

Wed Feb 01 10:24:25 2017

Sample Id	Sam/Ctr/c	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	0.3184	µg/l	2/1/2017 9:33:08	
5.OPPBCN	A	Reactive CN	P	5.3014	µg/l	2/1/2017 9:33:09	
10PPBCN	A	Reactive CN	P	10.5915	µg/l	2/1/2017 9:33:10	
50PPBCN	A	Reactive CN	P	48.8554	µg/l	2/1/2017 9:33:11	
100PPBCN	A	Reactive CN	P	99.7329	µg/l	2/1/2017 9:33:12	
250PPBCN	A	Reactive CN	P	250.0948	µg/l	2/1/2017 9:33:13	
500PPBCN	A	Reactive CN	P	500.1057	µg/l	2/1/2017 9:33:14	
ICV1	S	Reactive CN	P	95.1404	µg/l	2/1/2017 10:16:40	
ICB1	S	Reactive CN	P	1.2254	µg/l	2/1/2017 10:16:41	
CCV1	S	Reactive CN	P	245.336	µg/l	2/1/2017 10:16:42	
CCB1	S	Reactive CN	P	1.2904	µg/l	2/1/2017 10:16:43	
PB96571BLS	S	Reactive CN	P	0.7974	µg/l	2/1/2017 10:16:44	
I1572-01	S	Reactive CN	P	1.4227	µg/l	2/1/2017 10:16:45	
I1572-01D	S	Reactive CN	P	1.1169	µg/l	2/1/2017 10:16:46	
I1588-04	S	Reactive CN	P	0.8291	µg/l	2/1/2017 10:16:47	
I1597-21	S	Reactive CN	P	3.9241	µg/l	2/1/2017 10:16:48	
I1597-22	S	Reactive CN	P	0.359	µg/l	2/1/2017 10:16:49	
I1597-23	S	Reactive CN	P	0.3363	µg/l	2/1/2017 10:16:50	
I1597-24	S	Reactive CN	P	0.3438	µg/l	2/1/2017 10:24:15	
I1597-25	S	Reactive CN	P	0.6172	µg/l	2/1/2017 10:24:16	
I1597-26	S	Reactive CN	P	1.3143	µg/l	2/1/2017 10:24:17	
CCV2	S	Reactive CN	P	242.7468	µg/l	2/1/2017 10:24:18	
CCB2	S	Reactive CN	P	1.1252	µg/l	2/1/2017 10:24:19	
I1597-27	S	Reactive CN	P	0.3207	µg/l	2/1/2017 10:24:20	
I1597-28	S	Reactive CN	P	0.5169	µg/l	2/1/2017 10:24:21	
I1597-29	S	Reactive CN	P	0.8239	µg/l	2/1/2017 10:24:22	
I1597-30	S	Reactive CN	P	0.8435	µg/l	2/1/2017 10:24:23	
CCV3	S	Reactive CN	P	240.276	µg/l	2/1/2017 10:24:24	
CCB3	S	Reactive CN	P	1.0346	µg/l	2/1/2017 10:24:25	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

2/1/2017 9:42


Test Total CN

Accepted 2/1/2017 9:42

Factor 1085
Bias 0.004

Coeff. of det. 0.999991

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.005	0.3184	0.0000	
2	5.0PPBCN	0.009	5.3014	5.0000	
3	10PPBCN	0.014	10.5915	10.0000	
4	50PPBCN	0.049	48.8554	50.0000	
5	100PPBCN	0.096	99.7329	100.0000	
6	250PPBCN	0.235	250.0948	250.0000	
7	500PPBCN	0.465	500.1057	500.0000	