


Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/1/17
Analyst : AK
Data File : lb85619.csv

Approved By : 
Approved Date :
Worksheet # :

M9012 A-B Total Amenable of Reactive cyanide-18

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Reactive Cyanide												
0.0PPBCN	0.0PPBCN				W	2/1/17						
Reactive Cyanide		PASS	0.318		0.000						mg/L	
5.0PPBCN	5.0PPBCN				W	2/1/17						
Reactive Cyanide		PASS	5.301		0.005						mg/L	
10PPBCN	10PPBCN				W	2/1/17						
Reactive Cyanide		PASS	10.592		0.011						mg/L	
50PPBCN	50PPBCN				W	2/1/17						
Reactive Cyanide		PASS	48.855		0.049						mg/L	
100PPBCN	100PPBCN				W	2/1/17						
Reactive Cyanide		PASS	99.733		0.100						mg/L	
250PPBCN	250PPBCN				W	2/1/17						
Reactive Cyanide		PASS	250.095		0.250						mg/L	
500PPBCN	500PPBCN				W	2/1/17						
Reactive Cyanide		PASS	500.106		0.500						mg/L	
ICV1	ICV1				W	2/1/17						
Reactive Cyanide		PASS	95.140		0.10	100.0	85	115			mg/L	
ICB1	ICB1				W	2/1/17						
Reactive Cyanide		PASS	1.225		0.001			+/-0.0050			mg/L	
CCV1	CCV1				W	2/1/17						
Reactive Cyanide		PASS	245.336		0.25	100.0	90	110			mg/L	
CCB1	CCB1				W	2/1/17						
Reactive Cyanide		PASS	1.290		0.001			+/-0.0050			mg/L	
PB96571BLS	PB96571BLS				S	2/1/17						
Reactive Cyanide		PASS	0.797		0.008			+/-0.0500			mg/Kg	
I1572-01	COMPOSITE-1			1	S	2/1/17						
Reactive Cyanide		PASS	1.423		0.014						mg/Kg	
I1572-01D	COMPOSITE-1D			1	S	2/1/17						
Reactive Cyanide		PASS	1.117		0.011				0	20	mg/Kg	
I1588-04	WC-B15-B21-001			1	S	2/1/17						
Reactive Cyanide		PASS	0.829		0.008						mg/Kg	
I1597-21	SB-26-COMP			1	S	2/1/17						
Reactive Cyanide		PASS	3.924		0.039						mg/Kg	
I1597-22	SB-27-COMP			1	S	2/1/17						
Reactive Cyanide		PASS	0.359		0.004						mg/Kg	
I1597-23	SB-30-COMP			1	S	2/1/17						
Reactive Cyanide		PASS	0.336		0.003						mg/Kg	
I1597-24	SB-25-COMP			1	S	2/1/17						
Reactive Cyanide		PASS	0.344		0.003						mg/Kg	
I1597-25	SB-29-COMP			1	S	2/1/17						
Reactive Cyanide		PASS	0.617		0.006						mg/Kg	

Chemtech Consulting Group


Analytical Review Report

Date Printed : 2/1/17

Analyst : AK

Data File : lb85619.csv

Approved By : 

Approved Date : 

Worksheet # : 

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1	
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Reactive Cyanide												
I1597-26	SB-31-COMP			1	S	2/1/17						
Reactive Cyanide		PASS	1.314		0.013						mg/Kg	
CCV2	CCV2				W	2/1/17						
Reactive Cyanide		PASS	242.747		0.24	96.0	90	110			mg/L	
CCB2	CCB2				W	2/1/17						
Reactive Cyanide		PASS	1.125		0.001			+/-0.0050			mg/L	
I1597-27	SB-32-COMP			1	S	2/1/17						
Reactive Cyanide		PASS	0.321		0.003						mg/Kg	
I1597-28	SB-33-COMP			1	S	2/1/17						
Reactive Cyanide		PASS	0.517		0.005						mg/Kg	
I1597-29	SB-50-COMP			1	S	2/1/17						
Reactive Cyanide		PASS	0.824		0.008						mg/Kg	
I1597-30	SB-40-COMP			1	S	2/1/17						
Reactive Cyanide		PASS	0.844		0.008						mg/Kg	
CCV3	CCV3				W	2/1/17						
Reactive Cyanide		PASS	240.276		0.24	96.0	90	110			mg/L	
CCB3	CCB3				W	2/1/17						
Reactive Cyanide		PASS	1.035		0.001			+/-0.0050			mg/L	