

LB85628

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

2/1/2017 15:07

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.996	0.0	0.092	
ICB1	1.398	0.0	0.006	
CCV1	237.387	0.0	0.223	
CCB1	1.646	0.0	0.006	
PB96617BLS	1.839	0.0	0.006	
PB96617BSS	97.684	0.0	0.094	
LOWPB96617	11.122	0.0	0.015	
HIGHPB96617	485.756	0.0	0.452	
I1569-05	1.269	0.0	0.005	
I1569-05D	0.923	0.0	0.005	
I1569-05S	38.781	0.0	0.040	
I1569-05SD	38.286	0.0	0.040	
CCV2	242.637	0.0	0.228	
CCB2	2.685	0.0	0.007	

N	14
Mean	89.743
SD	141.4506
CV%	157.62

CB 85628

AK

Aquakem v. 7.2AQ1

Results from time period:

Wed Feb 01 14:14:46 2017

Wed Feb 01 14:47:12 2017

Sample Id	Sam/Ctr/c/	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	0.3184	µg/l	2/1/2017 9:33:08	
5.OPPBCN	A	Total CN	P	5.3014	µg/l	2/1/2017 9:33:09	
10PPBCN	A	Total CN	P	10.5915	µg/l	2/1/2017 9:33:10	
50PPBCN	A	Total CN	P	48.8554	µg/l	2/1/2017 9:33:11	
100PPBCN	A	Total CN	P	99.7329	µg/l	2/1/2017 9:33:12	
250PPBCN	A	Total CN	P	250.0948	µg/l	2/1/2017 9:33:13	
500PPBCN	A	Total CN	P	500.1057	µg/l	2/1/2017 9:33:14	
ICV1	S	Total CN	P	94.9961	µg/l	2/1/2017 14:14:46	
ICB1	S	Total CN	P	1.3978	µg/l	2/1/2017 14:14:47	
CCV1	S	Total CN	P	237.3873	µg/l	2/1/2017 14:44:24	
CCB1	S	Total CN	P	1.6464	µg/l	2/1/2017 14:44:25	
PB96617BLS	S	Total CN	P	1.8386	µg/l	2/1/2017 14:44:26	
PB96617BSS	S	Total CN	P	97.6836	µg/l	2/1/2017 14:44:27	
LOWPB96617	S	Total CN	P	11.122	µg/l	2/1/2017 14:44:28	
HIGHPB96617	S	Total CN	P	485.756	µg/l	2/1/2017 14:44:29	
I1569-05	S	Total CN	P	1.2686	µg/l	2/1/2017 14:44:30	
I1569-05D	S	Total CN	P	0.9229	µg/l	2/1/2017 14:44:31	
I1569-05S	S	Total CN	P	38.7805	µg/l	2/1/2017 14:44:32	
I1569-05SD	S	Total CN	P	38.2858	µg/l	2/1/2017 14:44:33	
CCV2	S	Total CN	P	242.6374	µg/l	2/1/2017 14:44:34	
CCB2	S	Total CN	P	2.6855	µg/l	2/1/2017 14:47:12	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

2/1/2017 9:42


Test Total CN

Accepted 2/1/2017 9:42

Factor 1085
Bias 0.004

Coeff. of det. 0.999991

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.005	0.3184	0.0000	
2	5.0PPBCN	0.009	5.3014	5.0000	
3	10PPBCN	0.014	10.5915	10.0000	
4	50PPBCN	0.049	48.8554	50.0000	
5	100PPBCN	0.096	99.7329	100.0000	
6	250PPBCN	0.235	250.0948	250.0000	
7	500PPBCN	0.465	500.1057	500.0000	