

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/1/17

Analyst : AK

Data File : lb85628.csv

Approved By : 

Approved Date : 

Worksheet # :

m9012 A-B Total Amenable d Reactive cyanide - 16

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep Method	Analysis Method		Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Line 2
Cyanide										
0.0PPBCN	0.0PPBCN				W	2/1/17				
Cyanide		PASS	0.318		0.000					mg/L
5.0PPBCN	5.0PPBCN				W	2/1/17				
Cyanide		PASS	5.301		0.005					mg/L
10PPBCN	10PPBCN				W	2/1/17				
Cyanide		PASS	10.592		0.011					mg/L
50PPBCN	50PPBCN				W	2/1/17				
Cyanide		PASS	48.855		0.049					mg/L
100PPBCN	100PPBCN				W	2/1/17				
Cyanide		PASS	99.733		0.100					mg/L
250PPBCN	250PPBCN				W	2/1/17				
Cyanide		PASS	250.095		0.250					mg/L
500PPBCN	500PPBCN				W	2/1/17				
Cyanide		PASS	500.106		0.500					mg/L
ICV1	ICV1				W	2/1/17				
Cyanide		PASS	94.996		0.10	100.0	90	110		mg/L
ICB1	ICB1				W	2/1/17				
Cyanide		PASS	1.398		0.001			+/-0.0030		mg/L
CCV1	CCV1				W	2/1/17				
Cyanide		PASS	237.387		0.24	96.0	90	110		mg/L
CCB1	CCB1				W	2/1/17				
Cyanide		PASS	1.646		0.002			+/-0.0030		mg/L
PB96617BLS	PB96617BLS				S	2/1/17				
Cyanide		PASS	1.839		0.092			+/-0.0330		mg/Kg
PB96617BSS	PB96617BSS				S	2/1/17				
Cyanide		PASS	97.684		4.88	98.0	85.00	115.00		mg/Kg
LOWPB96617	LOWPB96617				W	2/1/17				
Cyanide		PASS	11.122		0.011					mg/L
HIGHPB96617	HIGHPB96617				W	2/1/17				
Cyanide		PASS	485.756		0.486					mg/L
I1569-05	ELT-SB-03-10-12				S	2/1/17				
Cyanide		PASS	1.269		0.072					mg/Kg
I1569-05D	ELT-SB-03-10-12D				S	2/1/17				
Cyanide		FAIL	0.923		0.053				30.00 20	mg/Kg
I1569-05S	ELT-SB-03-10-12S				S	2/1/17				
Cyanide		PASS	38.780		2.210	94.0	75	125		mg/Kg
I1569-05SD	ELT-SB-03-10-12SD				S	2/1/17				
Cyanide		PASS	38.286		2.190	93.0	75	125	0.9 20	mg/Kg
CCV2	CCV2				W	2/1/17				
Cyanide		PASS	242.637		0.24	96.0	90	110		mg/L


Chemtech Consulting Group

Analytical Review Report


Date Printed : 2/1/17

Analyst : AK

Data File : lb85628.csv

Approved By : 

Approved Date : 2/1/17

Worksheet # : 

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Cyanide	CCB2			W	2/1/17						
Cyanide		PASS		2.685	0.003		+/-0.0030			mg/L	