

LB856 29

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

2/1/2017 13:43

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.957	0.0	0.115	
ICB1	-0.032	0.0	0.005	
CCV1	0.958	0.0	0.115	
CCB1	-0.027	0.0	0.006	
PB96579BLW	-0.033	0.0	0.005	
PB96579BSW	0.957	0.0	0.115	
I1272-17	2.707	0.0	0.309	Test limit high
I1272-17 2X	1.430	0.0	0.168	
CCV2	0.974	0.0	0.117	
CCB2	-0.036	0.0	0.005	
N	10			
Mean	0.785			
SD	0.8760			
CV%	111.52			

LB 856 29

AK

Aquakem v. 7.2AQ1

Results from time period:

Wed Feb 01 13:33:57 2017

Wed Feb 01 13:34:06 2017

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0463	mg/l	2/1/2017 11:26:53	
0.1PPM	A	Phenolics-	P	0.1053	mg/l	2/1/2017 11:26:54	
0.25PPM	A	Phenolics-	P	0.2516	mg/l	2/1/2017 11:26:55	
0.50PPM	A	Phenolics-	P	0.4964	mg/l	2/1/2017 11:26:56	
1.0PPM	A	Phenolics-	P	0.9997	mg/l	2/1/2017 11:26:57	
2.0PPM	A	Phenolics-	P	2.0007	mg/l	2/1/2017 11:26:58	
ICV1	S	Phenolics-	P	0.957	mg/l	2/1/2017 13:33:57	
ICB1	S	Phenolics-	P	-0.032	mg/l	2/1/2017 13:33:58	
CCV1	S	Phenolics-	P	0.9579	mg/l	2/1/2017 13:33:59	
CCB1	S	Phenolics-	P	-0.027	mg/l	2/1/2017 13:34:00	
PB96579BLW	S	Phenolics-	P	-0.0334	mg/l	2/1/2017 13:34:01	
PB96579BSW	S	Phenolics-	P	0.9574	mg/l	2/1/2017 13:34:02	
I1272-17	S	Phenolics-	P	2.7074	mg/l	2/1/2017 13:34:03	
I1272-17 2X	S	Phenolics-	P	1.4297	mg/l	2/1/2017 13:34:04	
CCV2	S	Phenolics-	P	0.9739	mg/l	2/1/2017 13:34:05	
CCB2	S	Phenolics-	P	-0.036	mg/l	2/1/2017 13:34:06	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

2/1/2017 11:31

Test Phenolics-


Accepted 2/1/2017 11:27

Factor 9.018

Bias 0.009

Coeff. of det. 0.999979

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.014	0.0463	0.0500	
2	Phenol-2	0.021	0.1053	0.1000	
3	Phenol-2	0.037	0.2516	0.2500	
4	Phenol-2	0.064	0.4964	0.5000	
5	Phenol-2	0.120	0.9997	1.0000	
6	Phenol-2	0.231	2.0007	2.0000	