

LB85680

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

2/1/2017 13:05

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.966	0.0	0.116	
ICB1	-0.027	0.0	0.006	
CCV1	0.973	0.0	0.117	
CCB1	-0.021	0.0	0.007	
PB96578BLW	-0.024	0.0	0.006	
PB96578BSW	0.974	0.0	0.117	
I1272-16	2.792	0.0	0.319	Test limit high
I1594-01	-0.018	0.0	0.007	
I1594-01D	-0.016	0.0	0.007	
I1594-01S	0.886	0.0	0.107	
I1594-01SD	0.872	0.0	0.106	
CCV2	0.969	0.0	0.116	
CCB2	-0.019	0.0	0.007	
I1272-16 2X	1.411	0.0	0.165	
CCV3	0.980	0.0	0.118	
CCB3	-0.028	0.0	0.006	

N 16
Mean 0.667
SD 0.7723
CV% 115.77

LB85630

AK

Aquakem v. 7.2AQ1

Results from time period:

Wed Feb 01 11:26:53 2017

Wed Feb 01 12:43:29 2017

Sample Id	Sam/Ctr/c	Test short n	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0463	mg/l	2/1/2017 11:26:53	
0.1PPM	A	Phenolics-	P	0.1053	mg/l	2/1/2017 11:26:54	
0.25PPM	A	Phenolics-	P	0.2516	mg/l	2/1/2017 11:26:55	
0.50PPM	A	Phenolics-	P	0.4964	mg/l	2/1/2017 11:26:56	
1.0PPM	A	Phenolics-	P	0.9997	mg/l	2/1/2017 11:26:57	
2.0PPM	A	Phenolics-	P	2.0007	mg/l	2/1/2017 11:26:58	
ICV1	S	Phenolics-	P	0.9663	mg/l	2/1/2017 12:38:32	
ICB1	S	Phenolics-	P	-0.0268	mg/l	2/1/2017 12:38:33	
CCV1	S	Phenolics-	P	0.9732	mg/l	2/1/2017 12:38:34	
CCB1	S	Phenolics-	P	-0.021	mg/l	2/1/2017 12:38:35	
PB96578BLW	S	Phenolics-	P	-0.0237	mg/l	2/1/2017 12:38:36	
PB96578BSW	S	Phenolics-	P	0.9742	mg/l	2/1/2017 12:38:37	
I1272-16	S	Phenolics-	P	2.7918	mg/l	2/1/2017 12:38:38	
I1594-01	S	Phenolics-	P	-0.0179	mg/l	2/1/2017 12:38:39	
I1594-01D	S	Phenolics-	P	-0.0156	mg/l	2/1/2017 12:38:40	
I1594-01S	S	Phenolics-	P	0.8865	mg/l	2/1/2017 12:38:41	
I1594-01SD	S	Phenolics-	P	0.8721	mg/l	2/1/2017 12:38:42	
CCV2	S	Phenolics-	P	0.9687	mg/l	2/1/2017 12:38:43	
CCB2	S	Phenolics-	P	-0.0189	mg/l	2/1/2017 12:43:26	
I1272-16 2X	S	Phenolics-	P	1.4111	mg/l	2/1/2017 12:43:27	
CCV3	S	Phenolics-	P	0.9805	mg/l	2/1/2017 12:43:28	
CCB3	S	Phenolics-	P	-0.0277	mg/l	2/1/2017 12:43:29	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

2/1/2017 11:27

Test Phenolics-


Accepted 2/1/2017 11:27

Factor 9.018

Bias 0.009

Coeff. of det. 0.999979

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.014	0.0463	0.0500	
2	Phenol-2	0.021	0.1053	0.1000	
3	Phenol-2	0.037	0.2516	0.2500	
4	Phenol-2	0.064	0.4964	0.5000	
5	Phenol-2	0.120	0.9997	1.0000	
6	Phenol-2	0.231	2.0007	2.0000	