


Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/1/17
 Analyst : AK
 Data File : lb85630.csv

Approved By : 
 Approved Date : 2/1/17
 Worksheet # :


m4201-Phenolics-9

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter		PPB		Final Conc	%Rec	Method	Method	LCL	UCL	Line 2
Phenol										
0.05PPM	0.05PPM			W	2/1/17					
Phenolics	PASS	0.046	0.046							mg/L
0.1PPM	0.1PPM			W	2/1/17					
Phenolics	PASS	0.105	0.105							mg/L
0.25PPM	0.25PPM			W	2/1/17					
Phenolics	PASS	0.252	0.252							mg/L
0.50PPM	0.50PPM			W	2/1/17					
Phenolics	PASS	0.496	0.496							mg/L
1.0PPM	1.0PPM			W	2/1/17					
Phenolics	PASS	1.000	1.000							mg/L
2.0PPM	2.0PPM			W	2/1/17					
Phenolics	PASS	2.001	2.00							mg/L
ICV1	ICV1			W	2/1/17					
Phenolics	PASS	0.966	0.97		97.0	90	110			mg/L
ICB1	ICB1			W	2/1/17					
Phenolics	PASS	-0.027	-0.027				+/-0.0020			mg/L
CCV1	CCV1			W	2/1/17					
Phenolics	PASS	0.973	0.97		97.0	90	110			mg/L
CCB1	CCB1			W	2/1/17					
Phenolics	PASS	-0.021	-0.021				+/-0.0020			mg/L
PB96578BLW	PB96578BLW			W	2/1/17					
Phenolics	PASS	-0.024	-0.024				+/-0.0020			mg/L
PB96578BSW	PB96578BSW			W	2/1/17					
Phenolics	PASS	0.974	0.97		97.0	80	120			mg/L
I1272-16	0009		1	W	2/1/17					
Phenolics	OR	2.792	2.79							mg/L
I1594-01	MH-6		1	W	2/1/17					
Phenolics	PASS	-0.018	-0.018							mg/L
I1594-01D	MH-6D		1	W	2/1/17					
Phenolics	PASS	-0.016	-0.016					0	20	mg/L
I1594-01S	MH-6S		1	W	2/1/17					
Phenolics	PASS	0.887	0.886		89.0	75	125			mg/L
I1594-01SD	MH-6SD		1	W	2/1/17					
Phenolics	PASS	0.872	0.872		87.0	75	125	1.7	20	mg/L
CCV2	CCV2			W	2/1/17					
Phenolics	PASS	0.969	0.97		97.0	90	110			mg/L
CCB2	CCB2			W	2/1/17					
Phenolics	PASS	-0.019	-0.019				+/-0.0020			mg/L
I1272-16DL	0009DL		2	W	2/1/17					
Phenolics	PASS	1.411	2.82							mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/1/17
Analyst : AK
Data File : lb85630.csv

Approved By : 
Approved Date : 2/1/17
Worksheet # :

Lab Sample ID	Client ID					Prep	Analysis				Line 1
Parameter		Raw Amt PPB	Dil	Matrix	A. Date	Method	Method				Line 2
			Final Conc		%Rec	LCL	UCL	RPD	Max RPD	Units	
Phenol											
CCV3	CCV3			W	2/1/17						
Phenolics		PASS	0.980	0.98	98.0	90	110			mg/L	
CCB3	CCB3			W	2/1/17						
Phenolics		PASS	-0.028	-0.028			+/-0.0020			mg/L	