

CB 85684

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AIC

2/1/2017 15:51

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.047	0.0	0.093	
ICB1	1.807	0.0	0.006	
CCV1	240.910	0.0	0.226	
CCB1	1.740	0.0	0.006	
PB96618BLW	1.883	0.0	0.006	
PB96618BSW	112.610	0.0	0.108	
MIDPB96618	243.254	0.0	0.228	
I1594-01	3.680	0.0	0.008	
I1594-01D	2.809	0.0	0.007	
I1594-01S	34.660	0.0	0.036	
I1594-01SD	34.745	0.0	0.036	
CCV2	257.810	0.0	0.242	
CCB2	2.910	0.0	0.007	

N	13
Mean	79.605
SD	102.3271
CV%	128.54

UB85634
AL

Aquakem v. 7.2AQ1

Results from time period:

Wed Feb 01 15:34:48 2017

Wed Feb 01 15:38:12 2017

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	0.3184	µg/l	2/1/2017 9:33:08	
5.OPPBCN	A	Total CN	P	5.3014	µg/l	2/1/2017 9:33:09	
10PPBCN	A	Total CN	P	10.5915	µg/l	2/1/2017 9:33:10	
50PPBCN	A	Total CN	P	48.8554	µg/l	2/1/2017 9:33:11	
100PPBCN	A	Total CN	P	99.7329	µg/l	2/1/2017 9:33:12	
250PPBCN	A	Total CN	P	250.0948	µg/l	2/1/2017 9:33:13	
500PPBCN	A	Total CN	P	500.1057	µg/l	2/1/2017 9:33:14	
ICV1	S	Total CN	P	96.0473	µg/l	2/1/2017 15:34:48	
ICB1	S	Total CN	P	1.8072	µg/l	2/1/2017 15:34:49	
CCV1	S	Total CN	P	240.9101	µg/l	2/1/2017 15:34:50	
CCB1	S	Total CN	P	1.7398	µg/l	2/1/2017 15:34:51	
PB96618BLW	S	Total CN	P	1.8828	µg/l	2/1/2017 15:34:52	
PB96618BSW	S	Total CN	P	112.6102	µg/l	2/1/2017 15:34:53	
MIDPB96618	S	Total CN	P	243.254	µg/l	2/1/2017 15:34:54	
I1594-01	S	Total CN	P	3.6798	µg/l	2/1/2017 15:34:55	
I1594-01D	S	Total CN	P	2.8091	µg/l	2/1/2017 15:34:56	
I1594-01S	S	Total CN	P	34.6604	µg/l	2/1/2017 15:34:57	
I1594-01SD	S	Total CN	P	34.7449	µg/l	2/1/2017 15:34:58	
CCV2	S	Total CN	P	257.81	µg/l	2/1/2017 15:38:11	
CCB2	S	Total CN	P	2.9095	µg/l	2/1/2017 15:38:12	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

2/1/2017 11:31


Test Phenolics-

Accepted 2/1/2017 11:27

Factor 9.018
Bias 0.009

Coeff. of det. 0.999979

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.014	0.0463	0.0500	
2	Phenol-2	0.021	0.1053	0.1000	
3	Phenol-2	0.037	0.2516	0.2500	
4	Phenol-2	0.064	0.4964	0.5000	
5	Phenol-2	0.120	0.9997	1.0000	
6	Phenol-2	0.231	2.0007	2.0000	