

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/1/17

Analyst : AK

Data File : LB85634.csv

Approved By : 

Approved Date : 

Worksheet # :

msmy500 - CN, CIE - Cyanide - 8

Lab Sample ID	Client ID		Raw Amt PPB	Dil	Matrix	A. Date	Prep Method	Analysis Method				Line 1
Parameter							LCL	UCL	RPD	Max RPD	Units	Line 2
Cyanide												
0.0PPBCN	0.0PPBCN				W	2/1/17						
Cyanide		PASS	0.318		0.000						mg/L	
5.0PPBCN	5.0PPBCN				W	2/1/17						
Cyanide		PASS	5.301		0.005						mg/L	
10PPBCN	10PPBCN				W	2/1/17						
Cyanide		PASS	10.592		0.011						mg/L	
50PPBCN	50PPBCN				W	2/1/17						
Cyanide		PASS	48.855		0.049						mg/L	
100PPBCN	100PPBCN				W	2/1/17						
Cyanide		PASS	99.733		0.100						mg/L	
250PPBCN	250PPBCN				W	2/1/17						
Cyanide		PASS	250.095		0.250						mg/L	
500PPBCN	500PPBCN				W	2/1/17						
Cyanide		PASS	500.106		0.500						mg/L	
ICV1	ICV1				W	2/1/17						
Cyanide		PASS	96.047		0.10	100.0	85	115			mg/L	
ICB1	ICB1				W	2/1/17						
Cyanide		PASS	1.807		0.002			+/-0.0010			mg/L	
CCV1	CCV1				W	2/1/17						
Cyanide		PASS	240.910		0.24	96.0	90	110			mg/L	
CCB1	CCB1				W	2/1/17						
Cyanide		PASS	1.740		0.002			+/-0.0010			mg/L	
PB96618BLW	PB96618BLW				W	2/1/17						
Cyanide		PASS	1.883		0.002			+/-0.0010			mg/L	
PB96618BSW	PB96618BSW				W	2/1/17						
Cyanide		PASS	112.610		0.113	113.0	85	115			mg/L	
MIDPB96618	MIDPB96618				W	2/1/17						
Cyanide		PASS	243.254		0.243						mg/L	
I1594-01	MH-6			1	W	2/1/17						
Cyanide		PASS	3.680		0.004						mg/L	
I1594-01D	MH-6D			1	W	2/1/17						
Cyanide		FAIL	2.809		0.003				28.6	20	mg/L	
I1594-01S	MH-6S			1	W	2/1/17						
Cyanide		PASS	34.660		0.035	78.0	75	125			mg/L	
I1594-01SD	MH-6SD			1	W	2/1/17						
Cyanide		PASS	34.745		0.035	78.0	75	125	0	20	mg/L	
CCV2	CCV2				W	2/1/17						
Cyanide		PASS	257.810		0.26	104.0	90	110			mg/L	
CCB2	CCB2				W	2/1/17						
Cyanide		PASS	2.909		0.003			+/-0.0010			mg/L	