

===== CB85648 =====
Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

2/2/2017 14:41

Test: Total CN

Sample Id	Result ^{ug/L}	Dil. 1 +	Response	Errors
ICV1	95.590	0.0	0.092	
ICB1	1.495	0.0	0.007	
CCV1	239.366	0.0	0.224	
CCB1	1.465	0.0	0.007	
PB96615BLS	1.467	0.0	0.007	
I1644-05	0.197	0.0	0.005	
I1644-05D	-0.360	0.0	0.005	
I1644-06	0.162	0.0	0.005	
I1658-02	-0.127	0.0	0.005	
I1658-07	-0.412	0.0	0.005	
I1658-12	-0.093	0.0	0.005	
I1658-17	-0.878	0.0	0.004	
I1659-07	-0.546	0.0	0.005	
I1661-03	0.019	0.0	0.005	
CCV2	242.870	0.0	0.227	
CCB2	2.202	0.0	0.007	
I1661-04	0.191	0.0	0.005	
I1664-02	-0.047	0.0	0.005	
I1665-02	0.379	0.0	0.006	
CCV3	247.107	0.0	0.231	
CCB3	1.640	0.0	0.007	

N	21
Mean	39.604
SD	87.6238
CV%	221.25

LB 85648

Aquakem v. 7.2AQ1

Results from time period:

Thu Feb 02 13:47:13 2017

Thu Feb 02 14:33:42 2017

AK

Sample Id	Sam/Ctr/c/	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	0.6596	µg/l	2/2/2017 13:47:13	
5.OPPBCN	A	Reactive CN	P	4.2187	µg/l	2/2/2017 13:47:14	
10PPBCN	A	Reactive CN	P	9.3607	µg/l	2/2/2017 13:47:15	
50PPBCN	A	Reactive CN	P	50.0313	µg/l	2/2/2017 13:47:16	
100PPBCN	A	Reactive CN	P	101.3669	µg/l	2/2/2017 13:47:17	
250PPBCN	A	Reactive CN	P	249.2376	µg/l	2/2/2017 13:47:18	
500PPBCN	A	Reactive CN	P	500.1253	µg/l	2/2/2017 13:47:19	
ICV1	S	Reactive CN	P	95.5899	µg/l	2/2/2017 14:26:26	
ICB1	S	Reactive CN	P	1.4952	µg/l	2/2/2017 14:26:27	
CCV1	S	Reactive CN	P	239.3663	µg/l	2/2/2017 14:26:28	
CCB1	S	Reactive CN	P	1.4652	µg/l	2/2/2017 14:26:29	
PB96615BLS	S	Reactive CN	P	1.467	µg/l	2/2/2017 14:26:30	
I1644-05	S	Reactive CN	P	0.1967	µg/l	2/2/2017 14:26:31	
I1644-05D	S	Reactive CN	P	-0.3602	µg/l	2/2/2017 14:26:32	
I1644-06	S	Reactive CN	P	0.1625	µg/l	2/2/2017 14:26:33	
I1658-02	S	Reactive CN	P	-0.1265	µg/l	2/2/2017 14:26:34	
I1658-07	S	Reactive CN	P	-0.4117	µg/l	2/2/2017 14:26:35	
I1658-12	S	Reactive CN	P	-0.093	µg/l	2/2/2017 14:26:36	
I1658-17	S	Reactive CN	P	-0.878	µg/l	2/2/2017 14:33:33	
I1659-07	S	Reactive CN	P	-0.5458	µg/l	2/2/2017 14:33:34	
I1661-03	S	Reactive CN	P	0.0193	µg/l	2/2/2017 14:33:35	
CCV2	S	Reactive CN	P	242.8696	µg/l	2/2/2017 14:33:36	
CCB2	S	Reactive CN	P	2.2021	µg/l	2/2/2017 14:33:37	
I1661-04	S	Reactive CN	P	0.1907	µg/l	2/2/2017 14:33:38	
I1664-02	S	Reactive CN	P	-0.0468	µg/l	2/2/2017 14:33:39	
I1665-02	S	Reactive CN	P	0.3793	µg/l	2/2/2017 14:33:40	
CCV3	S	Reactive CN	P	247.1069	µg/l	2/2/2017 14:33:41	
CCB3	S	Reactive CN	P	1.6398	µg/l	2/2/2017 14:33:42	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AK

2/2/2017 13:53


Test Total CN

Accepted 2/2/2017 13:53

Factor 1096
 Bias 0.005

Coeff. of det. 0.999981

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.006	0.6596	0.0000	
2	5.0PPBCN	0.009	4.2187	5.0000	
3	10PPBCN	0.014	9.3607	10.0000	
4	50PPBCN	0.051	50.0313	50.0000	
5	100PPBCN	0.098	101.3669	100.0000	
6	250PPBCN	0.233	249.2376	250.0000	
7	500PPBCN	0.462	500.1253	500.0000	