

LB 85669

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

2/3/2017 12:45

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	92.728	0.0	0.091	
ICB1	0.540	0.0	0.005	
CCV1	234.518	0.0	0.222	
CCB1	0.770	0.0	0.006	
PB96707BLS	0.062	0.0	0.005	
I1660-02	0.250	0.0	0.005	
I1660-03	2.297	0.0	0.007	
I1660-04	0.106	0.0	0.005	
I1660-05	-0.066	0.0	0.005	
I1660-06	0.302	0.0	0.005	
I1660-07	0.058	0.0	0.005	
I1660-08	0.418	0.0	0.005	
CCV2	242.886	0.0	0.230	
CCB2	0.732	0.0	0.006	
I1660-10	-0.038	0.0	0.005	
I1660-11	-0.037	0.0	0.005	
I1660-12	0.483	0.0	0.005	
I1687-03	-0.027	0.0	0.005	
I1687-04	0.279	0.0	0.005	
I1660-01	0.436	0.0	0.005	
I1660-01D	0.176	0.0	0.005	
I1660-09	0.495	0.0	0.005	
CCV3	253.381	0.0	0.240	
CCB3	0.128	0.0	0.005	

N	24
Mean	34.620
SD	82.8922
CV%	239.44

UB 85669

AK

Aquakem v. 7.2AQ1

Results from time period:

Fri Feb 03 11:55:22 2017

Fri Feb 03 12:27:10 2017

Sample Id	Sam/Ctr/c	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.4186	µg/l	2/3/2017 9:24:37	
5.OPPBCN	A	Reactive CN	P	4.488	µg/l	2/3/2017 9:24:38	
10PPBCN	A	Reactive CN	P	9.9232	µg/l	2/3/2017 9:24:39	
50PPBCN	A	Reactive CN	P	49.3802	µg/l	2/3/2017 9:24:40	
100PPBCN	A	Reactive CN	P	101.2609	µg/l	2/3/2017 9:24:41	
250PPBCN	A	Reactive CN	P	251.0998	µg/l	2/3/2017 9:24:42	
500PPBCN	A	Reactive CN	P	499.2666	µg/l	2/3/2017 9:24:43	
ICV1	S	Reactive CN	P	92.7277	µg/l	2/3/2017 11:55:22	
ICB1	S	Reactive CN	P	0.5398	µg/l	2/3/2017 11:55:23	
CCV1	S	Reactive CN	P	234.5182	µg/l	2/3/2017 11:55:24	
CCB1	S	Reactive CN	P	0.7704	µg/l	2/3/2017 11:55:25	
PB96707BLS	S	Reactive CN	P	0.0623	µg/l	2/3/2017 11:55:26	
I1660-02	S	Reactive CN	P	0.2505	µg/l	2/3/2017 11:55:29	
I1660-03	S	Reactive CN	P	2.2966	µg/l	2/3/2017 11:55:30	
I1660-04	S	Reactive CN	P	0.1059	µg/l	2/3/2017 11:55:31	
I1660-05	S	Reactive CN	P	-0.0662	µg/l	2/3/2017 11:55:32	
I1660-06	S	Reactive CN	P	0.3025	µg/l	2/3/2017 12:02:57	
I1660-07	S	Reactive CN	P	0.058	µg/l	2/3/2017 12:02:58	
I1660-08	S	Reactive CN	P	0.4178	µg/l	2/3/2017 12:02:59	
CCV2	S	Reactive CN	P	242.886	µg/l	2/3/2017 12:03:00	
CCB2	S	Reactive CN	P	0.7317	µg/l	2/3/2017 12:03:01	
I1660-10	S	Reactive CN	P	-0.0377	µg/l	2/3/2017 12:03:03	
I1660-11	S	Reactive CN	P	-0.0375	µg/l	2/3/2017 12:03:04	
I1660-12	S	Reactive CN	P	0.4828	µg/l	2/3/2017 12:03:05	
I1687-03	S	Reactive CN	P	-0.0266	µg/l	2/3/2017 12:03:06	
I1687-04	S	Reactive CN	P	0.2788	µg/l	2/3/2017 12:03:07	
I1660-01	S	Reactive CN	P	0.4363	µg/l	2/3/2017 12:27:06	
I1660-01D	S	Reactive CN	P	0.1756	µg/l	2/3/2017 12:27:07	
I1660-09	S	Reactive CN	P	0.4949	µg/l	2/3/2017 12:27:08	
CCV3	S	Reactive CN	P	253.381	µg/l	2/3/2017 12:27:09	
CCB3	S	Reactive CN	P	0.1282	µg/l	2/3/2017 12:27:10	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AK

2/3/2017 9:27


Test Total CN

Accepted 2/3/2017 9:27

Factor 1079
 Bias 0.005

Coeff. of det. 0.999980

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.005	-0.4186	0.0000	
2	5.0PPBCN	0.009	4.4880	5.0000	
3	10PPBCN	0.014	9.9232	10.0000	
4	50PPBCN	0.051	49.3802	50.0000	
5	100PPBCN	0.099	101.2609	100.0000	
6	250PPBCN	0.238	251.0998	250.0000	
7	500PPBCN	0.467	499.2666	500.0000	