

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/6/17
Analyst : AK
Data File : lb85706.csv

Approved By : 
Approved Date : 
Worksheet # :


m9012 A-13 Total Amenable Reactive Cyanide -16

Lab Sample ID	Client ID		Raw Amt PPB	Dil	Matrix	A. Date	Prep Method	Analysis Method			Line 1	
Parameter						%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Cyanide												
0.0PPBCN	0.0PPBCN				W	2/6/17						
Cyanide		PASS	-0.827		-0.001						mg/L	
5.0PPBCN	5.0PPBCN				W	2/6/17						
Cyanide		PASS	4.469		0.004						mg/L	
10PPBCN	10PPBCN				W	2/6/17						
Cyanide		PASS	9.522		0.010						mg/L	
50PPBCN	50PPBCN				W	2/6/17						
Cyanide		PASS	49.455		0.049						mg/L	
100PPBCN	100PPBCN				W	2/6/17						
Cyanide		PASS	101.084		0.101						mg/L	
250PPBCN	250PPBCN				W	2/6/17						
Cyanide		PASS	252.889		0.253						mg/L	
500PPBCN	500PPBCN				W	2/6/17						
Cyanide		PASS	498.408		0.498						mg/L	
ICV1	ICV1				W	2/6/17						
Cyanide		PASS	95.839		0.10	100.0	90	110			mg/L	
ICB1	ICB1				W	2/6/17						
Cyanide		PASS	0.723		0.001			+/-0.0030			mg/L	
CCV1	CCV1				W	2/6/17						
Cyanide		PASS	242.943		0.24	96.0	90	110			mg/L	
CCB1	CCB1				W	2/6/17						
Cyanide		PASS	0.110		0.000			+/-0.0030			mg/L	
PB96740BLS	PB96740BLS				S	2/6/17						
Cyanide		PASS	0.617		0.031			+/-0.0330			mg/Kg	
PB96740BSS	PB96740BSS				S	2/6/17						
Cyanide		PASS	97.650		4.88	98.0	85.00	115.00			mg/Kg	
LOWPB96740	LOWPB96740				W	2/6/17						
Cyanide		PASS	9.312		0.009						mg/L	
HIGHPB96740	HIGHPB96740				W	2/6/17						
Cyanide		PASS	487.877		0.488						mg/L	
I1690-01	MH-6-COMP		1		S	2/6/17						
Cyanide		PASS	0.095		0.005						mg/Kg	
I1696-01	FK-BPM-01-A5B-A6B-A7B-A8		1		S	2/6/17						
Cyanide		PASS	-0.115		-0.012						mg/Kg	
CCV2	CCV2				W	2/6/17						
Cyanide		PASS	243.421		0.24	96.0	90	110			mg/L	
CCB2	CCB2				W	2/6/17						
Cyanide		PASS	0.814		0.001			+/-0.0030			mg/L	
I1699-01	F-1		1		S	2/6/17						
Cyanide		PASS	3.699		0.202						mg/Kg	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/6/17
 Analyst : AK
 Data File : lb85706.csv

Approved By : 
 Approved Date :
 Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep Method	Analysis Method			Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units
Cyanide											
I1699-03	F-2			1	S	2/6/17					
Cyanide		PASS	2.168		0.117						mg/Kg
I1701-01	C-1			1	S	2/6/17					
Cyanide		PASS	1.667		0.089						mg/Kg
I1704-01	FK-SF-01-002-A			1	S	2/6/17					
Cyanide		PASS	1.125		0.058						mg/Kg
I1704-03	FK-SF-01-002-B			1	S	2/6/17					
Cyanide		PASS	0.678		0.036						mg/Kg
I1685-01	N-1A			1	S	2/6/17					
Cyanide		PASS	4.443		0.396						mg/Kg
I1685-01D	N-1AD			1	S	2/6/17					
Cyanide		PASS	5.211		0.465				16.0	20	mg/Kg
I1685-01S	N-1AS			1	S	2/6/17					
Cyanide		FAIL	33.803		3.050	74.0	75	125			mg/Kg
I1685-01SD	N-1ASD			1	S	2/6/17					
Cyanide		PASS	34.873		3.110	76.0	75	125	1.9	20	mg/Kg
CCV3	CCV3				W	2/6/17					
Cyanide		PASS	258.855		0.26	104.0	90	110			mg/L
CCB3	CCB3				W	2/6/17					
Cyanide		PASS	0.813		0.001			+/-0.0030			mg/L