

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/7/17
Analyst : AK
Data File : lb85723.csv

Approved By : 
Approved Date : 2/7/17
Worksheet # :

m9012 A-B Total Amenable & Reactive Cyanide

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis					
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 1	
												Line 2	
Reactive Cyanide													
0.0PPBCN	0.0PPBCN				W	2/7/17							
Reactive Cyanide		PASS	0.002		0.000						mg/L		
5.0PPBCN	5.0PPBCN				W	2/7/17							
Reactive Cyanide		PASS	3.880		0.004						mg/L		
10PPBCN	10PPBCN				W	2/7/17							
Reactive Cyanide		PASS	10.294		0.010						mg/L		
50PPBCN	50PPBCN				W	2/7/17							
Reactive Cyanide		PASS	47.994		0.048						mg/L		
100PPBCN	100PPBCN				W	2/7/17							
Reactive Cyanide		PASS	99.635		0.100						mg/L		
250PPBCN	250PPBCN				W	2/7/17							
Reactive Cyanide		PASS	255.833		0.256						mg/L		
500PPBCN	500PPBCN				W	2/7/17							
Reactive Cyanide		PASS	497.362		0.497						mg/L		
ICV1	ICV1				W	2/7/17							
Reactive Cyanide		PASS	90.217		0.09	90.0	85	115			mg/L		
ICB1	ICB1				W	2/7/17							
Reactive Cyanide		PASS	0.126		0.000			+/-0.0050			mg/L		
CCV1	CCV1				W	2/7/17							
Reactive Cyanide		PASS	243.644		0.24	96.0	90	110			mg/L		
CCB1	CCB1				W	2/7/17							
Reactive Cyanide		PASS	0.333		0.000			+/-0.0050			mg/L		
PB96754BLS	PB96754BLS				S	2/7/17							
Reactive Cyanide		PASS	0.223		0.002			+/-0.0500			mg/Kg		
I1692-02	ROLL-OFF-15-40-COMP		1		S	2/7/17							
Reactive Cyanide		PASS	-0.938		-0.009						mg/Kg		
I1692-02D	ROLL-OFF-15-40-COMPD		1		S	2/7/17							
Reactive Cyanide		PASS	-0.699		-0.007				0	20	mg/Kg		
I1692-05	ROLL-OFF-20-41-COMP		1		S	2/7/17							
Reactive Cyanide		PASS	-0.968		-0.010						mg/Kg		
I1706-09	WC-SCW-COMP-02		1		S	2/7/17							
Reactive Cyanide		PASS	-1.652		-0.016						mg/Kg		
I1706-11	WC-SCW-COMP-04		1		S	2/7/17							
Reactive Cyanide		PASS	-1.420		-0.014						mg/Kg		
I1709-03	SS-1		1		S	2/7/17							
Reactive Cyanide		PASS	-0.735		-0.007						mg/Kg		
I1728-04	ROLL-OFF-R2-COMP		1		S	2/7/17							
Reactive Cyanide		PASS	-0.991		-0.010						mg/Kg		
I1731-04	ROLL-OFF-1615-COMP		1		S	2/7/17							
Reactive Cyanide		PASS	-1.229		-0.012						mg/Kg		

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/7/17
Analyst : AK
Data File : lb85723.csv

Approved By : [Signature]
Approved Date : 2/7/17
Worksheet # : 2/7/17

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	Method	Method				Line 2
Reactive Cyanide											
I1712-04	DRUMS-SOIL-ROW-COMP		1	S	2/7/17						
Reactive Cyanide	PASS	-0.759		-0.008						mg/Kg	
CCV2	CCV2			W	2/7/17						
Reactive Cyanide	PASS	263.350		0.26	104.0	90	110			mg/L	
CCB2	CCB2			W	2/7/17						
Reactive Cyanide	PASS	0.640		0.001			+/-0.0050			mg/L	