

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/8/17

Analyst : AK

Data File : lb85748.csv

Approved By : *AB*

Approved Date : 2/10/18

Worksheet # :

m9012 A-B Total Amenable & Reactive cyanide - 16

Lab Sample ID	Client ID		Raw Amt PPB	Dil	Matrix	A. Date	Prep Method	Analysis Method			Line 1	
Parameter					Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Reactive Cyanide												
0.0PPBCN	0.0PPBCN				W	2/8/17						
Reactive Cyanide		PASS	1.563		0.002						mg/L	
5.0PPBCN	5.0PPBCN				W	2/8/17						
Reactive Cyanide		PASS	5.107		0.005						mg/L	
10PPBCN	10PPBCN				W	2/8/17						
Reactive Cyanide		PASS	10.215		0.010						mg/L	
50PPBCN	50PPBCN				W	2/8/17						
Reactive Cyanide		PASS	48.134		0.048						mg/L	
100PPBCN	100PPBCN				W	2/8/17						
Reactive Cyanide		PASS	98.505		0.099						mg/L	
250PPBCN	250PPBCN				W	2/8/17						
Reactive Cyanide		PASS	251.992		0.252						mg/L	
500PPBCN	500PPBCN				W	2/8/17						
Reactive Cyanide		PASS	499.484		0.499						mg/L	
ICV1	ICV1				W	2/8/17						
Reactive Cyanide		PASS	94.926		0.09	90.0	85	115			mg/L	
ICB1	ICB1				W	2/8/17						
Reactive Cyanide		PASS	0.512		0.001			+/-0.0050			mg/L	
CCV1	CCV1				W	2/8/17						
Reactive Cyanide		PASS	245.167		0.25	100.0	90	110			mg/L	
CCB1	CCB1				W	2/8/17						
Reactive Cyanide		PASS	0.440		0.000			+/-0.0050			mg/L	
PB96830BLS	PB96830BLS				S	2/8/17						
Reactive Cyanide		PASS	0.123		0.001			+/-0.0500			mg/Kg	
I1751-04	ROLL-OFF-COMP		1		S	2/8/17						
Reactive Cyanide		PASS	0.597		0.006						mg/Kg	
I1751-04D	ROLL-OFF-COMPD		1		S	2/8/17						
Reactive Cyanide		PASS	-0.048		0.000				0	20	mg/Kg	
CCV2	CCV2				W	2/8/17						
Reactive Cyanide		PASS	248.747		0.25	100.0	90	110			mg/L	
CCB2	CCB2				W	2/8/17						
Reactive Cyanide		PASS	0.004		0.000			+/-0.0050			mg/L	