

Chemtech Consulting Group


Analytical Review Report

Date Printed : 2/10/17

Analyst : AK

Data File : lb85775.csv

Approved By : 

Approved Date : 

Worksheet # :

m9012 A-B Total Amenable & Reactive cyanide-18

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis											Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method	LCL	UCL	RPD	Max RPD	Units						Line 2
Reactive Cyanide																			
0.0PPBCN	0.0PPBCN				W	2/10/17													
Reactive Cyanide		PASS	1.832		0.002								mg/L						
5.0PPBCN	5.0PPBCN				W	2/10/17													
Reactive Cyanide		PASS	4.706		0.005								mg/L						
10PPBCN	10PPBCN				W	2/10/17													
Reactive Cyanide		PASS	9.625		0.010								mg/L						
50PPBCN	50PPBCN				W	2/10/17													
Reactive Cyanide		PASS	47.966		0.048								mg/L						
100PPBCN	100PPBCN				W	2/10/17													
Reactive Cyanide		PASS	100.528		0.101								mg/L						
250PPBCN	250PPBCN				W	2/10/17													
Reactive Cyanide		PASS	250.470		0.250								mg/L						
500PPBCN	500PPBCN				W	2/10/17													
Reactive Cyanide		PASS	499.873		0.500								mg/L						
ICV1	ICV1				W	2/10/17													
Reactive Cyanide		PASS	94.620		0.09	90.0	85	115					mg/L						
ICB1	ICB1				W	2/10/17													
Reactive Cyanide		PASS	0.100		0.000					+/-0.0050			mg/L						
CCV1	CCV1				W	2/10/17													
Reactive Cyanide		PASS	242.321		0.24	96.0	90	110					mg/L						
CCB1	CCB1				W	2/10/17													
Reactive Cyanide		PASS	0.379		0.000					+/-0.0050			mg/L						
PB96863BLS	PB96863BLS				S	2/10/17													
Reactive Cyanide		PASS	-0.014		0.000					+/-0.0500			mg/Kg						
I1769-02	9717			1	S	2/10/17													
Reactive Cyanide		PASS	0.093		0.001								mg/Kg						
I1769-02D	9717D			1	S	2/10/17													
Reactive Cyanide		PASS	-0.038		0.000						0	20	mg/Kg						
CCV2	CCV2				W	2/10/17													
Reactive Cyanide		PASS	259.428		0.26	104.0	90	110					mg/L						
CCB2	CCB2				W	2/10/17													
Reactive Cyanide		PASS	0.939		0.001					+/-0.0050			mg/L						