

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/14/17
Analyst : AK
Data File : lb85806.csv

Approved By : 
Approved Date : 2/14/17
Worksheet # :

msm 4500 N org Bore TKN-7

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			
			PPB				Method	Method			Line 1
Parameter						%Rec	LCL	UCL	RPD	Max RPD	Units
Line 2											
TKN											
0.5PPM	0.5PPM				W	2/13/17					
TKN	PASS		0.503		0.503						mg/L
1.0PPM	1.0PPM				W	2/13/17					
TKN	PASS		0.989		0.989						mg/L
2.5PPM	2.5PPM				W	2/13/17					
TKN	PASS		2.380		2.38						mg/L
5.0PPM	5.0PPM				W	2/13/17					
TKN	PASS		5.019		5.02						mg/L
6.7PPM	6.7PPM				W	2/13/17					
TKN	PASS		6.928		6.93						mg/L
10.0PPM	10.0PPM				W	2/13/17					
TKN	PASS		9.847		9.85						mg/L
ICV1	ICV1				W	2/13/17					
TKN	PASS		5.278		5.28	106.0	80	120			mg/L
ICB1	ICB1				W	2/13/17					
TKN	PASS		0.042		0.042			+/-0.0960			mg/L
CCV1	CCV1				W	2/13/17					
TKN	PASS		4.780		4.78	96.0	80	120			mg/L
CCB1	CCB1				W	2/13/17					
TKN	PASS		0.042		0.042			+/-0.0960			mg/L
PB96890BLW	PB96890BLW				W	2/13/17					
TKN	PASS		0.042		0.084			+/-0.0960			mg/L
PB96890BSW	PB96890BSW				W	2/13/17					
TKN	PASS		5.258		5.26	105.0	80	120			mg/L
I1773-01	DSN001		1		W	2/13/17					
TKN	PASS		0.295		0.295						mg/L
I1773-01D	DSN001D		1		W	2/13/17					
TKN	PASS		0.299		0.300				3.4	20	mg/L
I1773-01S	DSN001S		1		W	2/13/17					
TKN	PASS		5.391		5.390	102.0	75	125			mg/L
I1773-01SD	DSN001SD		1		W	2/13/17					
TKN	PASS		5.436		5.440	103.0	75	125	0.9	20	mg/L
CCV2	CCV2				W	2/13/17					
TKN	PASS		4.693		4.69	94.0	80	120			mg/L
CCB2	CCB2				W	2/13/17					
TKN	PASS		0.044		0.044			+/-0.0960			mg/L