

LB85813

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

2/14/2017 12:47

Test: ~~Total CN~~ ^{Reactive}

Sample Id	Result ^{Mg/L}	Dil. 1 +	Response	Errors
ICV1	95.820	0.0	0.095	
ICB1	1.357	0.0	0.006	
CCV1	242.912	0.0	0.234	
CCB1	0.745	0.0	0.005	
PB96925BLS	1.021	0.0	0.005	
I1790-02	1.248	0.0	0.006	
I1790-02D	1.061	0.0	0.005	
CCV2	247.415	0.0	0.238	
CCB2	1.172	0.0	0.006	

N	9
Mean	65.861
SD	106.2837
CV%	161.38

LB85813

Aquakem v. 7.2AQ1

Alc

Results from time period:

Tue Feb 14 12:09:29 2017

Tue Feb 14 12:09:37 2017

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	0.9853	µg/l	2/14/2017 9:36:18	
5.0PPBCN	A	Reactive CN	P	5.5592	µg/l	2/14/2017 9:36:19	
10PPBCN	A	Reactive CN	P	10.4879	µg/l	2/14/2017 9:36:20	
50PPBCN	A	Reactive CN	P	47.7403	µg/l	2/14/2017 9:36:21	
100PPBCN	A	Reactive CN	P	100.145	µg/l	2/14/2017 9:36:22	
250PPBCN	A	Reactive CN	P	249.8013	µg/l	2/14/2017 9:36:23	
500PPBCN	A	Reactive CN	P	500.2809	µg/l	2/14/2017 9:36:24	
ICV1	S	Reactive CN	P	95.8195	µg/l	2/14/2017 12:09:29	
ICB1	S	Reactive CN	P	1.3573	µg/l	2/14/2017 12:09:30	
CCV1	S	Reactive CN	P	242.9119	µg/l	2/14/2017 12:09:31	
CCB1	S	Reactive CN	P	0.745	µg/l	2/14/2017 12:09:32	
PB96925BLS	S	Reactive CN	P	1.021	µg/l	2/14/2017 12:09:33	
I1790-02	S	Reactive CN	P	1.2483	µg/l	2/14/2017 12:09:34	
I1790-02D	S	Reactive CN	P	1.0611	µg/l	2/14/2017 12:09:35	
CCV2	S	Reactive CN	P	247.4147	µg/l	2/14/2017 12:09:36	
CCB2	S	Reactive CN	P	1.1716	µg/l	2/14/2017 12:09:37	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

2/14/2017 9:40


Test Total CN

Accepted 2/14/2017 9:40

Factor 1060
Bias 0.004

Coeff. of det. 0.999967

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.005	0.9853	0.0000	
2	5.0PPBCN	0.010	5.5592	5.0000	
3	10PPBCN	0.014	10.4879	10.0000	
4	50PPBCN	0.049	47.7403	50.0000	
5	100PPBCN	0.099	100.1450	100.0000	
6	250PPBCN	0.240	249.8013	250.0000	
7	500PPBCN	0.477	500.2809	500.0000	