

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/14/17

Analyst : AK

Data File : lb85813.csv

Approved By : 

Approved Date : 2/17/17

Worksheet # :

m9012 A-B Total Amenable & Reactive cyanide - 6

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units
Reactive Cyanide											
0.0PPBCN	0.0PPBCN				W	2/14/17					
Reactive Cyanide		PASS	0.985		0.001						mg/L
5.0PPBCN	5.0PPBCN				W	2/14/17					
Reactive Cyanide		PASS	5.559		0.006						mg/L
10PPBCN	10PPBCN				W	2/14/17					
Reactive Cyanide		PASS	10.488		0.010						mg/L
50PPBCN	50PPBCN				W	2/14/17					
Reactive Cyanide		PASS	47.740		0.048						mg/L
100PPBCN	100PPBCN				W	2/14/17					
Reactive Cyanide		PASS	100.145		0.100						mg/L
250PPBCN	250PPBCN				W	2/14/17					
Reactive Cyanide		PASS	249.801		0.250						mg/L
500PPBCN	500PPBCN				W	2/14/17					
Reactive Cyanide		PASS	500.281		0.500						mg/L
ICV1	ICV1				W	2/14/17					
Reactive Cyanide		PASS	95.820		0.10	100.0	85	115			mg/L
ICB1	ICB1				W	2/14/17					
Reactive Cyanide		PASS	1.357		0.001			+/-0.0050			mg/L
CCV1	CCV1				W	2/14/17					
Reactive Cyanide		PASS	242.912		0.24	96.0	90	110			mg/L
CCB1	CCB1				W	2/14/17					
Reactive Cyanide		PASS	0.745		0.001			+/-0.0050			mg/L
PB96925BLS	PB96925BLS				S	2/14/17					
Reactive Cyanide		PASS	1.021		0.010			+/-0.0500			mg/Kg
I1790-02	38-WCS021017-0-15		1		S	2/14/17					
Reactive Cyanide		PASS	1.248		0.012						mg/Kg
I1790-02D	38-WCS021017-0-15D		1		S	2/14/17					
Reactive Cyanide		PASS	1.061		0.011				0	20	mg/Kg
CCV2	CCV2				W	2/14/17					
Reactive Cyanide		PASS	247.415		0.25	100.0	90	110			mg/L
CCB2	CCB2				W	2/14/17					
Reactive Cyanide		PASS	1.172		0.001			+/-0.0050			mg/L