

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/14/17

Analyst : AK

Data File : lb85815.csv

Approved By : 

Approved Date : 2/15/17

Worksheet # :

msm4500 - NH₃ - B.G.H - Ammonia - 131

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	Method	Method	RPD	Max RPD	Units	Line 2
Ammonia											
0.1PPM	0.1PPM			W	2/14/17						
Ammonia as N	PASS	0.111	0.111							mg/L	
0.2PPM	0.2PPM			W	2/14/17						
Ammonia as N	PASS	0.206	0.206							mg/L	
0.4PPM	0.4PPM			W	2/14/17						
Ammonia as N	PASS	0.409	0.409							mg/L	
1.0PPM	1.0PPM			W	2/14/17						
Ammonia as N	PASS	0.958	0.958							mg/L	
1.3PPM	1.3PPM			W	2/14/17						
Ammonia as N	PASS	1.328	1.33							mg/L	
2.0PPM	2.0PPM			W	2/14/17						
Ammonia as N	PASS	2.022	2.02							mg/L	
ICV1	ICV1			W	2/14/17						
Ammonia as N	PASS	0.958	0.96		96.0	90	110			mg/L	
ICB1	ICB1			W	2/14/17						
Ammonia as N	PASS	0.015	0.015				+/-0.0340			mg/L	
CCV1	CCV1			W	2/14/17						
Ammonia as N	PASS	0.950	0.95		95.0	90	110			mg/L	
CCB1	CCB1			W	2/14/17						
Ammonia as N	PASS	0.018	0.018				+/-0.0340			mg/L	
PB96891BLW	PB96891BLW			W	2/14/17						
Ammonia as N	PASS	0.020	0.020				+/-0.0340			mg/L	
PB96891BSW	PB96891BSW			W	2/14/17						
Ammonia as N	PASS	0.933	0.93		93.0	80	120			mg/L	
I1775-01	001M-WILLETS-PTBLVD(FE	1	W	2/14/17							
Ammonia as N	PASS	0.304	0.304							mg/L	
I1775-01D	001M-WILLETS-PTBLVD(FE	1	W	2/14/17							
Ammonia as N	PASS	0.306	0.306					0.7	20	mg/L	
I1775-01S	001M-WILLETS-PTBLVD(FE	1	W	2/14/17							
Ammonia as N	PASS	1.198	1.200		90.0	75	125			mg/L	
I1775-01SD	001M-WILLETS-PTBLVD(FE	1	W	2/14/17							
Ammonia as N	PASS	1.232	1.230		93.0	75	125	2.5	20	mg/L	
I1775-02	002M-35THAVE(FEB)	1	W	2/14/17							
Ammonia as N	PASS	0.294	0.294							mg/L	
CCV2	CCV2			W	2/14/17						
Ammonia as N	PASS	0.945	0.95		95.0	90	110			mg/L	
CCB2	CCB2			W	2/14/17						
Ammonia as N	PASS	0.027	0.027				+/-0.0340			mg/L	