



Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/16/17
Analyst : AK
Data File : lb85851.csv

Approved By : 
Approved Date : 
Worksheet # :

m9012 A-B Total Amenable & Reactive Cyanide-B

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method				Line 2
Reactive Cyanide												
0.0PPBCN	0.0PPBCN				W	2/16/17						
Reactive Cyanide		PASS	-1.099		-0.001						mg/L	
5.0PPBCN	5.0PPBCN				W	2/16/17						
Reactive Cyanide		PASS	3.822		0.004						mg/L	
10PPBCN	10PPBCN				W	2/16/17						
Reactive Cyanide		PASS	9.144		0.009						mg/L	
50PPBCN	50PPBCN				W	2/16/17						
Reactive Cyanide		PASS	50.849		0.051						mg/L	
100PPBCN	100PPBCN				W	2/16/17						
Reactive Cyanide		PASS	101.840		0.102						mg/L	
250PPBCN	250PPBCN				W	2/16/17						
Reactive Cyanide		PASS	251.737		0.252						mg/L	
500PPBCN	500PPBCN				W	2/16/17						
Reactive Cyanide		PASS	498.708		0.499						mg/L	
ICV1	ICV1				W	2/16/17						
Reactive Cyanide		PASS	93.975		0.09	90.0	85	115			mg/L	
ICB1	ICB1				W	2/16/17						
Reactive Cyanide		PASS	-0.667		-0.001			+/-0.0050			mg/L	
CCV1	CCV1				W	2/16/17						
Reactive Cyanide		PASS	235.775		0.24	96.0	90	110			mg/L	
CCB1	CCB1				W	2/16/17						
Reactive Cyanide		PASS	-0.230		0.000			+/-0.0050			mg/L	
PB96968BLS	PB96968BLS				S	2/16/17						
Reactive Cyanide		PASS	-0.769		-0.008			+/-0.0500			mg/Kg	
I1822-01	38-WCS021317-15-70			1	S	2/16/17						
Reactive Cyanide		PASS	-0.265		-0.003						mg/Kg	
I1822-01D	38-WCS021317-15-70D			1	S	2/16/17						
Reactive Cyanide		PASS	-0.549		-0.005						mg/Kg	
I1825-02	AIRSIDELIFTSAMPLE1C			1	S	2/16/17						
Reactive Cyanide		PASS	-0.199		-0.002						mg/Kg	
I1826-04	TP1-2-COMP14			1	S	2/16/17						
Reactive Cyanide		PASS	-0.250		-0.002						mg/Kg	
I1832-02	WC-B1-002			1	S	2/16/17						
Reactive Cyanide		PASS	-0.406		-0.004						mg/Kg	
I1835-05	TEST-PIT-3-COMP			1	S	2/16/17						
Reactive Cyanide		PASS	-0.642		-0.006						mg/Kg	
I1836-04	TEST-PIT-A-COMP			1	S	2/16/17						
Reactive Cyanide		PASS	0.042		0.000						mg/Kg	
CCV2	CCV2				W	2/16/17						
Reactive Cyanide		PASS	257.906		0.26	104.0	90	110			mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/16/17
Analyst : AK
Data File : lb85851.csv

Approved By : 
Approved Date : 2/16/17
Worksheet # : 2/16/17

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis					Line 1
Parameter		PPB		Final Conc	%Rec	Method	Method					Line 2
Reactive Cyanide												
CCB2	CCB2			W	2/16/17							
Reactive Cyanide	PASS	-0.145		0.000				+/-0.0050			mg/L	