


Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/16/17

Analyst : AK

Data File : lb85852.csv

Approved By : 

Approved Date : 2/21/17

Worksheet # :

m9012 A-13, Total Amenable & Reactive Cyanide -16

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis						Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units			Line 2
Cyanide														
0.0PPBCN	0.0PPBCN				W	2/16/17								
Cyanide		PASS	-1.099		-0.001						mg/L			
5.0PPBCN	5.0PPBCN				W	2/16/17								
Cyanide		PASS	3.822		0.004						mg/L			
10PPBCN	10PPBCN				W	2/16/17								
Cyanide		PASS	9.144		0.009						mg/L			
50PPBCN	50PPBCN				W	2/16/17								
Cyanide		PASS	50.849		0.051						mg/L			
100PPBCN	100PPBCN				W	2/16/17								
Cyanide		PASS	101.840		0.102						mg/L			
250PPBCN	250PPBCN				W	2/16/17								
Cyanide		PASS	251.737		0.252						mg/L			
500PPBCN	500PPBCN				W	2/16/17								
Cyanide		PASS	498.708		0.499						mg/L			
ICV1	ICV1				W	2/16/17								
Cyanide		PASS	94.362		0.09	90.0	90	110			mg/L			
ICB1	ICB1				W	2/16/17								
Cyanide		PASS	-0.338		0.000			+/-0.0030			mg/L			
CCV1	CCV1				W	2/16/17								
Cyanide		PASS	238.119		0.24	96.0	90	110			mg/L			
CCB1	CCB1				W	2/16/17								
Cyanide		PASS	-0.437		0.000			+/-0.0030			mg/L			
PB96965BLS	PB96965BLS				S	2/16/17								
Cyanide		PASS	-0.838		-0.042			+/-0.0330			mg/Kg			
PB96965BSS	PB96965BSS				S	2/16/17								
Cyanide		PASS	98.393		4.92	98.0	85.00	115.00			mg/Kg			
LOWPB96965	LOWPB96965				W	2/16/17								
Cyanide		PASS	9.007		0.009						mg/L			
HIGHPB96965	HIGHPB96965				W	2/16/17								
Cyanide		PASS	480.027		0.480						mg/L			
I1816-01	H-1			1	S	2/16/17								
Cyanide		PASS	0.389		0.023						mg/Kg			
I1816-01D	H-1D			1	S	2/16/17								
Cyanide		PASS	-0.384		-0.023				0	20	mg/Kg			
I1816-01S	H-1S			1	S	2/16/17								
Cyanide		PASS	32.698		1.970	81.0	75	125			mg/Kg			
I1816-01SD	H-1SD			1	S	2/16/17								
Cyanide		PASS	31.122		1.870	77.0	75	125	5.2	20	mg/Kg			
I1825-01	AIRSIDELIFTSAMPLE1C			1	S	2/16/17								
Cyanide		PASS	-0.255		-0.026						mg/Kg			

Chemtech Consulting Group


Analytical Review Report

Date Printed : 2/16/17

Analyst : AK

Data File : lb85852.csv

Approved By : 

Approved Date : 

Worksheet # :

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Cyanide											
I1826-03	TP1-2-COMP14		1	S	2/16/17						
Cyanide	PASS	0.595		0.033						mg/Kg	
CCV2	CCV2			W	2/16/17						
Cyanide	PASS	239.085		0.24	96.0	90	110			mg/L	
CCB2	CCB2			W	2/16/17						
Cyanide	PASS	-0.184		0.000			+/-0.0030			mg/L	
I1835-01	TEST-PIT-3-COMP		1	S	2/16/17						
Cyanide	PASS	0.827		0.087						mg/Kg	
I1836-01	TEST-PIT-A-COMP		1	S	2/16/17						
Cyanide	PASS	-0.712		-0.075						mg/Kg	
I1838-01	FK-BPM-01-BB9-BB10-BB11-		1	S	2/16/17						
Cyanide	PASS	-0.474		-0.024						mg/Kg	
I1839-01	FK-BPM-01-AA9B-AA10B-AA		1	S	2/16/17						
Cyanide	PASS	-0.329		-0.017						mg/Kg	
I1837-01	P-1		1	S	2/16/17						
Cyanide	PASS	3.036		0.156						mg/Kg	
CCV3	CCV3			W	2/16/17						
Cyanide	PASS	258.938		0.26	104.0	90	110			mg/L	
CCB3	CCB3			W	2/16/17						
Cyanide	PASS	0.081		0.000			+/-0.0030			mg/L	