


Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/16/17
Analyst : AK
Data File : lb85853.csv

Approved By : 
Approved Date : 2/16/17
Worksheet # :

m9065 - Phenolics-11

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1	
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Phenol												
0.05PPM	0.05PPM				W	2/16/17						
Phenolics	PASS		0.050		0.050						mg/L	
0.1PPM	0.1PPM				W	2/16/17						
Phenolics	PASS		0.098		0.098						mg/L	
0.25PPM	0.25PPM				W	2/16/17						
Phenolics	PASS		0.254		0.254						mg/L	
0.50PPM	0.50PPM				W	2/16/17						
Phenolics	PASS		0.497		0.497						mg/L	
1.0PPM	1.0PPM				W	2/16/17						
Phenolics	PASS		1.002		1.00						mg/L	
2.0PPM	2.0PPM				W	2/16/17						
Phenolics	PASS		1.999		2.00						mg/L	
ICV1	ICV1				W	2/16/17						
Phenolics	PASS		0.913		0.91	91.0	90	110			mg/L	
ICB1	ICB1				W	2/16/17						
Phenolics	PASS		-0.004		-0.004			+/-0.0100			mg/L	
CCV1	CCV1				W	2/16/17						
Phenolics	PASS		0.984		0.98	98.0	90	110			mg/L	
CCB1	CCB1				W	2/16/17						
Phenolics	PASS		-0.009		-0.009			+/-0.0100			mg/L	
PB96970BLS	PB96970BLS				S	2/16/17						
Phenolics	PASS		-0.004		-0.220			+/-0.0780			mg/Kg	
PB96970BSS	PB96970BSS				S	2/16/17						
Phenolics	PASS		0.919		46.00	92.0	80.00	120.00			mg/Kg	
I1826-03	TP1-2-COMP14	1			S	2/16/17						
Phenolics	PASS		-0.002		-0.137						mg/Kg	
I1826-03D	TP1-2-COMP14D	1			S	2/16/17						
Phenolics	PASS		-0.007		-0.415				0	20	mg/Kg	
I1826-03S	TP1-2-COMP14S	1			S	2/16/17						
Phenolics	PASS		0.882		50.900	88.0	75	125			mg/Kg	
I1826-03SD	TP1-2-COMP14SD	1			S	2/16/17						
Phenolics	PASS		0.887		50.600	89.0	75	125	1	20	mg/Kg	
I1850-05	TP21-COMP15	1			S	2/16/17						
Phenolics	PASS		-0.001		-0.028						mg/Kg	
CCV2	CCV2				W	2/16/17						
Phenolics	PASS		0.979		0.98	98.0	90	110			mg/L	
CCB2	CCB2				W	2/16/17						
Phenolics	PASS		-0.004		-0.004			+/-0.0100			mg/L	