


Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/17/17

Analyst : AK

Data File : lb85880.csv

Approved By : 

Approved Date : 2/17/17

Worksheet # :



mg012 A-B Total Amenable @ Reactive cyanide

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				
			PPB				Method	Method				Line 1
Parameter				Final Conc		%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Cyanide												
0.0PPBCN	0.0PPBCN				W	2/17/17						
Cyanide		PASS	-0.552	-0.001							mg/L	
5.0PPBCN	5.0PPBCN				W	2/17/17						
Cyanide		PASS	3.877	0.004							mg/L	
10PPBCN	10PPBCN				W	2/17/17						
Cyanide		PASS	9.233	0.009							mg/L	
50PPBCN	50PPBCN				W	2/17/17						
Cyanide		PASS	51.515	0.052							mg/L	
100PPBCN	100PPBCN				W	2/17/17						
Cyanide		PASS	100.639	0.101							mg/L	
250PPBCN	250PPBCN				W	2/17/17						
Cyanide		PASS	251.081	0.251							mg/L	
500PPBCN	500PPBCN				W	2/17/17						
Cyanide		PASS	499.207	0.499							mg/L	
ICV1	ICV1				W	2/17/17						
Cyanide		PASS	96.311	0.10		100.0	90	110			mg/L	
ICB1	ICB1				W	2/17/17						
Cyanide		PASS	-0.410	0.000				+/-0.0030			mg/L	
CCV1	CCV1				W	2/17/17						
Cyanide		PASS	243.171	0.24		96.0	90	110			mg/L	
CCB1	CCB1				W	2/17/17						
Cyanide		PASS	-0.199	0.000				+/-0.0030			mg/L	
PB96982BLS	PB96982BLS				S	2/17/17						
Cyanide		PASS	-0.709	-0.035				+/-0.0330			mg/Kg	
PB96982BSS	PB96982BSS				S	2/17/17						
Cyanide		PASS	97.719	4.89		98.0	85.00	115.00			mg/Kg	
LOWPB96982	LOWPB96982				W	2/17/17						
Cyanide		PASS	9.449	0.009							mg/L	
HIGHPB96982	HIGHPB96982				W	2/17/17						
Cyanide		PASS	490.250	0.490							mg/L	
I1850-05	TP21-COMP15		1		S	2/17/17						
Cyanide		PASS	3.032	0.173							mg/Kg	
I1850-05D	TP21-COMP15D		1		S	2/17/17						
Cyanide		PASS	2.710	0.149					14.9	20	mg/Kg	
I1850-05S	TP21-COMP15S		1		S	2/17/17						
Cyanide		FAIL	28.040	1.600		63.0	75	125			mg/Kg	
I1850-05SD	TP21-COMP15SD		1		S	2/17/17						
Cyanide		FAIL	27.808	1.570		62.0	75	125	1.9	20	mg/Kg	
I1859-01	280514-COMP-0-2FT		1		S	2/17/17						
Cyanide		PASS	-0.187	-0.019							mg/Kg	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/17/17
Analyst : AK
Data File : lb85880.csv

Approved By : 
Approved Date : 
Worksheet # : 

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter		PPB		Final Conc	%Rec	Method	Method			Line 2
Cyanide										
I1859-06	280517-COMP-0-2FT		1	S	2/17/17					
Cyanide	PASS	0.222		0.012					mg/Kg	
CCV2	CCV2			W	2/17/17					
Cyanide	PASS	250.132		0.25	100.0	90	110		mg/L	
CCB2	CCB2			W	2/17/17					
Cyanide	PASS	-0.068		0.000			+/-0.0030		mg/L	
I1859-11	280385-COMP-0-2FT		1	S	2/17/17					
Cyanide	PASS	-0.036		-0.002					mg/Kg	
CCV3	CCV3			W	2/17/17					
Cyanide	PASS	248.904		0.25	100.0	90	110		mg/L	
CCB3	CCB3			W	2/17/17					
Cyanide	PASS	0.022		0.000			+/-0.0030		mg/L	
I1850-05A	TP21-COMP15A		1	S	2/17/17					
Cyanide	PASS	41.738		2.380	97.0	75	125		mg/Kg	
CCV4	CCV4			W	2/17/17					
Cyanide	PASS	258.415		0.26	104.0	90	110		mg/L	
CCB4	CCB4			W	2/17/17					
Cyanide	PASS	0.125		0.000			+/-0.0030		mg/L	