


Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/17/17

Analyst : AK

Data File : lb85881.csv

Approved By : 

Approved Date : 2/17/17

Worksheet # :


m9012-A-B Total Amenable & Reactive Cyanide -16

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis						Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units			Line 2
Reactive Cyanide														
0.0PPBCN	0.0PPBCN				W	2/17/17								
Reactive Cyanide		PASS	-0.552		-0.001						mg/L			
5.0PPBCN	5.0PPBCN				W	2/17/17								
Reactive Cyanide		PASS	3.877		0.004						mg/L			
10PPBCN	10PPBCN				W	2/17/17								
Reactive Cyanide		PASS	9.233		0.009						mg/L			
50PPBCN	50PPBCN				W	2/17/17								
Reactive Cyanide		PASS	51.515		0.052						mg/L			
100PPBCN	100PPBCN				W	2/17/17								
Reactive Cyanide		PASS	100.639		0.101						mg/L			
250PPBCN	250PPBCN				W	2/17/17								
Reactive Cyanide		PASS	251.081		0.251						mg/L			
500PPBCN	500PPBCN				W	2/17/17								
Reactive Cyanide		PASS	499.207		0.499						mg/L			
ICV1	ICV1				W	2/17/17								
Reactive Cyanide		PASS	95.188		0.10	100.0	85	115			mg/L			
ICB1	ICB1				W	2/17/17								
Reactive Cyanide		PASS	1.879		0.002			+/-0.0050			mg/L			
CCV1	CCV1				W	2/17/17								
Reactive Cyanide		PASS	241.861		0.24	96.0	90	110			mg/L			
CCB1	CCB1				W	2/17/17								
Reactive Cyanide		PASS	0.497		0.000			+/-0.0050			mg/L			
I1850-06	TP21-COMP15			1	S	2/17/17								
Reactive Cyanide		PASS	0.388		0.004						mg/Kg			
I1850-06D	TP21-COMP15D			1	S	2/17/17								
Reactive Cyanide		PASS	0.253		0.003				0	20	mg/Kg			
I1859-05	280514-COMP-0-2FT			1	S	2/17/17								
Reactive Cyanide		PASS	-0.173		-0.002						mg/Kg			
I1859-10	280517-COMP-0-2FT			1	S	2/17/17								
Reactive Cyanide		PASS	-0.140		-0.001						mg/Kg			
I1865-02	TILCON-SCREENINGS			1	S	2/17/17								
Reactive Cyanide		PASS	-0.737		-0.007						mg/Kg			
I1869-02	TILCON-NO-57			1	S	2/17/17								
Reactive Cyanide		PASS	-0.620		-0.006						mg/Kg			
PB96988BLS	PB96988BLS				S	2/17/17								
Reactive Cyanide		PASS	-0.520		-0.005			+/-0.0500			mg/Kg			
I1859-15	280385-COMP-0-2FT			1	S	2/17/17								
Reactive Cyanide		PASS	-0.557		-0.006						mg/Kg			
CCV2	CCV2				W	2/17/17								
Reactive Cyanide		PASS	262.336		0.26	104.0	90	110			mg/L			

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/17/17
Analyst : AK
Data File : lb85881.csv

Approved By : 
Approved Date :
Worksheet # : 

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis					
Parameter		PPB		Final Conc	%Rec	Method	Method	LCL	UCL	RPD	Max RPD	Units
Reactive Cyanide												
CCB2	CCB2			W	2/17/17							
Reactive Cyanide		PASS		-0.125	0.000				+/-0.0050			mg/L