


Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/20/17

Analyst : AK

Data File : lb85902.csv

Approved By : 

Approved Date : 2/21/17

Worksheet # :


m9012 A-B Total Amenable el Reactive Cyanide -16

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis						Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method	LCL	UCL	RPD	Max RPD	Units	Line 2
Reactive Cyanide														
0.0PPBCN	0.0PPBCN				W	2/20/17								
Reactive Cyanide		PASS	-0.813		-0.001								mg/L	
5.0PPBCN	5.0PPBCN				W	2/20/17								
Reactive Cyanide		PASS	4.057		0.004								mg/L	
10PPBCN	10PPBCN				W	2/20/17								
Reactive Cyanide		PASS	8.908		0.009								mg/L	
50PPBCN	50PPBCN				W	2/20/17								
Reactive Cyanide		PASS	51.878		0.052								mg/L	
100PPBCN	100PPBCN				W	2/20/17								
Reactive Cyanide		PASS	100.940		0.101								mg/L	
250PPBCN	250PPBCN				W	2/20/17								
Reactive Cyanide		PASS	250.749		0.251								mg/L	
500PPBCN	500PPBCN				W	2/20/17								
Reactive Cyanide		PASS	499.281		0.499								mg/L	
ICV1	ICV1				W	2/20/17								
Reactive Cyanide		PASS	94.730		0.09	90.0	85	115					mg/L	
ICB1	ICB1				W	2/20/17								
Reactive Cyanide		PASS	-0.678		-0.001			+/-0.0050					mg/L	
CCV1	CCV1				W	2/20/17								
Reactive Cyanide		PASS	239.632		0.24	96.0	90	110					mg/L	
CCB1	CCB1				W	2/20/17								
Reactive Cyanide		PASS	-0.773		-0.001			+/-0.0050					mg/L	
PB97020BL	PB97020BL				W	2/20/17								
Reactive Cyanide		PASS	-0.984		-0.001			+/-0.0050					mg/L	
I1884-14	MANHOLE-001			1	W	2/20/17								
Reactive Cyanide		PASS	-0.733		-0.001								mg/L	
I1884-14D	MANHOLE-001D			1	W	2/20/17								
Reactive Cyanide		PASS	-0.654		-0.001					0	20		mg/L	
I1884-17	MANHOLE-002			1	W	2/20/17								
Reactive Cyanide		PASS	-0.764		-0.001								mg/L	
I1884-20	GROUNDWATER			1	W	2/20/17								
Reactive Cyanide		PASS	-0.852		-0.001								mg/L	
PB97024BL	PB97024BL				S	2/20/17								
Reactive Cyanide		PASS	-0.951		-0.010			+/-0.0500					mg/Kg	
I1884-03	PIT-A-COMP			1	S	2/20/17								
Reactive Cyanide		PASS	-0.890		-0.009								mg/Kg	
I1884-03D	PIT-A-COMPD			1	S	2/20/17								
Reactive Cyanide		PASS	-1.373		-0.014					0	20		mg/Kg	
I1884-06	PIT-B-COMP			1	S	2/20/17								
Reactive Cyanide		PASS	-1.135		-0.011								mg/Kg	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 2/20/17
Analyst : AK
Data File : lb85902.csv

Approved By : 
Approved Date : 2/21/17
Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units
Line 2											
Reactive Cyanide											
I1884-09	PIT-C-COMP			1	S	2/20/17					
Reactive Cyanide		PASS	-0.878		-0.009						mg/Kg
CCV2	CCV2				W	2/20/17					
Reactive Cyanide		PASS	245.797		0.25	100.0	90	110			mg/L
CCB2	CCB2				W	2/20/17					
Reactive Cyanide		PASS	-0.635		-0.001			+/-0.0050			mg/L
I1884-12	PIT-D-COMP			1	S	2/20/17					
Reactive Cyanide		PASS	-0.703		-0.007						mg/Kg
CCV3	CCV3				W	2/20/17					
Reactive Cyanide		PASS	244.787		0.25	100.0	90	110			mg/L
CCB3	CCB3				W	2/20/17					
Reactive Cyanide		PASS	-0.685		-0.001			+/-0.0050			mg/L