

LB 85926

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

2/20/2017 14:41

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	93.959	0.0	0.102	
ICB1	-0.225	0.0	0.007	
CCV1	240.534	0.0	0.251	
CCB1	-0.518	0.0	0.006	
PB97049BL	-0.220	0.0	0.007	
I1892-06	-0.455	0.0	0.006	
I1892-06DUP	-0.991	0.0	0.006	
I1913-10	-0.808	0.0	0.006	
I1913-05	-0.924	0.0	0.006	
CCV2	256.152	0.0	0.267	
CCB2	-0.610	0.0	0.006	

N	11
Mean	53.263
SD	100.5462
CV%	188.77

LB 88926

Aquakem v. 7.2AQ1

Results from time period:

AK

Mon Feb 20 14:09:33 2017

Mon Feb 20 14:41:06 2017

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	-0.8127	µg/l	2/20/2017 9:33:37	
5.0PPBCN	A	Reactive CN	P	4.0571	µg/l	2/20/2017 9:33:38	
10PPBCN	A	Reactive CN	P	8.9081	µg/l	2/20/2017 9:33:39	
50PPBCN	A	Reactive CN	P	51.8776	µg/l	2/20/2017 9:33:40	
100PPBCN	A	Reactive CN	P	100.9398	µg/l	2/20/2017 9:33:41	
250PPBCN	A	Reactive CN	P	250.7491	µg/l	2/20/2017 9:33:42	
500PPBCN	A	Reactive CN	P	499.281	µg/l	2/20/2017 9:33:43	
ICV1	S	Reactive CN	P	93.9588	µg/l	2/20/2017 14:09:33	
ICB1	S	Reactive CN	P	-0.2247	µg/l	2/20/2017 14:09:34	
CCV1	S	Reactive CN	P	240.5336	µg/l	2/20/2017 14:09:35	
CCB1	S	Reactive CN	P	-0.5176	µg/l	2/20/2017 14:09:36	
PB97049BL	S	Reactive CN	P	-0.2196	µg/l	2/20/2017 14:09:37	
I1892-06	S	Reactive CN	P	-0.4551	µg/l	2/20/2017 14:09:38	
I1892-06DUP	S	Reactive CN	P	-0.9908	µg/l	2/20/2017 14:09:39	
I1913-10	S	Reactive CN	P	-0.8083	µg/l	2/20/2017 14:09:41	
I1913-05	S	Reactive CN	P	-0.9239	µg/l	2/20/2017 14:41:03	
CCV2	S	Reactive CN	P	256.152	µg/l	2/20/2017 14:41:05	
CCB2	S	Reactive CN	P	-0.6099	µg/l	2/20/2017 14:41:06	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

2/20/2017 10:17

Test Total CN


Accepted 2/20/2017 9:34

Factor 985.7

Bias 0.007

Coeff. of det. 0.999960

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.006	-0.8127	0.0000	
2	5.0PPBCN	0.011	4.0571	5.0000	
3	10PPBCN	0.016	8.9081	10.0000	
4	50PPBCN	0.059	51.8776	50.0000	
5	100PPBCN	0.109	100.9398	100.0000	
6	250PPBCN	0.261	250.7491	250.0000	
7	500PPBCN	0.513	499.2810	500.0000	