

686031

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092

Reviewed by : AK

2/24/2017 10:36

Test: ~~Total~~ CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	93.684	0.0	0.097	
ICB1	0.984	0.0	0.007	
CCV1	242.807	0.0	0.242	
CCB1	1.078	0.0	0.007	
PB97133BL	0.477	0.0	0.007	
PB97134BL	0.550	0.0	0.007	
I1956-02	-0.970	0.0	0.005	
I1956-02D	-0.454	0.0	0.006	
I1948-06	-0.509	0.0	0.006	
I1948-06D	-0.595	0.0	0.006	
CCV2	256.932	0.0	0.255	
CCB2	1.379	0.0	0.008	

N 12
Mean 49.614
SD 97.3347
CV% 196.19

LB86081

Aquakem v. 7.2AQ1

Results from time period:

Fri Feb 24 09:30:21 2017

Fri Feb 24 10:34:21 2017

AK

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	0.8249	µg/l	2/24/2017 9:47:37	
5.OPPBCN	A	Reactive CN	P	4.1678	µg/l	2/24/2017 9:47:38	
10PPBCN	A	Reactive CN	P	9.4035	µg/l	2/24/2017 9:47:39	
50PPBCN	A	Reactive CN	P	48.9443	µg/l	2/24/2017 9:47:40	
100PPBCN	A	Reactive CN	P	100.9926	µg/l	2/24/2017 9:47:41	
250PPBCN	A	Reactive CN	P	251.4793	µg/l	2/24/2017 9:47:42	
500PPBCN	A	Reactive CN	P	499.1877	µg/l	2/24/2017 9:47:43	
ICV1	S	Reactive CN	P	93.6835	µg/l	2/24/2017 10:13:25	
ICB1	S	Reactive CN	P	0.9844	µg/l	2/24/2017 10:13:26	
CCV1	S	Reactive CN	P	242.8071	µg/l	2/24/2017 10:13:27	
CCB1	S	Reactive CN	P	1.078	µg/l	2/24/2017 10:13:28	
PB97133BL	S	Reactive CN	P	0.4766	µg/l	2/24/2017 10:13:29	
PB97134BL	S	Reactive CN	P	0.5501	µg/l	2/24/2017 10:13:32	
I1956-02	S	Reactive CN	P	-0.9704	µg/l	2/24/2017 10:13:33	
I1956-02D	S	Reactive CN	P	-0.4545	µg/l	2/24/2017 10:13:34	
I1948-06	S	Reactive CN	P	-0.5091	µg/l	2/24/2017 10:34:18	
I1948-06D	S	Reactive CN	P	-0.5946	µg/l	2/24/2017 10:34:19	
CCV2	S	Reactive CN	P	256.9324	µg/l	2/24/2017 10:34:20	
CCB2	S	Reactive CN	P	1.3787	µg/l	2/24/2017 10:34:21	

CB86031

=====
Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

2/24/2017 9:48


Test Total CN

Accepted 2/24/2017 9:48

Factor 1032
Bias 0.006

Coeff. of det. 0.999968

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.007	0.8249	0.0000	
2	5.0PPBCN	0.010	4.1678	5.0000	
3	10PPBCN	0.015	9.4035	10.0000	
4	50PPBCN	0.054	48.9443	50.0000	
5	100PPBCN	0.104	100.9926	100.0000	
6	250PPBCN	0.250	251.4793	250.0000	
7	500PPBCN	0.490	499.1877	500.0000	