

4386044

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

2/27/2017 10:02

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.903	0.0	0.104	
ICB1	-0.007	0.0	0.006	
CCV1	0.969	0.0	0.111	
CCB1	0.015	0.0	0.008	
PB97176BL	-0.006	0.0	0.006	
PB97176BS	0.909	0.0	0.105	
I1948-05	-0.009	0.0	0.005	
I1948-05DUP	-0.009	0.0	0.005	
I1948-05MS	0.957	0.0	0.110	
I1948-05MSD	0.951	0.0	0.110	
CCV2	0.969	0.0	0.111	
CCB2	-0.005	0.0	0.006	

N	12
Mean	0.470
SD	0.4947
CV%	105.31

C386044

Aquakem v. 7.2AQ1

AK

Results from time period:

Mon Feb 27 09:25:34 2017

Mon Feb 27 10:05:09 2017

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0517	mg/l	2/27/2017 9:25:34	
0.1PPM	A	Phenolics-	P	0.1004	mg/l	2/27/2017 9:25:35	
0.25PPM	A	Phenolics-	P	0.248	mg/l	2/27/2017 9:25:36	
0.50PPM	A	Phenolics-	P	0.4975	mg/l	2/27/2017 9:25:37	
1.0PPM	A	Phenolics-	P	1.0031	mg/l	2/27/2017 9:25:38	
2.0PPM	A	Phenolics-	P	1.9993	mg/l	2/27/2017 9:25:39	
ICV1	S	Phenolics-	P	0.9032	mg/l	2/27/2017 10:01:10	
ICB1	S	Phenolics-	P	-0.0065	mg/l	2/27/2017 10:01:11	
CCV1	S	Phenolics-	P	0.9689	mg/l	2/27/2017 10:01:12	
CCB1	S	Phenolics-	P	0.0147	mg/l	2/27/2017 10:01:13	
PB97176BL	S	Phenolics-	P	-0.0058	mg/l	2/27/2017 10:01:14	
PB97176BS	S	Phenolics-	P	0.9093	mg/l	2/27/2017 10:01:15	
I1948-05	S	Phenolics-	P	-0.0086	mg/l	2/27/2017 10:01:16	
I1948-05DUP	S	Phenolics-	P	-0.0092	mg/l	2/27/2017 10:01:17	
I1948-05MS	S	Phenolics-	P	0.9574	mg/l	2/27/2017 10:01:18	
I1948-05MSD	S	Phenolics-	P	0.9512	mg/l	2/27/2017 10:01:19	
CCV2	S	Phenolics-	P	0.9685	mg/l	2/27/2017 10:01:20	
CCB2	S	Phenolics-	P	-0.0052	mg/l	2/27/2017 10:01:21	

=====

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

2/27/2017 9:35


Test Phenolics-

Accepted 2/27/2017 9:35

Factor 9.212
Bias 0.006

Coeff. of det. 0.999992

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.012	0.0517	0.0500	
2	Phenol-2	0.017	0.1004	0.1000	
3	Phenol-2	0.033	0.2480	0.2500	
4	Phenol-2	0.060	0.4975	0.5000	
5	Phenol-2	0.115	1.0031	1.0000	
6	Phenol-2	0.223	1.9993	2.0000	