

LB86121

CHEMTECH

284 Sheffield Street, Mountainside, NJ 07092 Phone: 908-789-8900 Fax: 908-789-892

Analysis Method: 300 Nitrate+Nitrite
Parameter: Nitrate+Nitrite
Run Number: LB86121
Instrument: IC Metrohm

m300.0 - Inorganic Anions-16

Analytical Summary Report

Analysis Method: 300 Nitrate+Nitrite
Parameter: Nitrate+Nitrite
Run Number: LB86121
Instrument: IC Metrohm


Seq	Lab ID	Client ID	Sample Type	Inst Conc. (ppm/ppb)*	Dil	Analytical Date
1	STD1	STD1	CAL	0.000	1	2/14/17
2	STD2	STD2	CAL	0.759	1	2/14/17
3	STD3	STD3	CAL	1.452	1	2/14/17
4	STD4	STD4	CAL	2.362	1	2/14/17
5	STD5	STD5	CAL	5.279	1	2/14/17
6	STD6	STD6	CAL	10.410	1	2/14/17
7	STD7	STD7	CAL	13.003	1	2/14/17
8	STD8	STD8	CAL	28.061	1	2/14/17
9	ICV	ICV	ICV	13.260	1	2/14/17
10	ICB	ICB	ICB	0.000	1	2/14/17
11	CCV	CCV	CCV	13.089	1	3/2/17
12	CCB	CCB	CCB	0.000	1	3/2/17
13	LB86121BLW	LB86121BLW	MB	0.000	1	3/2/17
14	LB86121BSW	LB86121BSW	LCS	12.942	1	3/2/17
15	I2058-01	DSN001	SAM	0.642	1	3/2/17
16	I2058-01D	DSN001D	DUP	0.641	1	3/2/17
17	I2058-01S	DSN001S	MS	13.485	1	3/2/17
18	CCV	CCV	CCV	13.137	1	3/2/17
19	CCB	CCB	CCB	0.000	1	3/2/17

* Cal. results are in raw amount.

Chemtech Consulting Group

Analytical Review Report

Date Printed : 3/3/17
Analyst : AK
Data File : LB86121.MDB

Approved By : 
Approved Date : 3/6/17
Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis					Line 1	
Parameter			PPB		Final Conc	%Rec	Method	Method	LCL	UCL	RPD	Max RPD	Units	Line 2
Nitrate+Nitrite														
STD1	STD1				W	2/14/17								
Nitrate+Nitrite		PASS	0.000		0.000								mg/L	
STD2	STD2				W	2/14/17								
Nitrate+Nitrite		PASS	0.759		0.760								mg/L	
STD3	STD3				W	2/14/17								
Nitrate+Nitrite		PASS	1.452		1.45								mg/L	
STD4	STD4				W	2/14/17								
Nitrate+Nitrite		PASS	2.362		2.36								mg/L	
STD5	STD5				W	2/14/17								
Nitrate+Nitrite		PASS	5.279		5.28								mg/L	
STD6	STD6				W	2/14/17								
Nitrate+Nitrite		PASS	10.410		10.40								mg/L	
STD7	STD7				W	2/14/17								
Nitrate+Nitrite		PASS	13.003		13.00								mg/L	
STD8	STD8				W	2/14/17								
Nitrate+Nitrite		PASS	28.061		28.10								mg/L	
ICV1	ICV1				W	2/14/17								
Nitrate+Nitrite		PASS	13.260		13.30	96.0	90	110					mg/L	
ICB1	ICB1				W	2/14/17								
Nitrate+Nitrite		PASS	0.000		0.000			+/-0.2500					mg/L	
CCV1	CCV1				W	3/2/17								
Nitrate+Nitrite		PASS	13.089		13.10	99.0	90	110					mg/L	
CCB1	CCB1				W	3/2/17								
Nitrate+Nitrite		PASS	0.000		0.000			+/-0.2500					mg/L	
LB86121BLW	LB86121BLW				W	3/2/17								
Nitrate+Nitrite		PASS	0.000		0.000			+/-0.2500					mg/L	
LB86121BSW	LB86121BSW				W	3/2/17								
Nitrate+Nitrite		PASS	12.942		12.90	93.0	90	110					mg/L	
I2058-01	DSN001		1		W	3/2/17								
Nitrate+Nitrite		PASS	0.642		0.640								mg/L	
I2058-01D	DSN001D		1		W	3/2/17								
Nitrate+Nitrite		PASS	0.641		0.64			0		20			mg/L	
I2058-01S	DSN001S		1		W	3/2/17								
Nitrate+Nitrite		PASS	13.485		13.50	97.0	80	120					mg/L	
CCV2	CCV2				W	3/2/17								
Nitrate+Nitrite		PASS	13.137		13.10	99.0	90	110					mg/L	
CCB2	CCB2				W	3/2/17								
Nitrate+Nitrite		PASS	0.000		0.000			+/-0.2500					mg/L	

Cal results
are in raw
amount.