

LB86161

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

3/3/2017 15:54

Test: ~~Total~~ CNReactive AK
36-17
MSL

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.597	0.0	0.096	
ICB1	-0.267	0.0	0.005	
CCV1	242.051	0.0	0.235	
CCB1	-0.177	0.0	0.005	
PB97327BL	-0.570	0.0	0.004	
I2089-03	-0.629	0.0	0.004	
I2089-03DUP	-0.772	0.0	0.004	
I2089-04	-0.814	0.0	0.004	
CCV2	245.755	0.0	0.238	
CCB2	-0.841	0.0	0.004	

N	10
Mean	57.933
SD	102.5037
CV%	176.93

LB 86161

Aquakem v. 7.2AQ1

Results from time period:

Fri Mar 03 15:53:24 2017

A12

Fri Mar 03 15:53:33 2017

Sample Id	Sam/Ctr/c	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.8452	µg/l	3/3/2017 13:21:42	
5.OPPBCN	A	Reactive CN	P	3.8111	µg/l	3/3/2017 13:21:43	
10PPBCN	A	Reactive CN	P	9.3132	µg/l	3/3/2017 13:21:44	
50PPBCN	A	Reactive CN	P	49.4582	µg/l	3/3/2017 13:21:45	
100PPBCN	A	Reactive CN	P	101.8636	µg/l	3/3/2017 13:21:46	
250PPBCN	A	Reactive CN	P	253.3841	µg/l	3/3/2017 13:21:47	
500PPBCN	A	Reactive CN	P	498.015	µg/l	3/3/2017 13:21:48	
ICV1	S	Reactive CN	P	95.5973	µg/l	3/3/2017 15:53:24	
ICB1	S	Reactive CN	P	-0.267	µg/l	3/3/2017 15:53:25	
CCV1	S	Reactive CN	P	242.051	µg/l	3/3/2017 15:53:26	
CCB1	S	Reactive CN	P	-0.1775	µg/l	3/3/2017 15:53:27	
PB97327BL	S	Reactive CN	P	-0.5702	µg/l	3/3/2017 15:53:28	
I2089-03	S	Reactive CN	P	-0.6291	µg/l	3/3/2017 15:53:29	
I2089-03DUP	S	Reactive CN	P	-0.7717	µg/l	3/3/2017 15:53:30	
I2089-04	S	Reactive CN	P	-0.814	µg/l	3/3/2017 15:53:31	
CCV2	S	Reactive CN	P	245.7547	µg/l	3/3/2017 15:53:32	
CCB2	S	Reactive CN	P	-0.8409	µg/l	3/3/2017 15:53:33	

LB86161

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

3/3/2017 13:39


Test Total CN

Accepted 3/3/2017 13:39

Factor 1053
Bias 0.005

Coeff. of det. 0.999894

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.8452	0.0000	
2	5.0PPBCN	0.009	3.8111	5.0000	
3	10PPBCN	0.014	9.3132	10.0000	
4	50PPBCN	0.052	49.4582	50.0000	
5	100PPBCN	0.102	101.8636	100.0000	
6	250PPBCN	0.246	253.3841	250.0000	
7	500PPBCN	0.478	498.0150	500.0000	