

AK 3.9.17

Sample Results Report Print Date/Time: 2017/03/09 12:25:59

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
=====	=====	=====	=====	=====	=====
CCV	9.7367			TOC	
CCB	0.1095			TOC	
LB86234BLW	0.0915	0.0166	18.11	TOC	
LB86234BSW.....	10.8339...	1.6225..	14.98...	TOC	..
I2185-01	1.1337	0.1990	17.55	TOC	
I2185-01S	11.0657	0.2557	2.31	TOC	
I2185-01SD.....	10.9561...	0.1224..	1.12...	TOC	..
CCV	10.1237			TOC	
CCB	0.2988			TOC	

AK
3-9-17

Method ID	Sample Type	Vial	Timestamp	Message
TOC 0 - 20 ppmC	Sample	2	2017/03/08 14:16	
TOC 0 - 20 ppmC	Sample	3	2017/03/08 14:40	
TOC 0 - 20 ppmC	Sample	4	2017/03/08 15:31	
TOC 0 - 20 ppmC	...Sample	.. 5..	2017/03/08 16:01	..
TOC 0 - 20 ppmC	Sample	6	2017/03/08 16:52	
TOC 0 - 20 ppmC	Sample	7	2017/03/08 17:21	
TOC 0 - 20 ppmC	...Sample	.. 8..	2017/03/08 19:25	..
TOC 0 - 20 ppmC	Sample	2	2017/03/08 20:21	
TOC 0 - 20 ppmC	Sample	3	2017/03/08 20:34	

Cal. Curve ID: TOC WATER 0-20PPM
 Created: 2017/03/02 12:42
 Calibration Factor (m): 5.668e+04
 Y Intercept (b): 7065
 r-squared: 0.99641

Standard ID	Y Raw Data	X Expected ug C	Measured ug C	Message	Date & Time
0.0PPM	7646	0.000	0.010		2017/03/02 11:12
0.5PPM	22480	0.250	0.272		2017/03/02 11:27
1.0PPM	33398	0.500	0.465		2017/03/02 11:44
2.0PPM	60621	1.000	0.945		2017/03/02 11:56
5.0PPM	135654	2.500	2.269		2017/03/02 12:09
10.0PPM	316004	5.000	5.450		2017/03/02 12:23
20.0PPM	564780	10.000	9.839		2017/03/02 12:37

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
=====	=====	=====	=====	=====	=====
0.0PPM	7646			TOC	
0.5PPM	22480			TOC	
1.0PPM	33398			TOC	
2.0PPM.....	60621...	..	...	TOC	..
5.0PPM	135654			TOC	
10.0PPM	316004			TOC	
20.0PPM.....	564780...	..	...	TOC	..
ICV	9.4860			TOC	
ICB	0.3294			TOC	

Method ID	Sample Type	Vial	Timestamp	Message
=====	=====	=====	=====	=====
TOC 0 - 20 ppmC	TOC Standard	1	2017/03/02 11:12	
TOC 0 - 20 ppmC	TOC Standard	2	2017/03/02 11:27	
TOC 0 - 20 ppmC	TOC Standard	3	2017/03/02 11:44	
TOC 0 - 20 ppmC	...TOC Standard	.. 4..	2017/03/02 11:56	..
TOC 0 - 20 ppmC	TOC Standard	5	2017/03/02 12:09	
TOC 0 - 20 ppmC	TOC Standard	6	2017/03/02 12:23	
TOC 0 - 20 ppmC	...TOC Standard	.. 7..	2017/03/02 12:37	..
TOC 0 - 20 ppmC	Sample	8	2017/03/02 12:56	
TOC 0 - 20 ppmC	Sample	9	2017/03/02 13:23	