

LB86234

CHEMTECH

284 Sheffield Street, Mountainside, NJ 07092 Phone: 908-789-8900 Fax: 908-789-892

Analysis Method: 9060 TOC
Parameter: TOC
Run Number: LB86234
Instrument: Appolo-9000

M9060-A-TOC-18

Analytical Summary Report

Analysis Method: 9060 TOC
Parameter: TOC
Run Number: LB86234
Instrument: Appolo-9000


Seq	Lab ID	Client ID	Sample Type	Inst Conc. ppm/ppb*	Dil	Analytical Date
1	0.0PPM	0.0PPM	CAL	7646.000	1	3/2/17
2	0.5PPM	0.5PPM	CAL	22480.000	1	3/2/17
3	1.0PPM	1.0PPM	CAL	33398.000	1	3/2/17
4	2.0PPM	2.0PPM	CAL	60621.000	1	3/2/17
5	5.0PPM	5.0PPM	CAL	135654.000	1	3/2/17
6	10.0PPM	10.0PPM	CAL	316004.000	1	3/2/17
7	20.0PPM	20.0PPM	CAL	564780.000	1	3/2/17
8	ICV	ICV	ICV	9.486	1	3/2/17
9	ICB	ICB	ICB	0.000	1	3/2/17
10	CCV	CCV	CCV	9.737	1	3/8/17
11	CCB	CCB	CCB	0.000	1	3/8/17
12	LB86234BLW	LB86234BLW	MB	0.000	1	3/8/17
13	LB86234BSW	LB86234BSW	LCS	10.834	1	3/8/17
14	I2185-01	EFF-170307-02	SAM	1.134	1	3/8/17
15	I2185-01S	EFF-170307-02S	MS	11.066	1	3/8/17
16	I2185-01SD	EFF-170307-02SD	MSD	10.956	1	3/8/17
17	CCV	CCV	CCV	10.124	1	3/8/17
18	CCB	CCB	CCB	0.000	1	3/8/17

* Cal results are in raw amount.

Chemtech Consulting Group

Analytical Review Report

Date Printed : 3/9/17
 Analyst : AK
 Data File : LB86234.MDB

Approved By : 
 Approved Date : 3/9/17
 Worksheet # :

Lab Sample ID	Client ID		Raw Amt PPB	Dil	Matrix	A. Date	Prep Method	Analysis Method			Line 1
Parameter					Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units
TOC											
0.0PPM	0.0PPM				W	3/2/17					
TOC		PASS	7646.000		7650						mg/L
0.5PPM	0.5PPM				W	3/2/17					
TOC		PASS	22480.000		22500						mg/L
1.0PPM	1.0PPM				W	3/2/17					
TOC		PASS	33398.000		33400						mg/L
2.0PPM	2.0PPM				W	3/2/17					
TOC		PASS	60621.000		60600						mg/L
5.0PPM	5.0PPM				W	3/2/17					
TOC		PASS	135654.000		136000						mg/L
10.0PPM	10.0PPM				W	3/2/17					
TOC		PASS	316004.000		316000						mg/L
20.0PPM	20.0PPM				W	3/2/17					
TOC		PASS	564780.000		565000						mg/L
ICV1	ICV1				W	3/2/17					
TOC		PASS	9.486		9.49	95.0	90	110			mg/L
ICB1	ICB1				W	3/2/17					
TOC		PASS	0.000		0.000			+/-0.0970			mg/L
CCV1	CCV1				W	3/8/17					
TOC		PASS	9.737		9.74	97.0	90	110			mg/L
CCB1	CCB1				W	3/8/17					
TOC		PASS	0.000		0.000			+/-0.0970			mg/L
LB86234BLW	LB86234BLW				W	3/8/17					
TOC		PASS	0.000		0.000			+/-0.0970			mg/L
LB86234BSW	LB86234BSW				W	3/8/17					
TOC		PASS	10.834		10.80	108.0	90	110			mg/L
I2185-01	EFF-170307-02			1	W	3/8/17					
TOC		PASS	1.134		1.13						mg/L
I2185-01S	EFF-170307-02S			1	W	3/8/17					
TOC		PASS	11.066		11.10	100.0	75	125			mg/L
I2185-01SD	EFF-170307-02SD			1	W	3/8/17					
TOC		PASS	10.956		11.00	99.0	75	125	0.9	20	mg/L
CCV2	CCV2				W	3/8/17					
TOC		PASS	10.124		10.10	101.0	90	110			mg/L
CCB2	CCB2				W	3/8/17					
TOC		PASS	0.000		0.000			+/-0.0970			mg/L

Cal results are
in raw
amount