

Chemtech Consulting Group

Analytical Review Report

Date Printed : 3/13/17

Analyst : JS

Data File : lb86289.csv

Approved By : AS

Approved Date : 3/13/17


Worksheet # : 3/13/17

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1	
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
ASTM Ammonia												
0.1PPM	0.1PPM				W	3/10/17						
ASTM Ammonia		PASS	0.118		0.118						mg/L	
0.2PPM	0.2PPM				W	3/10/17						
ASTM Ammonia		PASS	0.193		0.193						mg/L	
0.4PPM	0.4PPM				W	3/10/17						
ASTM Ammonia		PASS	0.399		0.399						mg/L	
1.0PPM	1.0PPM				W	3/10/17						
ASTM Ammonia		PASS	0.996		0.996						mg/L	
1.3PPM	1.3PPM				W	3/10/17						
ASTM Ammonia		PASS	1.309		1.310						mg/L	
2.0PPM	2.0PPM				W	3/10/17						
ASTM Ammonia		PASS	2.018		2.020						mg/L	
ICV1	ICV1				W	3/10/17						
ASTM Ammonia		PASS	0.944		0.94	94.0	90	110			mg/L	
ICB1	ICB1				W	3/10/17						
ASTM Ammonia		PASS	0.021		0.021			+/-0.0340			mg/L	
CCV1	CCV1				W	3/10/17						
ASTM Ammonia		PASS	0.917		0.92	92.0	90	110			mg/L	
CCB1	CCB1				W	3/10/17						
ASTM Ammonia		PASS	0.014		0.014			+/-0.0340			mg/L	
PB97499BL	PB97499BL				W	3/10/17						
ASTM Ammonia		PASS	0.012		0.012			+/-0.0340			mg/L	
PB97499BS	PB97499BS				W	3/10/17						
ASTM Ammonia		PASS	0.946		0.95	95.0	80	120			mg/L	
I2128-01	D-1		1		W	3/10/17						
ASTM Ammonia		PASS	0.088		0.088						mg/L	
I2128-01D	D-1D		1		W	3/10/17						
ASTM Ammonia		PASS	0.089		0.089				1.1	20	mg/L	
I2128-01S	D-1S		1		W	3/10/17						
ASTM Ammonia		PASS	1.045		1.040	95.0	75	125			mg/L	
I2128-01SD	D-1SD		1		W	3/10/17						
ASTM Ammonia		PASS	1.045		1.040	95.0	75	125	0	20	mg/L	
I2128-02	D-8		1		W	3/10/17						
ASTM Ammonia		PASS	0.365		0.365						mg/L	
I2128-03	D-12		1		W	3/10/17						
ASTM Ammonia		PASS	0.063		0.063						mg/L	
I2128-04	D-14		1		W	3/10/17						
ASTM Ammonia		PASS	-0.001		-0.001						mg/L	
I2128-05	D-18		1		W	3/10/17						
ASTM Ammonia		PASS	0.009		0.008						mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 3/13/17
 Analyst : 
 Data File : lb86289.csv

Approved By : 
 Approved Date : 3/13/17
 Worksheet # : 

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis		Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method		Line 2
ASTM Ammonia										
CCV2	CCV2				W	3/10/17				
ASTM Ammonia		PASS	0.922	0.92	92.0	90	110			mg/L
CCB2	CCB2				W	3/10/17				
ASTM Ammonia		PASS	0.014	0.014			+/-0.0340			mg/L
I2128-06	D-21			1	W	3/10/17				
ASTM Ammonia		PASS	0.012	0.012						mg/L
I2128-07	D-23			1	W	3/10/17				
ASTM Ammonia		PASS	0.087	0.087						mg/L
I2128-08	D-25			1	W	3/10/17				
ASTM Ammonia		PASS	0.201	0.201						mg/L
I2128-09	D-26			1	W	3/10/17				
ASTM Ammonia		PASS	0.989	0.989						mg/L
I2130-01	D-27			1	W	3/10/17				
ASTM Ammonia		OR	2.185	2.190						mg/L
I2130-02	D-30			1	W	3/10/17				
ASTM Ammonia		PASS	1.683	1.680						mg/L
I2130-03	D-33			1	W	3/10/17				
ASTM Ammonia		PASS	1.640	1.640						mg/L
I2130-04	D-35			1	W	3/10/17				
ASTM Ammonia		OR	2.055	2.050						mg/L
CCV3	CCV3				W	3/10/17				
ASTM Ammonia		PASS	0.966	0.97	97.0	90	110			mg/L
CCB3	CCB3				W	3/10/17				
ASTM Ammonia		PASS	0.018	0.018			+/-0.0340			mg/L
I2130-01DL	D-27DL			2	W	3/10/17				
ASTM Ammonia		PASS	1.106	2.210						mg/L
I2130-04DL	D-35DL			2	W	3/10/17				
ASTM Ammonia		PASS	1.051	2.100						mg/L
CCV4	CCV4				W	3/10/17				
ASTM Ammonia		PASS	0.974	0.97	97.0	90	110			mg/L
CCB4	CCB4				W	3/10/17				
ASTM Ammonia		PASS	0.018	0.018			+/-0.0340			mg/L