

284 Sheffield Street, Mountainside, New Jersey 07092 Phone : 908 789 8900 Fax : 908 789 8922

Daily Analysis Runlog For Sequence/QC Batch ID # LB86346

Review By

Review On

SubDirectory

LB86346

Test :

ASTM Ammonia

STD. NAME		STD REF.#				
ICAL Standard		WP53483				
ICV Standard		WP53485				
CCV Standard		WP53484				
ICSA Standard						
CRI Standard						
LCS Standard		WP53080				
Chk Standard		WP51279,WP53185,WP52890,WP52891				

Sr#	SampleID	ClientID	QcType	Date	Comment	Status
1	0.1PPM	0.1PPM	CAL1	03/13/17 14:46		OK
2	0.2PPM	0.2PPM	CAL2	03/13/17 14:46		OK
3	0.4PPM	0.4PPM	CAL3	03/13/17 14:46		OK
4	1.0PPM	1.0PPM	CAL4	03/13/17 14:46		OK
5	1.3PPM	1.3PPM	CAL5	03/13/17 14:46		OK
6	2.0PPM	2.0PPM	CAL6	03/13/17 14:46		OK
7	ICV1	ICV1	ICV	03/13/17 15:29		OK
8	ICB1	ICB1	ICB	03/13/17 15:29		OK
9	CCV1	CCV1	CCV	03/13/17 15:29		OK
10	CCB1	CCB1	CCB	03/13/17 15:29		OK
11	PB97569BL	PB97569BL	MB	03/13/17 15:29		OK
12	PB97569BS	PB97569BS	LCS	03/13/17 15:29		OK
13	I2173-01	D-38	SAM	03/13/17 15:29		OK
14	I2173-01DUP	D-38DUP	DUP	03/13/17 15:29		OK
15	I2173-01MS	D-38MS	MS	03/13/17 15:29		OK
16	I2173-01MSD	D-38MSD	MSD	03/13/17 15:29		OK
17	I2173-04	D-47	SAM	03/13/17 15:40		OK
18	I2173-05	D-49	SAM	03/13/17 15:40		OK
19	I2173-09	D-57	SAM	03/13/17 15:40		OK
20	I2173-10	D-61	SAM	03/13/17 15:40		OK
21	CCV2	CCV2	CCV	03/13/17 15:51		OK

Daily Analysis Runlog For Sequence/QC Batch ID # LB86346

Review By			Review On			
SubDirectory		LB86346		Test : ASTM Ammonia		
STD. NAME		STD REF.#				
ICAL Standard		WP53483				
ICV Standard		WP53485				
CCV Standard		WP53484				
ICSA Standard						
CRI Standard						
LCS Standard		WP53080				
Chk Standard		WP51279,WP53185,WP52890,WP52891				
22	CCB2	CCB2	CCB	03/13/17 15:51		OK
23	I2173-02	D-41	SAM	03/13/17 16:29		OK
24	I2173-03	D-44	SAM	03/13/17 16:29		OK
25	I2173-06	D-51	SAM	03/13/17 16:29		OK
26	I2173-07	D-53	SAM	03/13/17 16:29		OK
27	I2173-08	D-55	SAM	03/13/17 16:29		OK
28	I2175-01	D-64	SAM	03/13/17 16:29		OK
29	I2175-02	D-67	SAM	03/13/17 16:29		OK
30	I2175-03	D-69	SAM	03/13/17 16:29		OK
31	I2175-04	D-71	SAM	03/13/17 16:29		OK
32	I2175-05	D-74	SAM	03/13/17 16:29		OK
33	CCV3	CCV3	CCV	03/13/17 16:29		OK
34	CCB3	CCB3	CCB	03/13/17 16:29		OK
35	I2175-06	D-77	SAM	03/13/17 16:36		OK
36	I2175-07	D-79	SAM	03/13/17 16:36		OK
37	I2175-08	D-81	SAM	03/13/17 16:36		OK
38	I2175-09	D-85	SAM	03/13/17 16:36		OK
39	CCV4	CCV4	CCV	03/13/17 16:36		OK
40	CCB4	CCB4	CCB	03/13/17 16:36		OK



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/7/2017 12:00:00 AM
Project:	LB86346	Date Received:	3/7/2017 12:00:00 AM
Client Sample ID:	D-38	SDG No.:	LB86346
Lab Sample ID:	I2173-01	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	ASTM Ammonia	0.08	J	1	0.034		0.1	mg/L	03/13/2017	03/13/2017	SM4500-NH3



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/7/2017 12:00:00 AM
Project:	LB86346	Date Received:	3/7/2017 12:00:00 AM
Client Sample ID:	D-41	SDG No.:	LB86346
Lab Sample ID:	I2173-02	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	ASTM Ammonia	0.088	J	1	0.034		0.1	mg/L	03/13/2017	03/13/2017	SM4500-NH3



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/7/2017 12:00:00 AM
Project:	LB86346	Date Received:	3/7/2017 12:00:00 AM
Client Sample ID:	D-44	SDG No.:	LB86346
Lab Sample ID:	I2173-03	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	ASTM Ammonia	0.41		1	0.034		0.1	mg/L	03/13/2017	03/13/2017	SM4500-NH3



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/7/2017 12:00:00 AM
Project:	LB86346	Date Received:	3/7/2017 12:00:00 AM
Client Sample ID:	D-47	SDG No.:	LB86346
Lab Sample ID:	I2173-04	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	ASTM Ammonia	1.6		1	0.034		0.1	mg/L	03/13/2017	03/13/2017	SM4500-NH3



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/7/2017 12:00:00 AM
Project:	LB86346	Date Received:	3/7/2017 12:00:00 AM
Client Sample ID:	D-49	SDG No.:	LB86346
Lab Sample ID:	I2173-05	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	ASTM Ammonia	1.5		1	0.034		0.1	mg/L	03/13/2017	03/13/2017	SM4500-NH3



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/7/2017 12:00:00 AM
Project:	LB86346	Date Received:	3/7/2017 12:00:00 AM
Client Sample ID:	D-51	SDG No.:	LB86346
Lab Sample ID:	I2173-06	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	ASTM Ammonia	0.85		1	0.034		0.1	mg/L	03/13/2017	03/13/2017	SM4500-NH3



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/7/2017 12:00:00 AM
Project:	LB86346	Date Received:	3/7/2017 12:00:00 AM
Client Sample ID:	D-53	SDG No.:	LB86346
Lab Sample ID:	I2173-07	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	ASTM Ammonia	0.63		1	0.034		0.1	mg/L	03/13/2017	03/13/2017	SM4500-NH3



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/7/2017 12:00:00 AM
Project:	LB86346	Date Received:	3/7/2017 12:00:00 AM
Client Sample ID:	D-55	SDG No.:	LB86346
Lab Sample ID:	I2173-08	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	ASTM Ammonia	0.48		1	0.034		0.1	mg/L	03/13/2017	03/13/2017	SM4500-NH3



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/7/2017 12:00:00 AM
Project:	LB86346	Date Received:	3/7/2017 12:00:00 AM
Client Sample ID:	D-57	SDG No.:	LB86346
Lab Sample ID:	I2173-09	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	ASTM Ammonia	1.4		1	0.034		0.1	mg/L	03/13/2017	03/13/2017	SM4500-NH3



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/7/2017 12:00:00 AM
Project:	LB86346	Date Received:	3/7/2017 12:00:00 AM
Client Sample ID:	D-61	SDG No.:	LB86346
Lab Sample ID:	I2173-10	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	ASTM Ammonia	0.17		1	0.034		0.1	mg/L	03/13/2017	03/13/2017	SM4500-NH3



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/7/2017 12:00:00 AM
Project:	LB86346	Date Received:	3/7/2017 12:00:00 AM
Client Sample ID:	D-64	SDG No.:	LB86346
Lab Sample ID:	I2175-01	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	ASTM Ammonia	0.15		1	0.034		0.1	mg/L	03/13/2017	03/13/2017	SM4500-NH3



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/7/2017 12:00:00 AM
Project:	LB86346	Date Received:	3/7/2017 12:00:00 AM
Client Sample ID:	D-67	SDG No.:	LB86346
Lab Sample ID:	I2175-02	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	ASTM Ammonia	0.094	J	1	0.034		0.1	mg/L	03/13/2017	03/13/2017	SM4500-NH3



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/7/2017 12:00:00 AM
Project:	LB86346	Date Received:	3/7/2017 12:00:00 AM
Client Sample ID:	D-69	SDG No.:	LB86346
Lab Sample ID:	I2175-03	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	ASTM Ammonia	0.074	J	1	0.034		0.1	mg/L	03/13/2017	03/13/2017	SM4500-NH3



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/7/2017 12:00:00 AM
Project:	LB86346	Date Received:	3/7/2017 12:00:00 AM
Client Sample ID:	D-71	SDG No.:	LB86346
Lab Sample ID:	I2175-04	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	ASTM Ammonia	0.17		1	0.034		0.1	mg/L	03/13/2017	03/13/2017	SM4500-NH3



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/7/2017 12:00:00 AM
Project:	LB86346	Date Received:	3/7/2017 12:00:00 AM
Client Sample ID:	D-74	SDG No.:	LB86346
Lab Sample ID:	I2175-05	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	ASTM Ammonia	0.092	J	1	0.034		0.1	mg/L	03/13/2017	03/13/2017	SM4500-NH3



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/7/2017 12:00:00 AM
Project:	LB86346	Date Received:	3/7/2017 12:00:00 AM
Client Sample ID:	D-77	SDG No.:	LB86346
Lab Sample ID:	I2175-06	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	ASTM Ammonia	0.11		1	0.034		0.1	mg/L	03/13/2017	03/13/2017	SM4500-NH3



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/7/2017 12:00:00 AM
Project:	LB86346	Date Received:	3/7/2017 12:00:00 AM
Client Sample ID:	D-79	SDG No.:	LB86346
Lab Sample ID:	I2175-07	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	ASTM Ammonia	0.57		1	0.034		0.1	mg/L	03/13/2017	03/13/2017	SM4500-NH3



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/7/2017 12:00:00 AM
Project:	LB86346	Date Received:	3/7/2017 12:00:00 AM
Client Sample ID:	D-81	SDG No.:	LB86346
Lab Sample ID:	I2175-08	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	ASTM Ammonia	0.49		1	0.034		0.1	mg/L	03/13/2017	03/13/2017	SM4500-NH3



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/7/2017 12:00:00 AM
Project:	LB86346	Date Received:	3/7/2017 12:00:00 AM
Client Sample ID:	D-85	SDG No.:	LB86346
Lab Sample ID:	I2175-09	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	ASTM Ammonia	1.0		1	0.034		0.1	mg/L	03/13/2017	03/13/2017	SM4500-NH3



Report of Analysis

Client:		Date Collected:	3/7/2017 12:00:00 AM
Project:	LB86346	Date Received:	3/7/2017 12:00:00 AM
Client Sample ID:	D-85	SDG No.:	LB86346
Lab Sample ID:	I2175-09	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
-----	-----------	-------	------	----	-----	-----	------------	-------	-----------	-----------	----------

Color Before:

Clarity Before:

Texture:

Color After:

Clarity After:

Artifacts:

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Sample ID	Analyte	Result	True Value	% Recovery	% RSD	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
		mg/L								
ICV1	ASTM Ammonia	0.95	1	95	0	90 - 110		03/13/2017	15:29	LB86346



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB86346

SDG No.: LB86346

Contract: _____ Lab Code: CHEM

Case No.: LB86346

SAS No.: LB86346

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	% RSD	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV1	ASTM Ammonia	0.96	1	96	0	90 - 110		03/13/2017	15:29	LB86346
CCV2	ASTM Ammonia	0.95	1	95	0	90 - 110		03/13/2017	15:51	LB86346
CCV3	ASTM Ammonia	0.98	1	98	0	90 - 110		03/13/2017	16:29	LB86346
CCV4	ASTM Ammonia	1.0	1	100	0	90 - 110		03/13/2017	16:36	LB86346



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB86346 **SDG No.:** LB86346
Contract: **Lab Code:** CHEM **Case No.:** LB86346 **SAS No.:** LB86346

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB1	ASTM Ammonia	0.028	+/-0.1	U		0.1		03/13/2017	15:29	LB86346



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB86346

SDG No.: LB86346

Contract:

Lab Code: CHEM

Case No.: LB86346

SAS No.: LB86346

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB1	ASTM Ammonia	0.024	+/-0.1	U			0.1	03/13/2017	15:29	LB86346
CCB2	ASTM Ammonia	0.032	+/-0.1	U			0.1	03/13/2017	15:51	LB86346
CCB3	ASTM Ammonia	0.035	+/-0.1	J			0.1	03/13/2017	16:29	LB86346
CCB4	ASTM Ammonia	0.037	+/-0.1	J			0.1	03/13/2017	16:36	LB86346



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:

LB86346

SDG No.:

LB86346

Contract:

Lab Code:

CHEM

Case No.:

LB86346

SAS No.:

LB86346

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	-----	------	---	------------------	------------------	---------------



GENCHEM
- 3b -
PREPARATION BLANK SUMMARY

Client: LB86346

SDG No.: LB86346

Instrument: Konelab 20

Sample ID	Analyte	Result (mg/L)	Acceptance Limit	Conc Qual	LOD mg/L	CRQL mg/L	M	Analysis Date	Analysis Time	Run
PB97569BL	ASTM Ammonia	0.031	<0.1	U						
		WATER		Batch Number:		PB97569		Prep Date:	03/13/2017	
						0.1		03/13/2017	15:29	LB86346



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB86346 level: low sdg no.: LB86346
contract: lab code: CHEM case no.: LB86346 sas no.: LB86346
matrix: WATER sample id: I2173-01 client id: D-38MS
Percent Solids for Sample: 0 Spiked ID: I2173-01MS Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
ASTM Ammonia	mg/L	75 - 125	0.97		0.08	J	1	89		



GENCHEM
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: LB86346 level: low sdg no.: LB86346
contract: lab code: CHEM case no.: LB86346 sas no.: LB86346
matrix: WATER sample id: I2173-01 client id: D-38MSD
Percent Solids for Sample: 0 Spiked ID: I2173-01MSD Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
ASTM Ammonia	mg/L	75 - 125	0.98		0.08	J	1	90		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB86346 **Level:** LOW **SDG No.:** LB86346
Contract: **Lab Code:** CHEM **Case No.:** LB86346 **SAS No.:** LB86346
Matrix: WATER **Sample ID:** I2173-01 **Client ID:** D-38DUP
Percent Solids for Sample: 0 **Duplicate ID** I2173-01DUP **Percent Solids for Spike Sample:** 0

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
ASTM Ammonia	mg/L	20	0.08	J	0.081	J	1		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB86346 **Level:** LOW **SDG No.:** LB86346
Contract: **Lab Code:** CHEM **Case No.:** LB86346 **SAS No.:** LB86346
Matrix: WATER **Sample ID:** I2173-01MS **Client ID:** D-38MSD
Percent Solids for Sample: 0 **Duplicate ID** I2173-01MSD **Percent Solids for Spike Sample:** 0

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
ASTM Ammonia	mg/L	20	0.97		0.98		1		



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB86346

SDG No.: LB86346

Contract:

Lab Code: CHEM

Case No.: LB86346

SAS No.: LB86346

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB97569BS							
ASTM Ammonia	mg/L	1	0.95		95	80 - 120	