

CB 86425

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

3/17/2017 13:16

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.179	0.0	0.095	
ICB1	-0.018	0.0	0.004	
CCV1	245.016	0.0	0.238	
CCB1	-0.427	0.0	0.004	
PB97627BL	-0.256	0.0	0.004	
I2244-03	-0.463	0.0	0.004	
I2244-03DUP	-0.321	0.0	0.004	
CCV2	248.430	0.0	0.241	
CCB2	-0.180	0.0	0.004	

N	9
Mean	65.218
SD	107.5463
CV%	164.90

LB 86425

Aquakem v. 7.2AQ1

Results from time period:

Fri Mar 17 13:09:24 2017

AK

Fri Mar 17 13:09:33 2017

Sample Id	Sam/Ctr/c/	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.3514	µg/l	3/17/2017 9:40:54	
5.OPPBCN	A	Reactive CN	P	4.5246	µg/l	3/17/2017 9:40:55	
10PPBCN	A	Reactive CN	P	9.61	µg/l	3/17/2017 9:40:56	
50PPBCN	A	Reactive CN	P	49.5026	µg/l	3/17/2017 9:40:57	
100PPBCN	A	Reactive CN	P	100.4949	µg/l	3/17/2017 9:40:58	
250PPBCN	A	Reactive CN	P	252.5119	µg/l	3/17/2017 9:40:59	
500PPBCN	A	Reactive CN	P	498.7074	µg/l	3/17/2017 9:41:00	
ICV1	S	ReactiveCN	P	95.1788	µg/l	3/17/2017 13:09:24	
ICB1	S	ReactiveCN	P	-0.018	µg/l	3/17/2017 13:09:25	
CCV1	S	ReactiveCN	P	245.0164	µg/l	3/17/2017 13:09:26	
CCB1	S	ReactiveCN	P	-0.4273	µg/l	3/17/2017 13:09:27	
PB97627BL	S	ReactiveCN	P	-0.2558	µg/l	3/17/2017 13:09:28	
I2244-03	S	ReactiveCN	P	-0.4626	µg/l	3/17/2017 13:09:29	
I2244-03DUP	S	ReactiveCN	P	-0.3205	µg/l	3/17/2017 13:09:30	
CCV2	S	ReactiveCN	P	248.4297	µg/l	3/17/2017 13:09:31	
CCB2	S	ReactiveCN	P	-0.1802	µg/l	3/17/2017 13:09:32	

LB 86425

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AL

3/17/2017 9:49

Test Total CN


Accepted 3/17/2017 9:49

Factor 1048

Bias 0.004

Coeff. of det. 0.999956

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.3514	0.0000	
2	5.0PPBCN	0.009	4.5246	5.0000	
3	10PPBCN	0.013	9.6100	10.0000	
4	50PPBCN	0.051	49.5026	50.0000	
5	100PPBCN	0.100	100.4949	100.0000	
6	250PPBCN	0.245	252.5119	250.0000	
7	500PPBCN	0.480	498.7074	500.0000	