

Sample Results Report Print Date/Time: 2017/03/21 16:16:24

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
=====					
0.0PPM	8869			TOC	
0.5PPM	32419			TOC	
1.0PPM	63052			TOC	
2.0PPM.....	120171...	..	...	TOC	..
5.0PPM	280664			TOC	
10.0PPM	535837			TOC	
20.0PPM.....	1034598...	..	...	TOC	..
ICV	10.6158			TOC	
ICB	0.2279			TOC	
CCV.....	10.3383...	..	...	TOC	..
CCB	0.2033			TOC	
LB86451BLW	0.1240	0.0227	18.31	TOC	
LB86451BSW.....	9.7421...	0.9537..	9.79...	TOC	..
I2212-03	66.8940	3.9452	5.90	TOC	
I2212-03DLX10	8.8720	0.8609	9.70	TOC	
I2297-01.....	2.1519...	0.3371..	15.66...	TOC	..
I2297-01DUP	2.4812	0.1650	6.65	TOC	
I2297-02MS	11.9077	0.2208	1.85	TOC	
I2297-03MSD.....	12.0943...	0.1884..	1.56...	TOC	..
I2297-04	5.8042	0.8603	14.82	TOC	
I2297-05	2.7165	0.2817	10.37	TOC	
CCV.....	9.4439...	..	...	TOC	..
CCB	0.2607			TOC	
I2297-06	0.2083	0.0028	1.32	TOC	
I2297-08.....	5.6497...	0.3126..	5.53...	TOC	..
CCV	9.6957			TOC	
CCB	0.3168			TOC	

Method ID	Sample Type	Vial	Timestamp	Message
=====	=====	=====	=====	=====
TOC 0 - 20 ppmC	TOC Standard	1	2017/03/20 11:05	
TOC 0 - 20 ppmC	TOC Standard	2	2017/03/20 11:19	
TOC 0 - 20 ppmC	TOC Standard	3	2017/03/20 11:32	
TOC 0 - 20 ppmC	...TOC Standard	.. 4..	2017/03/20 11:45	..
TOC 0 - 20 ppmC	TOC Standard	5	2017/03/20 11:59	
TOC 0 - 20 ppmC	TOC Standard	6	2017/03/20 12:13	
TOC 0 - 20 ppmC	...TOC Standard	.. 7..	2017/03/20 12:27	..
TOC 0 - 20 ppmC	Sample	8	2017/03/20 12:42	
TOC 0 - 20 ppmC	Sample	9	2017/03/20 12:54	
TOC 0 - 20 ppmC	...Sample	.. 2..	2017/03/20 13:10	..
TOC 0 - 20 ppmC	Sample	3	2017/03/20 13:22	
TOC 0 - 20 ppmC	Sample	4	2017/03/20 13:39	
TOC 0 - 20 ppmC	...Sample	.. 5..	2017/03/20 13:56	..
TOC 0 - 20 ppmC	Sample	6	2017/03/20 14:17	
TOC 0 - 20 ppmC	Sample	7	2017/03/20 14:36	
TOC 0 - 20 ppmC	...Sample	.. 8..	2017/03/20 15:23	..
TOC 0 - 20 ppmC	Sample	9	2017/03/20 15:41	
TOC 0 - 20 ppmC	Sample	10	2017/03/20 15:59	
TOC 0 - 20 ppmC	...Sample	.. 11..	2017/03/20 16:18	..
TOC 0 - 20 ppmC	Sample	12	2017/03/20 16:36	
TOC 0 - 20 ppmC	Sample	13	2017/03/20 16:54	
TOC 0 - 20 ppmC	...Sample	.. 2..	2017/03/20 17:08	..
TOC 0 - 20 ppmC	Sample	3	2017/03/20 17:21	
TOC 0 - 20 ppmC	Sample	14	2017/03/20 17:37	
TOC 0 - 20 ppmC	...Sample	.. 15..	2017/03/20 17:55	..
TOC 0 - 20 ppmC	Sample	2	2017/03/20 18:09	
TOC 0 - 20 ppmC	Sample	3	2017/03/20 18:22	

Cal. Curve ID: TOC WATER 0-20PPM
Created: 2017/03/20 12:28
Calibration Factor (m): 1.027e+05
Y Intercept (b): 14049
r-squared: 0.99964

Standard ID	Y	X Expected	Measured	Message	Date & Time
	Raw Data	ug C	ug C		
0.0PPM	8869	0.000	-0.050		2017/03/20 11:05
0.5PPM	32419	0.250	0.179		2017/03/20 11:19
1.0PPM	63052	0.500	0.477		2017/03/20 11:32
2.0PPM	120171	1.000	1.033		2017/03/20 11:45
5.0PPM	280664	2.500	2.596		2017/03/20 11:59
10.0PPM	535837	5.000	5.080		2017/03/20 12:13
20.0PPM	1034598	10.000	9.936		2017/03/20 12:27