

m300.0
IC-2
AK 3-22-17

Clear table

ident	concentrati on F-	concentrati on CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2017-03-16_	3/16/17 13:51
STD2	0.2400	0.4300	0.3940	1.3890	0.3510	0.6790	2.0740	_2017-03-16_	3/16/17 14:22
STD3	0.4810	0.8130	0.7500	2.6280	0.6560	1.3630	3.8490	_2017-03-16_	3/16/17 14:59
STD4	0.8410	1.3110	1.2540	4.3200	1.0720	2.3230	6.3000	_2017-03-16_	3/16/17 15:30
STD5	1.9260	2.8310	2.8670	9.4750	2.3640	4.9930	13.8440	_2017-03-16_	3/16/17 16:02
STD6	3.7740	5.6290	5.7140	18.8990	4.7210	8.8390	27.8000	_2017-03-16_	3/16/17 16:34
STD7	4.9030	7.1200	7.2290	23.8130	5.9720	11.9580	37.4430	_2017-03-16_	3/16/17 17:06
STD8	10.1340	15.3150	15.2430	50.9770	12.7390	25.5950	75.9400	_2017-03-16_	3/16/17 17:37
ICV	4.9030	7.1450	7.3950	23.9630	6.0120	12.4810	35.2460	_2017-03-16_	3/16/17 19:13
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2017-03-16_	3/16/17 19:44
CCV	4.8370	6.9800	7.2410	23.6700	5.9000	12.1490	35.1950	_2017-03-20_	3/20/17 9:50
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2017-03-20_	3/20/17 10:28
LB86458BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2017-03-20_	3/20/17 11:01
LB86458BSW	4.8290	7.0450	7.3320	23.8430	5.9650	12.8370	35.1340	_2017-03-20_	3/20/17 11:34
I2301-01	0.0000	0.3560	0.0000	0.0000	0.0000	0.0000	0.0000	_2017-03-20_	3/20/17 12:40
I2301-01DUP	0.0000	0.3560	0.0000	0.0000	0.0000	0.0000	0.0000	_2017-03-20_	3/20/17 13:13
I2301-01MS	5.1000	7.3270	7.5400	24.4150	6.1140	12.4590	39.9370	_2017-03-20_	3/20/17 13:46
I2297-04	0.0000	11.8380	0.0000	0.0000	0.0000	0.0000	11.7930	_2017-03-20_	3/20/17 14:55
I2297-06	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2017-03-20_	3/20/17 15:28
I2297-08	0.0000	11.1520	0.0000	0.0000	0.0000	0.0000	11.5830	_2017-03-20_	3/20/17 16:01
I2263-01	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2017-03-20_	3/20/17 16:34
CCV	5.0250	7.1360	7.4000	23.9510	5.9830	13.0600	35.7260	_2017-03-20_	3/20/17 17:07
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2017-03-20_	3/20/17 17:40
I2263-02	0.3180	124.7830	0.0000	1.4280	10.1100	0.7930	87.3210	_2017-03-20_	3/20/17 18:13
I2263-03	0.2840	105.5260	0.0000	1.2300	12.5770	0.0000	83.0140	_2017-03-20_	3/20/17 18:46
I2263-04	0.0000	204.5460	0.0000	2.7870	10.2610	0.0000	145.3570	_2017-03-20_	3/20/17 19:19
I2263-05	0.0000	24.1770	0.0000	0.0000	0.9100	0.0000	7.5700	_2017-03-20_	3/20/17 19:52
I2263-05DUP	0.0000	24.1580	0.0000	0.0000	0.9060	0.0000	7.5480	_2017-03-20_	3/20/17 20:25
I2263-06MS	5.1570	26.5840	7.5730	24.6920	6.7530	14.6230	45.8710	_2017-03-20_	3/20/17 20:58

M300.0
IC-2
AP 3-22-17


I2263-07MSD	5.1690	26.5950	7.6040	24.7720	6.7740	14.6240	45.9890	3/20/17 21:30
I2263-08	0.2800	110.9850	0.0000	0.0000	11.3270	0.0000	85.8320	3/20/17 22:02
I2263-09	0.7190	107.3080	0.0000	1.7420	6.5760	0.9350	79.9010	3/20/17 22:34
I2263-10	0.3150	124.7820	0.0000	1.4420	10.0790	0.0000	88.0630	3/20/17 23:06
CCV	5.0740	7.2660	7.3990	24.0730	6.0130	13.0260	35.8930	3/20/17 23:38
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	3/21/17 0:10
LB86458BLW2	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	3/21/17 0:42
LB86458BSW2	5.1170	7.2190	7.4790	24.2230	6.0730	13.2620	35.8260	3/21/17 1:14
I2263-12	0.9480	156.7590	0.0000	2.2760	0.3310	1.2940	25.7250	3/21/17 1:46
I2263-13	1.1380	237.6950	0.0000	2.8070	0.0000	1.4430	31.6130	3/21/17 2:18
I2263-14	0.4800	146.6710	0.0000	3.1010	0.0000	174.0270	0.0000	3/21/17 2:50
I2263-15	0.5580	296.7490	0.0000	3.2760	0.0000	0.8060	57.5570	3/21/17 3:22
I2263-16	0.4770	315.0930	2.0630	6.2980	0.9220	0.0000	187.7390	3/21/17 3:54
I2263-17	1.2870	259.1040	0.0000	4.0630	3.6390	1.0680	168.2060	3/21/17 4:26
I2263-19	0.7360	353.8280	0.0000	7.2580	0.0000	1.3730	5.5980	3/21/17 4:58
I2263-20	0.6180	147.4480	0.0000	2.0800	0.0000	2.6800	23.5970	3/21/17 5:30
CCV	5.1000	7.4000	7.4660	24.3300	6.0650	13.1440	36.1360	3/21/17 6:02
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	3/21/17 6:34
I2263-21	0.0000	24.9810	0.0000	1.2040	0.3810	0.0000	14.1430	3/21/17 7:06
I2263-23	0.9960	141.6800	0.0000	2.0700	0.3020	0.0000	20.1730	3/21/17 7:38
I2263-24	0.0000	676.3160	0.0000	33.5390	0.0000	0.0000	330.1160	3/21/17 8:10
I2263-24DUP	0.0000	648.5590	0.0000	33.8180	0.0000	0.0000	344.2940	3/21/17 8:42
I2263-02DLX50	0.0000	5.8230	0.0000	0.0000	0.4170	0.0000	2.8750	3/21/17 9:45
I2325-01	0.0000	0.5130	0.0000	0.0000	0.3010	0.0000	0.0000	3/21/17 10:18
CCV	5.1880	7.4300	7.5240	24.4260	6.1090	12.4440	36.1070	3/21/17 10:49
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	3/21/17 11:22
LB86458BLW3	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	3/21/17 12:10
LB86458BSW3	5.1700	7.3760	7.5870	24.5500	6.1530	12.6160	36.0750	3/21/17 12:41
I2263-03DLX50	0.0000	4.1290	0.0000	0.0000	0.4590	0.0000	2.8300	3/21/17 13:12
I2263-04DLX10C	0.0000	6.9960	0.0000	0.0000	0.3390	0.0000	2.8210	3/21/17 13:46
I2263-05DLX2	0.0000	11.7250	0.0000	0.0000	0.5790	0.0000	4.5220	3/21/17 14:17
I2343-01	0.0000	0.5890	0.0000	0.0000	0.0000	0.0000	0.0000	3/21/17 14:48
I2343-01DUP	0.0000	0.5870	0.0000	0.0000	0.0000	0.0000	0.0000	3/21/17 15:19
I2343-01MS	4.9930	7.5380	7.5160	24.3740	6.0990	12.4810	39.4240	3/21/17 15:50
I2263-08DLX50	0.0000	4.2780	0.0000	0.0000	0.4390	0.0000	3.0200	3/21/17 16:21
I2263-09DLX50	0.0000	5.5750	0.0000	0.0000	0.3640	0.0000	2.9190	3/21/17 16:52
CCV	5.0990	7.1780	7.3790	23.9440	5.9890	12.3150	35.3010	3/21/17 17:23
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	3/21/17 17:54

W300.0
IC-2
AK 3-22-17

I2263-10DLX50	0.0000	5.7670	0.0000	0.0000	0.4210	0.0000	2.9120	2017-03-21	3/21/17 18:25
I2263-12DLX50	0.0000	7.6760	0.0000	0.0000	0.0000	0.0000	1.9470	2017-03-21	3/21/17 18:56
I2263-13DLX10C	0.0000	5.6900	0.0000	0.0000	0.0000	0.0000	1.7670	2017-03-21	3/21/17 19:27
I2263-14DLX50	0.0000	13.0480	0.0000	0.0000	0.0000	0.0000	5.0170	2017-03-21	3/21/17 19:58
I2263-15DLX10C	0.0000	8.5340	0.0000	0.0000	0.0000	0.0000	1.9770	2017-03-21	3/21/17 20:29
I2263-16DLX20C	0.0000	8.4430	0.0000	0.0000	0.0000	0.0000	2.3060	2017-03-21	3/21/17 21:00
I2263-17DLX10C	0.0000	10.3540	0.0000	0.0000	0.0000	0.0000	2.8140	2017-03-21	3/21/17 21:31
I2263-19DLX50C	0.0000	3.4120	0.0000	0.0000	0.0000	0.0000	0.0000	2017-03-21	3/21/17 22:02
I2263-20DLX50	0.0000	6.4330	0.0000	0.0000	0.0000	0.0000	1.8980	2017-03-21	3/21/17 22:33
I2263-21DLX2	0.0000	11.8480	0.0000	0.0000	0.3150	0.0000	7.7210	2017-03-21	3/21/17 23:04
CCV	5.1250	7.2560	7.4480	24.1790	6.0420	12.4530	35.6950	2017-03-21	3/21/17 23:35
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2017-03-22	3/22/17 0:06
LB86458BLW4	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2017-03-22	3/22/17 0:37
LB86458BSW4	5.1260	7.2760	7.5150	24.3050	6.0890	12.5850	35.6270	2017-03-22	3/22/17 1:08
I2263-23DLX50	0.0000	6.4130	0.0000	0.0000	0.0000	0.0000	1.8450	2017-03-22	3/22/17 1:39
I2263-24DLX10C	0.0000	9.1390	0.0000	0.0000	0.0000	0.0000	2.0660	2017-03-22	3/22/17 2:10
CCV	5.1540	7.2540	7.4920	24.2220	6.0740	12.5770	35.4960	2017-03-22	3/22/17 2:41
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2017-03-22	3/22/17 3:12

CALIBRATION OF COMPONENT F-

Method: AnionIC2-031617.mtw
 Equation: $Q = 0.0172477 \cdot A + 3.21771$
 RSD: 4.695 %
 Correlation coefficient: 0.999252



K3 = 0 K2 = 0 K1 = 0.0172477 K0 = 3.21771

Base: Area

Ref.channel: ch1

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	6.475	91.81	0.1	20	5.175		
3	20.58	371.3	0.4	20	5.175		
4	43.5	789	0.8	20	5.175		
5	113	2046	2	20	5.175		
6	221.5	4190	4	20	5.175		
7	287.8	5499	5	20	5.175		
8	632.3	1.157e+04	10	20	5.175		

CALIBRATION OF COMPONENT CL-

Method: AnionIC2-031617.mtw
 Equation: $Q = 0.0246174 \cdot A + 6.10341$
 RSD: 6.915 %
 Correlation coefficient: 0.998376



K3 = 0 K2 = 0 K1 = 0.0246174 K0 = 6.10341

Base: Area

Ref.channel: ch1

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	7.885	101.5	0.15	20	9.032		
3	30.88	412.5	0.6	20	9.032		
4	61.79	817.5	1.2	20	9.032		
5	157.4	2052	3	20	9.032		
6	338.3	4326	6	20	9.032		
7	430.1	5537	7.5	20	9.032		
8	962.4	1.219e+04	15	20	9.032		

CALIBRATION OF COMPONENT NO2

Method: AnionIC2-031617.mtw
 Equation: $Q = 0.0106527 \cdot A + 5.51398$
 RSD: 5.275 %
 Correlation coefficient: 0.999055



K3 = 0 K2 = 0 K1 = 0.0106527 K0 = 5.51398

Base: Area

Ref.channel: ch1

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	13.99	222.8	0.15	20	11.72		
3	55.93	890.3	0.6	20	11.72		
4	114.6	1837	1.2	20	11.72		
5	301.8	4864	3	20	11.72		
6	639.8	1.021e+04	6	20	11.72		
7	809.4	1.305e+04	7.5	20	11.72		
8	1646	2.81e+04	15	20	11.72		

CALIBRATION OF COMPONENT BR-

Method: AnionIC2-031617.mtw
 Equation: $Q = 0.0554211 \cdot A + 19.8449$
 RSD: 6.363 %
 Correlation coefficient: 0.998625



K3 = 0 K2 = 0 K1 = 0.0554211 K0 = 19.8449

Base: Area

Ref.channel: chl

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	6.891	143.1	0.5	20	15.11		
3	28.25	590.1	2	20	15.11		
4	57.3	1201	4	20	15.11		
5	146.7	3061	10	20	15.11		
6	310.5	6462	20	20	15.11		
7	393.8	8235	25	20	15.11		
8	867.6	1.804e+04	50	20	15.11		

CALIBRATION OF COMPONENT NO3

Method: AnionIC2-031617.mtw
 Equation: $Q = 0.00984921 \cdot A + 5.07064$
 RSD: 6.266 %
 Correlation coefficient: 0.998667



K3 = 0 K2 = 0 K1 = 0.00984921 K0 = 5.07064

Base: Area

Ref.channel: ch1

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	8.19	198.5	0.125	20	17.96		
3	33.37	818.1	0.5	20	17.96		
4	67.22	1661	1	20	17.96		
5	169.8	4286	2.5	20	17.96		
6	352.1	9072	5	20	17.96		
7	443.5	1.161e+04	6.25	20	17.96		
8	928.4	2.535e+04	12.5	20	17.96		

CALIBRATION OF COMPONENT HPO4

Method: AnionIC2-031617.mtw
 Equation: $Q = 0.0283193 \cdot A + 7.72924$
 RSD: 8.727 %
 Correlation coefficient: 0.997412




K3 = 0 K2 = 0 K1 = 0.0283193 K0 = 7.72924
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	3.863	206.4	0.25	20	23.35		
3	14.85	689.5	1	20	23.35		
4	33.4	1368	2	20	23.35		
5	83.92	3253	5	20	23.35		
6	146.2	5970	10	20	23.35		
7	208.9	8172	12.5	20	23.35		
8	454.9	1.78e+04	25	20	23.35		

CALIBRATION OF COMPONENT SO4

Method: AnionIC2-031617.mtw
 Equation: $Q = 0.0318159 \cdot A + 29.6515$
 RSD: 5.808 %
 Correlation coefficient: 0.998855



K3 = 0 K2 = 0 K1 = 0.0318159 K0 = 29.6515
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	10.01	371.7	0.75	20	25.64		
3	39.87	1488	3	20	25.64		
4	80.64	3029	6	20	25.64		
5	207.7	7771	15	20	25.64		
6	441.4	1.654e+04	30	20	25.64		
7	600.3	2.261e+04	37.5	20	25.64		
8	1212	4.681e+04	75	20	25.64		

Report date: 3/16/2017 6:13:07 PM
Printed by: wet


Ident: STD1
Analysis from: 3/16/2017 1:51:24 PM
File: _2017-03-16_

Last save: 3/16/2017 6:12:31 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 13959

Last save: 3/16/2017 1:17:1

SAMPLE:
: HM/AP
Vial number: 2
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 3/16/2017 6:13:17 PM
Printed by: wet


Ident: STD2
Analysis from: 3/16/2017 2:22:21 PM
File: _2017-03-16_

Last save: 3/16/2017 6:12:31 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 13960

Last save: 3/16/2017 1:17:1

SAMPLE:
: HM/AP
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.14	0.190	6.47	11.30	91.815	6.87
2	8.97	0.194	7.88	13.76	101.517	7.60
3	11.50	0.243	14.00	24.42	222.752	16.68
4	15.08	0.316	6.89	12.02	143.081	10.71
5	17.98	0.370	8.19	14.29	198.482	14.86
6	23.35	0.770	3.86	6.74	206.361	15.45
7	26.18	0.572	10.01	17.47	371.680	27.83
7	28.00	0.379	57.31	99.99	1335.688	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/16/2017 6:13:26 PM
Printed by: wet


Ident: STD3
Analysis from: 3/16/2017 2:59:11 PM
File: _2017-03-16_

Last save: 3/16/2017 6:12:31 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 13961

Last save: 3/16/2017 1:17:1

SAMPLE:
: HM/AP
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.17	0.233	20.58	9.20	371.251	7.06
2	8.96	0.196	30.87	13.80	412.543	7.84
3	11.49	0.240	55.93	25.00	890.303	16.93
4	15.05	0.313	28.25	12.63	590.141	11.22
5	17.93	0.370	33.37	14.92	818.147	15.56
6	23.32	0.673	14.85	6.64	689.451	13.11
7	26.19	0.573	39.87	17.82	1487.513	28.28
7	28.00	0.371	223.72	100.00	5259.349	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/16/2017 6:13:33 PM
Printed by: wet


Ident: STD4
Analysis from: 3/16/2017 3:30:10 PM
File: _2017-03-16_

Last save: 3/16/2017 6:12:31 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 13962

Last save: 3/16/2017 1:17:1

SAMPLE:
: HM/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.18	0.235	43.50	9.49	789.039	7.37
2	8.97	0.196	61.79	13.48	817.477	7.64
3	11.51	0.240	114.61	25.00	1836.726	17.16
4	15.04	0.313	57.30	12.50	1200.861	11.22
5	17.91	0.371	67.21	14.66	1661.060	15.52
6	23.30	0.601	33.40	7.29	1367.657	12.78
7	26.21	0.573	80.64	17.59	3028.529	28.30
7	29.00	0.361	458.45	100.00	10701.350	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/16/2017 6:13:41 PM
Printed by: wet


Ident: STD5
Analysis from: 3/16/2017 4:02:08 PM
File: _2017-03-16_

Last save: 3/16/2017 6:12:31 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 13963

Last save: 3/16/2017 1:17:1

SAMPLE:
: HM/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.16	0.221	112.99	9.57	2046.384	7.49
2	8.95	0.194	157.36	13.33	2051.686	7.51
3	11.50	0.240	301.77	25.57	4864.186	17.80
4	14.98	0.310	146.70	12.43	3061.347	11.20
5	17.81	0.374	169.82	14.39	4285.697	15.68
6	23.23	0.576	83.92	7.11	3253.177	11.90
7	26.15	0.570	207.67	17.60	7770.583	28.43
7	29.00	0.355	1180.24	100.00	27333.061	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/16/2017 6:13:53 PM
Printed by: wet


Ident: STD6
Analysis from: 3/16/2017 4:34:06 PM
File: _2017-03-16_

Last save: 3/16/2017 6:12:32 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 13964

Last save: 3/16/2017 1:17:1

SAMPLE:
: HM/AP
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.19	0.240	221.47	9.04	4189.835	7.38
2	8.95	0.190	338.27	13.81	4325.595	7.62
3	11.51	0.238	639.80	26.12	10209.251	17.98
4	14.93	0.307	310.54	12.68	6462.124	11.38
5	17.73	0.378	352.07	14.37	9071.578	15.98
6	23.21	0.604	146.25	5.97	5969.711	10.52
7	26.13	0.571	441.41	18.02	16543.403	29.14
7	29.00	0.361	2449.81	100.00	56771.497	100.00

Report date: 3/16/2017 6:14:01 PM
Printed by: wet


Ident: STD7
Analysis from: 3/16/2017 5:06:04 PM
File: _2017-03-16_

Last save: 3/16/2017 6:12:32 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 13965

Last save: 3/16/2017 1:17:1

SAMPLE:
: HM/AP
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.20	0.248	287.77	9.07	5499.278	7.36
2	8.96	0.191	430.14	13.55	5536.771	7.41
3	11.53	0.241	809.41	25.50	13053.916	17.47
4	14.93	0.308	393.83	12.41	8235.290	11.02
5	17.71	0.383	443.51	13.97	11612.347	15.54
6	23.20	0.586	208.93	6.58	8172.162	10.94
7	26.13	0.577	600.30	18.91	22605.193	30.26
7	28.50	0.362	3173.89	100.00	74714.956	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/16/2017 6:14:07 PM
Printed by: wet


Ident: STD8
Analysis from: 3/16/2017 5:37:32 PM
File: _2017-03-16_

Last save: 3/16/2017 6:12:36 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 13966

Last save: 3/16/2017 1:17:1

SAMPLE:
: HM/AP
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.20	0.236	632.30	9.43	11565.063	7.23
2	8.94	0.188	962.36	14.36	12194.823	7.63
3	11.56	0.259	1645.45	24.55	28100.452	17.58
4	14.85	0.306	867.60	12.94	18038.043	11.28
5	17.57	0.397	928.32	13.85	25352.392	15.86
6	23.06	0.594	454.93	6.79	17803.422	11.14
7	26.02	0.594	1212.19	18.08	46805.269	29.28
7	28.50	0.368	6703.16	100.00	159859.464	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/17/2017 9:51:19 AM
Printed by: wet


Ident: ICV
Analysis from: 3/16/2017 7:13:47 PM
File: _2017-03-16_

Last save: 3/16/2017 7:41:48 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 13969

Last save: 3/16/2017 6:39:2

SAMPLE:
: AK/AP
Vial number: 12
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.18	0.225	309.10	9.69	5482.274	7.40
2	8.94	0.191	432.78	13.57	5557.879	7.50
3	11.51	0.244	819.87	25.70	13366.261	18.03
4	14.90	0.309	396.64	12.43	8289.698	11.18
5	17.69	0.384	446.66	14.00	11693.863	15.78
6	23.16	0.582	220.22	6.90	8511.850	11.48
7	26.10	0.578	564.79	17.70	21224.268	28.63
7	28.00	0.359	3190.07	100.00	74126.094	100.00

Report date: 3/17/2017 9:51:28 AM
Printed by: wet


Ident: ICB
Analysis from: 3/16/2017 7:44:47 PM
File: _2017-03-16_

Last save: 3/16/2017 8:12:48 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 13970

Last save: 3/16/2017 6:39:2

SAMPLE:
: AK/AP
Vial number: 13
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 9:14:14 AM
Printed by: wet


Ident: CCV
Analysis from: 3/20/2017 9:50:19 AM
File: _2017-03-20_

Last save: 3/20/2017 10:25:20 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14021

Last save: 3/20/2017 9:47:3

SAMPLE: *AC*
: HM/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.25	0.258	266.59	9.45	5404.945	7.40
2	9.09	0.212	377.81	13.39	5423.964	7.43
3	11.78	0.265	736.62	26.11	13076.365	17.91
4	15.29	0.344	351.96	12.48	8183.732	11.21
5	18.18	0.429	394.94	14.00	11466.564	15.70
6	22.94	0.616	193.32	6.85	8275.350	11.33
7	27.06	0.635	499.81	17.72	21192.129	29.02
7	30.00	0.394	2821.05	100.00	73023.049	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 9:14:26 AM
Printed by: wet


Ident: CCB
Analysis from: 3/20/2017 10:28:35 AM
File: _2017-03-20_

Last save: 3/20/2017 10:58:35 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14022

Last save: 3/20/2017 10:25:

SAMPLE:
:
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 9:14:32 AM
Printed by: wet


Ident: LB86458BLW
Analysis from: 3/20/2017 11:01:33 AM
File: _2017-03-20_

Last save: 3/21/2017 9:13:50 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14023

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 4
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 9:14:37 AM
Printed by: wet


Ident: LB86458BSW
Analysis from: 3/20/2017 11:34:32 AM
File: _2017-03-20_

Last save: 3/21/2017 9:13:50 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14024

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.26	0.256	271.68	9.41	5396.236	7.30
2	9.09	0.210	387.20	13.42	5476.943	7.41
3	11.76	0.263	756.88	26.23	13247.693	17.93
4	15.25	0.341	359.41	12.45	8246.347	11.16
5	18.14	0.427	403.31	13.97	11597.681	15.70
6	22.97	0.605	209.66	7.26	8765.459	11.86
7	27.12	0.636	497.76	17.25	21153.832	28.63
7	30.00	0.391	2885.90	100.00	73884.192	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 9:16:54 AM
Printed by: wet


Ident: I2301-01
Analysis from: 3/20/2017 12:40:53 PM
File: _2017-03-20_

Last save: 3/21/2017 9:16:53 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14026

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.95	0.210	3.24	55.37	44.338	34.21
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	0.00	0.000	0.00	0.00	0.000	0.00
7	30.00	0.030	3.24	55.37	44.338	34.21

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 9:16:20 AM
Printed by: wet


Ident: I2301-01DUP
Analysis from: 3/20/2017 1:13:52 PM
File: _2017-03-20_

Last save: 3/21/2017 9:16:09 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14027

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.96	0.210	3.23	55.61	44.403	35.03
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	0.00	0.000	0.00	0.00	0.000	0.00
7	30.00	0.030	3.23	55.61	44.403	35.03

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 9:17:08 AM
Printed by: wet


Ident: I2301-01MS
Analysis from: 3/20/2017 1:46:50 PM
File: _2017-03-20_

Last save: 3/20/2017 2:16:51 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14028

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.20	0.265	279.83	9.19	5712.691	7.32
2	8.95	0.207	406.83	13.37	5705.541	7.31
3	11.53	0.259	787.35	25.87	13637.481	17.47
4	14.94	0.338	370.44	12.17	8452.633	10.83
5	17.74	0.422	416.59	13.69	11900.310	15.24
6	22.83	0.622	199.40	6.55	8495.932	10.88
7	26.01	0.618	583.30	19.16	24172.926	30.96
7	30.00	0.390	3043.75	100.00	78077.515	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 9:17:17 AM
Printed by: wet


Ident: I2297-04
Analysis from: 3/20/2017 2:55:34 PM
File: _2017-03-20_

Last save: 3/20/2017 3:25:34 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14030

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.96	0.219	640.43	76.21	9369.613	57.06
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	26.13	0.608	158.34	18.84	6481.186	39.47
7	30.00	0.118	798.78	95.06	15850.799	96.53

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 9:17:27 AM
Printed by: wet


Ident: I2297-06
Analysis from: 3/20/2017 3:28:32 PM
File: _2017-03-20_

Last save: 3/20/2017 3:58:32 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14031

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 10
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 9:17:49 AM
Printed by: wet


Ident: I2297-08
Analysis from: 3/20/2017 4:01:30 PM
File: _2017-03-20_

Last save: 3/20/2017 4:31:30 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14032

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.95	0.218	610.01	75.99	8812.532	56.14
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	26.13	0.610	154.19	19.21	6349.458	40.45
7	30.00	0.118	764.20	95.20	15161.990	96.58

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 9:19:50 AM
Printed by: wet


Ident: I2263-01
Analysis from: 3/20/2017 4:34:28 PM
File: _2017-03-20_

Last save: 3/20/2017 5:04:28 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14033

Last save: 3/20/2017 10:25:

SAMPLE:
:
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 9:19:57 AM
Printed by: wet


Ident: CCV
Analysis from: 3/20/2017 5:07:32 PM
File: _2017-03-20_

Last save: 3/20/2017 5:37:32 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14034

Last save: 3/20/2017 10:25:

SAMPLE: ~~AK~~
: HM/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.20	0.263	280.14	9.52	5625.280	7.51
2	8.97	0.208	397.35	13.51	5550.869	7.41
3	11.56	0.259	775.59	26.36	13376.147	17.85
4	14.98	0.342	361.42	12.28	8285.191	11.06
5	17.80	0.428	403.80	13.72	11634.827	15.53
6	22.84	0.631	204.71	6.96	8924.604	11.91
7	26.04	0.615	519.09	17.64	21525.799	28.73
7	30.00	0.392	2942.08	100.00	74922.717	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 9:20:04 AM
Printed by: wet


Ident: CCB
Analysis from: 3/20/2017 5:40:30 PM
File: _2017-03-20_

Last save: 3/20/2017 6:10:30 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14035

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 9:26:39 AM
Printed by: wet


Ident: I2263-02
Analysis from: 3/20/2017 6:13:33 PM
File: _2017-03-20_

Last save: 3/21/2017 9:25:43 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14036

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 13
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.15	0.319	5.57	0.14	126.115	0.07
2	9.29	0.784	1933.24	49.96	101092.806	57.59
3	0.00	0.000	0.00	0.00	0.000	0.00
4	15.19	0.344	6.83	0.18	157.141	0.09
5	17.78	0.437	672.40	17.38	20015.672	11.40
6	23.08	0.857	3.12	0.08	181.859	0.10
7	25.94	0.652	1248.41	32.26	53959.287	30.74
7	30.00	0.485	3869.57	100.00	175532.880	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 9:45:57 AM
Printed by: wet


Ident: I2263-03
Analysis from: 3/20/2017 6:46:33 PM
File: _2017-03-20_

Last save: 3/21/2017 9:45:33 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14037

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 14
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.17	0.316	3.87	0.10	86.269	0.05
2	9.13	0.655	1935.22	48.49	85454.480	52.78
3	0.00	0.000	0.00	0.00	0.000	0.00
4	15.24	0.339	3.86	0.10	85.836	0.05
5	17.78	0.438	842.00	21.10	25023.882	15.46
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.98	0.643	1205.87	30.22	51251.701	31.66
7	30.00	0.342	3990.83	100.00	161902.168	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 9:48:40 AM
Printed by: wet


Ident: I2263-04
Analysis from: 3/20/2017 7:19:30 PM
File: _2017-03-20_

Last save: 3/21/2017 9:48:21 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14038

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 15
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	9.36	1.372	1935.21	42.12	165869.090	59.82
3	0.00	0.000	0.00	0.00	0.000	0.00
4	15.18	0.334	29.90	0.65	647.687	0.23
5	17.80	0.430	696.07	15.15	20321.424	7.33
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.86	0.710	1932.83	42.07	90441.970	32.62
7	30.00	0.407	4594.00	99.99	277280.172	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 9:48:59 AM
Printed by: wet


Ident: I2263-05
Analysis from: 3/20/2017 7:52:29 PM
File: _2017-03-20_

Last save: 3/21/2017 9:46:26 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14039

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 16
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.98	0.215	1316.96	90.50	19390.122	78.98
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	18.08	0.425	46.97	3.23	1332.692	5.43
6	0.00	0.000	0.00	0.00	0.000	0.00
7	26.05	0.620	91.25	6.27	3826.526	15.59
7	30.00	0.180	1455.18	100.00	24549.339	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 2:23:01 PM
Printed by: wet


Ident: I2263-05DUP
Analysis from: 3/20/2017 8:25:26 PM
File: _2017-03-20_

Last save: 3/22/2017 2:20:18 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14040

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 17
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.98	0.216	1312.68	90.22	19374.719	78.74
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	18.08	0.426	46.75	3.21	1325.875	5.39
6	0.00	0.000	0.00	0.00	0.000	0.00
7	26.07	0.620	90.98	6.25	3813.006	15.50
7	30.00	0.180	1450.41	99.69	24513.600	99.63

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 9:49:16 AM
Printed by: wet


Ident: I2263-06MS
Analysis from: 3/20/2017 8:58:26 PM
File: _2017-03-20_

Last save: 3/21/2017 9:46:26 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14041

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 18
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.18	0.262	281.52	6.66	5778.799	5.75
2	8.96	0.213	1472.53	34.86	21344.624	21.24
3	11.58	0.266	773.36	18.31	13699.660	13.63
4	15.01	0.349	364.29	8.62	8552.455	8.51
5	17.83	0.439	444.11	10.51	13198.121	13.13
6	22.79	0.653	221.56	5.25	10038.529	9.99
7	25.94	0.626	666.66	15.78	27903.507	27.76
7	29.00	0.401	4224.03	100.00	100515.695	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 9:49:27 AM
Printed by: wet


Ident: I2263-07MSD
Analysis from: 3/20/2017 9:30:24 PM
File: _2017-03-20_

Last save: 3/21/2017 9:46:26 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14042

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 19
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.19	0.275	270.28	6.39	5793.336	5.75
2	8.96	0.213	1472.95	34.83	21353.192	21.20
3	11.58	0.262	785.23	18.57	13758.274	13.66
4	15.02	0.348	365.61	8.65	8581.512	8.52
5	17.83	0.439	445.60	10.54	13239.702	13.14
6	22.80	0.655	221.23	5.23	10039.343	9.97
7	25.96	0.627	668.01	15.80	27977.212	27.77
7	29.00	0.403	4228.89	100.00	100742.570	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 9:53:55 AM
Printed by: wet


Ident: I2263-08
Analysis from: 3/20/2017 10:02:24 PM
File: _2017-03-20_

Last save: 3/21/2017 9:46:26 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14043

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 20
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.17	0.315	3.69	0.09	81.668	0.05
2	9.04	0.701	1934.24	49.39	89887.383	54.32
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	17.80	0.440	748.28	19.11	22485.240	13.59
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.98	0.655	1230.01	31.41	53023.234	32.04
7	29.00	0.302	3916.23	100.00	165477.525	100.00

Report date: 3/21/2017 9:54:20 AM
Printed by: wet


Ident: I2263-09
Analysis from: 3/20/2017 10:34:22 PM
File: _2017-03-20_

Last save: 3/21/2017 9:54:19 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14044

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 21
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.14	0.289	26.39	0.74	593.821	0.40
2	9.22	0.651	1933.59	54.05	86901.481	57.86
3	0.00	0.000	0.00	0.00	0.000	0.00
4	15.16	0.348	11.96	0.33	270.418	0.18
5	17.83	0.432	439.97	12.30	12837.703	8.55
6	22.92	0.698	5.88	0.16	283.282	0.19
7	25.91	0.640	1159.59	32.41	49294.788	32.82
7	29.00	0.437	3577.38	100.00	150181.493	100.00

Report date: 3/21/2017 12:43:16 PM
Printed by: wet


Ident: I2263-10
Analysis from: 3/20/2017 11:06:20 PM
File: _2017-03-20_

Last save: 3/21/2017 9:46:26 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14045

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 22
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.16	0.317	5.48	0.14	122.071	0.07
2	9.21	0.781	1933.08	49.91	101091.873	57.52
3	0.00	0.000	0.00	0.00	0.000	0.00
4	15.20	0.347	7.02	0.18	162.195	0.09
5	17.79	0.440	665.94	17.19	19951.083	11.35
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.95	0.654	1261.43	32.57	54425.992	30.97
7	29.00	0.363	3872.95	100.00	175753.214	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 12:43:30 PM
Printed by: wet


Ident: CCV
Analysis from: 3/20/2017 11:38:25 PM
File: _2017-03-20_

Last save: 3/21/2017 12:07:26 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14046

Last save: 3/20/2017 8:56:2

SAMPLE: *WV*
: HM/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.20	0.270	273.09	9.49	5681.567	7.55
2	8.98	0.212	391.85	13.62	5656.456	7.52
3	11.58	0.264	752.43	26.15	13373.002	17.77
4	15.04	0.351	351.49	12.22	8329.207	11.07
5	17.87	0.439	393.32	13.67	11694.742	15.54
6	22.84	0.642	200.79	6.98	8900.576	11.83
7	26.03	0.626	514.36	17.88	21631.222	28.74
7	29.00	0.401	2877.34	100.00	75266.772	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 12:43:38 PM
Printed by: wet


Ident: CCB
Analysis from: 3/21/2017 12:10:24 AM
File: _2017-03-21_

Last save: 3/21/2017 12:39:24 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14047

Last save: 3/20/2017 8:56:2

SAMPLE:
: AK/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 12:43:44 PM
Printed by: wet


Ident: LB86458BLW2
Analysis from: 3/21/2017 12:42:22 AM
File: _2017-03-21_

Last save: 3/21/2017 1:11:22 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14048

Last save: 3/20/2017 8:56:2

SAMPLE:
: AK/AP
Vial number: 4
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 12:43:50 PM
Printed by: wet


Ident: LB86458BSW2
Analysis from: 3/21/2017 1:14:20 AM
File: _2017-03-21_

Last save: 3/21/2017 1:43:20 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14049

Last save: 3/20/2017 8:56:2

SAMPLE:
: AK/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.21	0.266	279.93	9.53	5732.159	7.57
2	8.98	0.211	395.56	13.46	5618.466	7.42
3	11.58	0.262	772.74	26.30	13524.117	17.86
4	15.02	0.348	358.73	12.21	8383.389	11.07
5	17.84	0.437	401.65	13.67	11816.153	15.60
6	22.84	0.629	209.07	7.11	9068.516	11.97
7	26.03	0.616	520.78	17.72	21588.535	28.51
7	29.00	0.396	2938.45	100.00	75731.334	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 12:44:37 PM
Printed by: wet


Ident: I2263-12
Analysis from: 3/21/2017 1:46:25 AM
File: _2017-03-21_

Last save: 3/21/2017 12:44:22 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14050

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 23
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.14	0.312	36.78	1.55	861.269	0.60
2	9.50	1.001	1932.01	81.34	127060.596	88.04
3	0.00	0.000	0.00	0.00	0.000	0.00
4	15.16	0.344	20.66	0.87	463.341	0.32
5	18.12	0.416	5.53	0.23	156.427	0.11
6	23.01	0.785	9.87	0.42	539.293	0.37
7	26.05	0.613	370.37	15.59	15239.094	10.56
7	29.00	0.496	2375.22	99.99	144320.020	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 12:47:21 PM
Printed by: wet


Ident: I2263-13
Analysis from: 3/21/2017 2:18:22 AM
File: _2017-03-21_

Last save: 3/21/2017 12:47:18 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14051

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 24
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.14	0.361	41.26	1.67	1083.387	0.51
2	9.75	1.609	1932.72	78.25	192789.534	90.04
3	0.00	0.000	0.00	0.00	0.000	0.00
4	15.22	0.338	28.65	1.16	654.865	0.31
5	0.00	0.000	0.00	0.00	0.000	0.00
6	23.11	0.868	10.88	0.44	645.159	0.30
7	26.09	0.622	456.18	18.47	18940.604	8.85
7	29.00	0.543	2469.69	99.99	214113.549	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 12:48:22 PM
Printed by: wet


Ident: I2263-14
Analysis from: 3/21/2017 2:50:20 AM
File: _2017-03-21_

Last save: 3/21/2017 12:48:21 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14052

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 25
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.09	0.311	15.78	0.40	314.794	0.13
2	9.09	0.905	1934.22	49.38	118868.362	48.78
3	0.00	0.000	0.00	0.00	0.000	0.00
4	15.04	0.369	30.43	0.78	761.070	0.31
5	0.00	0.000	0.00	0.00	0.000	0.00
6	25.50	0.972	1930.02	49.27	123641.683	50.74
7	0.00	0.000	0.00	0.00	0.000	0.00
7	29.00	0.365	3910.45	99.82	243585.908	99.95

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 12:49:01 PM
Printed by: wet


Ident: I2263-15
Analysis from: 3/21/2017 3:22:18 AM
File: _2017-03-21_

Last save: 3/21/2017 12:49:00 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14053

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 26
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.13	0.417	13.00	0.46	406.407	0.15
2	8.95	2.025	1938.33	68.97	240747.836	86.78
3	0.00	0.000	0.00	0.00	0.000	0.00
4	15.23	0.342	37.22	1.32	823.990	0.30
5	0.00	0.000	0.00	0.00	0.000	0.00
6	23.33	0.940	3.12	0.11	191.652	0.07
7	26.05	0.655	818.64	29.13	35249.503	12.71
7	29.00	0.625	2810.31	99.99	277419.389	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 2:54:07 PM
Printed by: wet


Ident: I2263-16
Analysis from: 3/21/2017 3:54:16 AM
File: _2017-03-21_

Last save: 3/21/2017 2:54:05 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14054

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 27
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.04	0.466	9.76	0.23	310.927	0.08
2	8.72	2.212	1933.60	45.48	255645.326	67.33
3	11.85	0.214	248.97	5.86	3355.301	0.88
4	15.13	0.330	89.28	2.10	1914.847	0.50
5	18.03	0.415	50.01	1.18	1357.910	0.36
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.78	0.927	1925.45	45.28	117083.687	30.84
7	29.00	0.652	4257.08	100.12	379667.997	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 12:51:43 PM
Printed by: wet


Ident: I2263-17
Analysis from: 3/21/2017 4:26:13 AM
File: _2017-03-21_

Last save: 3/21/2017 12:50:40 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14055

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 28
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.12	0.398	43.78	1.04	1257.529	0.39
2	8.85	1.741	1936.84	45.98	210176.255	64.75
3	0.00	0.000	0.00	0.00	0.000	0.00
4	15.15	0.340	49.47	1.17	1108.102	0.34
5	17.92	0.425	241.74	5.74	6875.135	2.12
6	23.01	0.739	7.58	0.18	378.355	0.12
7	25.76	0.820	1933.10	45.89	104804.879	32.29
7	29.00	0.638	4212.51	99.99	324600.255	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 12:58:05 PM
Printed by: wet

Ident: I2263-19
Analysis from: 3/21/2017 4:58:11 AM
File: _2017-03-21_

Last save: 3/21/2017 12:58:00 PM


Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14056

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 29
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000

*Before
Peak was not integrated
properly*

*ok
3/21/17*



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.05	0.502	17.96	0.93	613.872	0.75
2	8.72	0.729	1730.98	89.64	76003.997	92.62
3	11.25	0.002	-0.01	-0.00	-0.001	-0.00
4	15.17	0.324	107.86	5.59	2261.311	2.76
5	0.00	0.000	0.00	0.00	0.000	0.00
6	23.06	0.771	11.36	0.59	595.815	0.73
7	26.14	0.629	61.55	3.19	2587.304	3.15
7	29.00	0.422	1929.71	99.93	82062.301	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 12:59:28 PM
Printed by: wet

Ident: I2263-19
Analysis from: 3/21/2017 4:58:11 AM
File: _2017-03-21_

Last save: 3/21/2017 12:59:27 PM

Manual peaks!

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14056


Last save: 3/20/2017 10:25:

SAMPLE:

: AK/AP
Vial number: 29
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000

After

3/21/17



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.05	0.502	17.96	0.84	613.872	0.21
2	8.72	2.414	1941.64	91.09	287102.261	97.93
3	0.00	0.000	0.00	0.00	0.000	0.00
4	15.17	0.324	107.86	5.06	2261.311	0.77
5	0.00	0.000	0.00	0.00	0.000	0.00
6	23.06	0.771	11.36	0.53	595.815	0.20
7	26.14	0.629	61.55	2.89	2587.304	0.88
7	29.00	0.663	2140.37	100.41	293160.564	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 1:01:57 PM
Printed by: wet


Ident: I2263-20
Analysis from: 3/21/2017 5:30:09 AM
File: _2017-03-21_

Last save: 3/21/2017 9:46:27 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14057

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 30
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.16	0.327	16.35	0.70	476.547	0.35
2	9.20	0.936	1937.24	82.86	119499.380	88.00
3	0.00	0.000	0.00	0.00	0.000	0.00
4	15.20	0.350	16.01	0.68	392.542	0.29
5	0.00	0.000	0.00	0.00	0.000	0.00
6	23.02	0.737	29.39	1.26	1527.208	1.12
7	26.12	0.610	338.75	14.49	13901.509	10.24
7	29.00	0.423	2337.73	99.99	135797.186	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 1:02:42 PM
Printed by: wet


Ident: CCV
Analysis from: 3/21/2017 6:02:10 AM
File: _2017-03-21_

Last save: 3/21/2017 6:31:10 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14058

Last save: 3/20/2017 8:56:2

SAMPLE: *RY*
: HM/AP
Vial number: 36
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.21	0.265	281.22	9.53	5712.433	7.52
2	8.98	0.211	408.74	13.85	5765.315	7.59
3	11.58	0.262	772.81	26.19	13499.288	17.77
4	15.02	0.349	359.06	12.17	8421.903	11.09
5	17.85	0.439	399.64	13.55	11800.825	15.53
6	22.85	0.631	205.54	6.97	8984.351	11.83
7	26.06	0.617	523.41	17.74	21783.423	28.67
7	29.00	0.396	2950.41	100.00	75967.538	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 1:03:04 PM
Printed by: wet


Ident: CCB
Analysis from: 3/21/2017 6:34:07 AM
File: _2017-03-21_

Last save: 3/21/2017 7:03:07 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14059

Last save: 3/20/2017 8:56:2

SAMPLE:
: AK/AP
Vial number: 37
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 1:03:20 PM
Printed by: wet


Ident: I2263-21
Analysis from: 3/21/2017 7:06:08 AM
File: _2017-03-21_

Last save: 3/21/2017 9:46:27 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14060

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 31
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.97	0.216	1383.07	87.13	20042.967	70.73
3	0.00	0.000	0.00	0.00	0.000	0.00
4	15.17	0.360	3.23	0.20	76.313	0.27
5	18.08	0.423	9.08	0.57	258.500	0.91
6	0.00	0.000	0.00	0.00	0.000	0.00
7	26.09	0.610	191.96	12.09	7958.416	28.09
7	29.00	0.230	1587.34	100.00	28336.196	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 1:06:19 PM
Printed by: wet

Ident: I2263-24
Analysis from: 3/21/2017 8:10:04 AM
File: _2017-03-21_

Last save: 3/21/2017 1:06:17 PM


Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14062

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 33
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000

Before

peak was not integrated properly
3/21/17



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.38	1.865	1854.55	43.23	207782.455	48.76
3	0.00	0.000	0.00	0.00	0.000	0.00
4	15.66	0.361	522.17	12.17	11745.232	2.76
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	24.90	1.704	1917.18	44.69	206584.120	48.48
7	29.00	0.561	4293.90	100.09	426111.807	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 2:27:51 PM
Printed by: wet

Ident: I2263-24
Analysis from: 3/21/2017 8:10:04 AM
File: _2017-03-21_

Last save: 3/22/2017 2:27:48 PM

Manual peaks!


Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14062

Last save: 3/20/2017 10:25:

SAMPLE:

: AK/AP
Vial number: 33
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000

As per 3/22/17



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.38	4.820	1935.30	44.35	548997.850	71.55
3	0.00	0.000	0.00	0.00	0.000	0.00
4	15.66	0.361	522.17	11.97	11745.232	1.53
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	24.90	1.704	1917.18	43.94	206584.120	26.92
7	29.00	0.984	4374.65	100.26	767327.202	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 2:26:13 PM
Printed by: wet


Ident: I2263-24DUP
Analysis from: 3/21/2017 8:42:02 AM
File: _2017-03-21_

Last save: 3/22/2017 2:26:12 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14063

Last save: 3/20/2017 10:25:

SAMPLE:
: AK/AP
Vial number: 34
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.36	4.512	1956.34	44.24	526456.606	69.84
3	0.00	0.000	0.00	0.00	0.000	0.00
4	15.40	0.349	543.01	12.28	11846.078	1.57
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	24.45	1.756	1928.66	43.62	215496.487	28.59
7	29.00	0.945	4428.01	100.14	753799.171	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:58:54 AM
Printed by: wet


Ident: I2263-02DLX50
Analysis from: 3/21/2017 9:45:59 AM
File: _2017-03-21_

Last save: 3/22/2017 10:58:46 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14065

Last save: 3/20/2017 8:56:2

SAMPLE:
: AK/AP
Vial number: 38
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.90	0.207	325.06	90.57	4484.305	78.79
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	17.83	0.415	12.34	3.44	331.756	5.83
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.59	0.602	21.52	5.99	875.330	15.38
7	29.00	0.175	358.92	100.00	5691.391	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 1:09:58 PM
Printed by: wet


Ident: I2325-01
Analysis from: 3/21/2017 10:18:00 AM
File: _2017-03-21_

Last save: 3/21/2017 10:47:00 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14066

Last save: 3/20/2017 8:56:2

SAMPLE:
: AK/AP
Vial number: 35
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.91	0.215	12.28	77.31	172.361	63.95
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	17.86	0.417	3.60	22.69	97.154	36.05
6	0.00	0.000	0.00	0.00	0.000	0.00
7	0.00	0.000	0.00	0.00	0.000	0.00
7	29.00	0.090	15.88	100.00	269.515	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 1:10:06 PM
Printed by: wet


Ident: CCV
Analysis from: 3/21/2017 10:49:58 AM
File: _2017-03-21_

Last save: 3/21/2017 11:19:08 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14067

Last save: 3/20/2017 8:56:2

SAMPLE: *ATC*
: HM/AP
Vial number: 36
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.19	0.264	287.61	9.60	5815.697	7.66
2	8.92	0.210	414.11	13.82	5789.274	7.63
3	11.46	0.258	792.92	26.45	13608.959	17.93
4	14.84	0.347	365.57	12.20	8456.764	11.14
5	17.61	0.435	407.76	13.60	11890.867	15.67
6	23.16	0.659	190.35	6.35	8485.715	11.18
7	25.56	0.605	535.80	17.88	21765.461	28.68
7	29.00	0.397	2994.12	99.89	75812.739	99.89

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 1:10:14 PM
Printed by: wet


Ident: CCB
Analysis from: 3/21/2017 11:22:10 AM
File: _2017-03-21_

Last save: 3/21/2017 11:50:10 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14068

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 37
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 1:10:21 PM
Printed by: wet


Ident: LB86458BLW3
Analysis from: 3/21/2017 12:10:30 PM
File: _2017-03-21_

Last save: 3/21/2017 12:38:30 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14069

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 4
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 1:10:28 PM
Printed by: wet


Ident: LB86458BSW3
Analysis from: 3/21/2017 12:41:29 PM
File: _2017-03-21_

Last save: 3/21/2017 1:09:53 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14070

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.18	0.262	290.50	9.62	5793.993	7.60
2	8.91	0.208	413.82	13.71	5745.412	7.54
3	11.45	0.257	803.09	26.60	13725.817	18.02
4	14.83	0.345	368.62	12.21	8501.301	11.16
5	17.60	0.433	411.78	13.64	11979.989	15.72
6	23.16	0.660	192.77	6.39	8608.084	11.30
7	25.55	0.606	535.18	17.73	21745.620	28.54
7	28.00	0.396	3015.75	99.89	76100.216	99.89

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 5:17:35 PM
Printed by: wet


Ident: I2263-03DLX50
Analysis from: 3/21/2017 1:12:35 PM
File: _2017-03-21_

Last save: 3/21/2017 5:17:32 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14071

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 39
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.92	0.210	222.20	86.09	3108.939	71.08
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	17.87	0.417	15.29	5.92	417.790	9.55
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.65	0.606	20.62	7.99	847.042	19.37
7	28.00	0.176	258.10	100.00	4373.770	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 5:17:44 PM
Printed by: wet


Ident: I2263-04DLX100
Analysis from: 3/21/2017 1:46:05 PM
File: _2017-03-21_

Last save: 3/21/2017 2:14:05 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14072

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 40
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.94	0.207	393.93	93.60	5437.352	84.27
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	17.94	0.417	6.45	1.53	173.321	2.69
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.72	0.608	20.51	4.87	841.420	13.04
7	28.00	0.176	420.88	100.00	6452.093	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 5:17:51 PM
Printed by: wet


Ident: I2263-05DLX2
Analysis from: 3/21/2017 2:17:04 PM
File: _2017-03-21_

Last save: 3/21/2017 2:45:05 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14073

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 41
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.95	0.208	671.00	89.86	9277.682	77.48
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	17.92	0.418	24.27	3.25	661.130	5.52
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.73	0.605	46.73	6.26	1910.466	15.95
7	28.00	0.176	742.00	99.37	11849.277	98.95

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 5:17:59 PM
Printed by: wet


Ident: I2343-01
Analysis from: 3/21/2017 2:48:04 PM
File: _2017-03-21_

Last save: 3/21/2017 3:16:04 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14074

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 44
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.93	0.220	15.35	82.18	233.872	69.33
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	0.00	0.000	0.00	0.00	0.000	0.00
7	28.00	0.031	15.35	82.18	233.872	69.33

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 5:18:06 PM
Printed by: wet


Ident: I2343-01DUP
Analysis from: 3/21/2017 3:19:01 PM
File: _2017-03-21_

Last save: 3/21/2017 3:47:02 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14075

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 45
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.93	0.219	15.39	82.07	232.133	69.14
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	0.00	0.000	0.00	0.00	0.000	0.00
7	28.00	0.031	15.39	82.07	232.133	69.14

This report has been created by IC Net
METROHM LTD

Report date: 3/21/2017 5:18:12 PM
Printed by: wet


Ident: I2343-01MS
Analysis from: 3/21/2017 3:50:14 PM
File: _2017-03-21_

Last save: 3/21/2017 5:16:48 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14076

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 46
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.18	0.261	289.79	9.46	5587.634	7.18
2	8.91	0.208	423.89	13.84	5877.618	7.55
3	11.45	0.257	794.41	25.94	13592.643	17.47
4	14.83	0.345	365.27	11.93	8437.928	10.84
5	17.60	0.433	407.65	13.31	11870.192	15.25
6	23.14	0.655	192.08	6.27	8512.122	10.94
7	25.53	0.607	586.04	19.14	23850.381	30.65
7	28.00	0.395	3059.14	99.89	77728.517	99.89

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:59:53 AM
Printed by: wet


Ident: I2263-08DLX50
Analysis from: 3/21/2017 4:21:18 PM
File: _2017-03-21_

Last save: 3/22/2017 10:59:53 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14077

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 42
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.90	0.209	231.84	86.03	3229.807	70.63
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	17.81	0.415	13.99	5.19	376.809	8.24
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.60	0.603	23.64	8.77	966.382	21.13
7	28.00	0.175	269.47	100.00	4572.998	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:33:10 AM
Printed by: wet


Ident: I2263-09DLX50
Analysis from: 3/21/2017 4:52:16 PM
File: _2017-03-21_

Last save: 3/22/2017 10:33:09 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14078

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 43
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.90	0.209	308.29	91.01	4282.876	79.16
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	17.84	0.415	8.33	2.46	224.456	4.15
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.62	0.604	22.12	6.53	903.063	16.69
7	28.00	0.175	338.74	100.00	5410.395	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:32:32 AM
Printed by: wet


Ident: CCV
Analysis from: 3/21/2017 5:23:19 PM
File: _2017-03-21_

Last save: 3/21/2017 5:55:49 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14079

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 36
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.18	0.264	282.35	9.64	5710.696	7.69
2	8.90	0.210	399.09	13.62	5584.808	7.52
3	11.44	0.258	777.33	26.53	13336.268	17.95
4	14.82	0.346	358.20	12.22	8282.605	11.15
5	17.58	0.434	400.05	13.65	11647.395	15.68
6	23.16	0.663	187.05	6.38	8393.870	11.30
7	25.56	0.606	522.76	17.84	21258.876	28.61
7	28.00	0.397	2926.84	99.88	74214.519	99.88

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:32:39 AM
Printed by: wet


Ident: CCB
Analysis from: 3/21/2017 5:54:17 PM
File: _2017-03-21_

Last save: 3/21/2017 6:22:17 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14080

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 37
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:32:55 AM
Printed by: wet


Ident: I2263-10DLX50
Analysis from: 3/21/2017 6:25:23 PM
File: _2017-03-21_

Last save: 3/22/2017 10:32:54 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14081

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 47
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.90	0.209	319.29	90.26	4439.073	78.19
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	17.80	0.416	12.55	3.55	339.797	5.99
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.64	0.605	21.92	6.20	898.552	15.83
7	28.00	0.176	353.75	100.00	5677.422	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:34:16 AM
Printed by: wet


Ident: I2263-12DLX50
Analysis from: 3/21/2017 6:56:21 PM
File: _2017-03-21_

Last save: 3/22/2017 10:34:15 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14082

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 48
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.91	0.208	433.03	98.37	5989.686	95.35
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.67	0.605	7.15	1.62	292.185	4.65
7	28.00	0.116	440.18	100.00	6281.872	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:34:41 AM
Printed by: wet


Ident: I2263-13DLX100
Analysis from: 3/21/2017 7:27:19 PM
File: _2017-03-21_

Last save: 3/22/2017 10:34:40 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14083

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 49
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.90	0.209	314.10	98.61	4376.159	96.08
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.61	0.606	4.41	1.38	178.683	3.92
7	28.00	0.117	318.51	100.00	4554.842	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:34:55 AM
Printed by: wet


Ident: I2263-14DLX50
Analysis from: 3/21/2017 7:58:17 PM
File: _2017-03-21_

Last save: 3/22/2017 10:34:55 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14084

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 50
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.91	0.203	766.13	93.40	10351.720	82.33
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.58	0.602	54.14	6.60	2221.678	17.67
7	28.00	0.115	820.26	100.00	12573.398	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:35:14 AM
Printed by: wet


Ident: I2263-15DLX100
Analysis from: 3/21/2017 8:29:15 PM
File: _2017-03-21_

Last save: 3/22/2017 10:35:12 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14085

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 51
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.92	0.207	484.63	98.45	6685.941	95.55
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.60	0.604	7.64	1.55	311.069	4.45
7	28.00	0.116	492.27	100.00	6997.010	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:35:29 AM
Printed by: wet


Ident: I2263-16DLX200
Analysis from: 3/21/2017 9:00:13 PM
File: _2017-03-21_

Last save: 3/22/2017 10:35:28 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14086

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 52
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.91	0.208	476.86	97.41	6612.216	92.74
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.54	0.603	12.66	2.59	517.591	7.26
7	28.00	0.116	489.51	100.00	7129.807	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:35:46 AM
Printed by: wet


Ident: I2263-17DLX100
Analysis from: 3/21/2017 9:31:11 PM
File: _2017-03-21_

Last save: 3/22/2017 10:35:45 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14087

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 53
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.91	0.207	591.24	96.65	8164.129	90.70
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.53	0.602	20.50	3.35	837.074	9.30
7	28.00	0.116	611.73	100.00	9001.204	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:35:53 AM
Printed by: wet


Ident: I2263-19DLX500
Analysis from: 3/21/2017 10:02:09 PM
File: _2017-03-21_

Last save: 3/21/2017 10:30:09 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14088

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 54
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.93	0.213	177.83	100.00	2526.254	100.00
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	0.00	0.000	0.00	0.00	0.000	0.00
7	28.00	0.030	177.83	100.00	2526.254	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:36:10 AM
Printed by: wet


Ident: I2263-20DLX50
Analysis from: 3/21/2017 10:33:06 PM
File: _2017-03-21_

Last save: 3/22/2017 10:36:09 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14089

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 55
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.91	0.211	354.98	98.22	4980.243	95.02
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.52	0.603	6.43	1.78	260.830	4.98
7	28.00	0.116	361.40	100.00	5241.072	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:36:27 AM
Printed by: wet


Ident: I2263-21DLX2
Analysis from: 3/21/2017 11:04:05 PM
File: _2017-03-21_

Last save: 3/22/2017 10:36:26 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14090

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 56
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.91	0.209	671.97	86.46	9377.088	69.24
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	17.95	0.423	4.57	0.59	125.120	0.92
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.51	0.600	96.25	12.38	3921.304	28.96
7	28.00	0.176	772.79	99.43	13423.512	99.12

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:36:36 AM
Printed by: wet


Ident: CCV
Analysis from: 3/21/2017 11:35:11 PM
File: _2017-03-21_

Last save: 3/22/2017 9:01:24 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14091

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 36
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.17	0.265	282.93	9.63	5741.981	7.65
2	8.91	0.210	401.84	13.68	5648.544	7.53
3	11.47	0.261	777.10	26.45	13465.247	17.94
4	14.89	0.352	355.35	12.10	8367.581	11.15
5	17.68	0.443	395.41	13.46	11753.809	15.66
6	23.10	0.660	190.42	6.48	8491.812	11.31
7	25.46	0.602	531.42	18.09	21506.446	28.65
7	28.00	0.399	2934.47	99.89	74975.420	99.89

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:36:44 AM
Printed by: wet


Ident: CCB
Analysis from: 3/22/2017 12:06:09 AM
File: _2017-03-22_

Last save: 3/22/2017 12:34:09 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14092

Last save: 3/21/2017 11:19:

SAMPLE:
:
Vial number: 37
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:36:50 AM
Printed by: wet


Ident: LB86458BLW4
Analysis from: 3/22/2017 12:37:14 AM
File: _2017-03-22_

Last save: 3/22/2017 1:05:14 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14093

Last save: 3/21/2017 11:19:

SAMPLE:
:
Vial number: 4
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:37:07 AM
Printed by: wet


Ident: LB86458BSW4
Analysis from: 3/22/2017 1:08:12 AM
File: _2017-03-22_

Last save: 3/22/2017 10:37:06 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14094

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.18	0.265	284.54	9.64	5742.385	7.62
2	8.92	0.210	403.29	13.67	5664.658	7.51
3	11.47	0.260	784.31	26.58	13592.366	18.03
4	14.88	0.352	356.81	12.09	8412.947	11.16
5	17.67	0.444	397.64	13.48	11848.927	15.72
6	23.12	0.658	192.89	6.54	8585.735	11.39
7	25.49	0.603	527.84	17.89	21463.566	28.47
7	28.00	0.399	2947.33	99.89	75310.584	99.88

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:37:21 AM
Printed by: wet


Ident: I2263-23DLX50
Analysis from: 3/22/2017 1:39:15 AM
File: _2017-03-22_

Last save: 3/22/2017 10:37:20 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14095

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 57
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.91	0.211	353.67	98.45	4963.867	95.61
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.57	0.607	5.55	1.55	227.777	4.39
7	28.00	0.117	359.22	100.00	5191.643	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:37:37 AM
Printed by: wet


Ident: I2263-24DLX1000
Analysis from: 3/22/2017 2:10:13 AM
File: _2017-03-22_

Last save: 3/22/2017 10:37:36 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14096

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 58
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.92	0.208	518.60	98.30	7177.600	95.13
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.58	0.603	8.98	1.70	367.057	4.87
7	28.00	0.116	527.58	100.00	7544.657	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:37:43 AM
Printed by: wet


Ident: CCV
Analysis from: 3/22/2017 2:41:18 AM
File: _2017-03-22_

Last save: 3/22/2017 9:01:00 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14097

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.18	0.266	283.21	9.63	5775.192	7.68
2	8.92	0.211	400.81	13.63	5646.295	7.51
3	11.46	0.261	781.67	26.58	13548.926	18.01
4	14.87	0.352	356.15	12.11	8383.081	11.14
5	17.65	0.443	397.27	13.51	11819.665	15.71
6	23.12	0.658	192.63	6.55	8580.167	11.41
7	25.51	0.602	526.30	17.89	21381.241	28.42
7	28.00	0.399	2938.04	99.89	75134.566	99.88

This report has been created by IC Net
METROHM LTD

Report date: 3/22/2017 10:37:52 AM
Printed by: wet


Ident: CCB
Analysis from: 3/22/2017 3:12:16 AM
File: _2017-03-22_

Last save: 3/22/2017 3:40:17 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14098

Last save: 3/21/2017 11:19:

SAMPLE:
: AK/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD