

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
=====	=====	=====	=====	=====	=====
CCV	10.5572			TOC	
CCB	0.2669			TOC	
LB86464BLW	0.2024	0.0482	23.79	TOC	
LB86464BSW.....	9.5598...	0.4742..	4.96...	TOC	..
I2212-04	66.2124	1.2982	1.96	TOC	
I2212-04DLX10	8.4813	0.1830	2.16	TOC	
I2296-05.....	0.3802...	0.0467..	12.30...	TOC	..
I2296-06	1.1422	0.0752	6.58	TOC	
CCV	9.3178			TOC	
CCB.....	0.3068...	..	...	TOC	..

Method ID	Sample Type	Vial	Timestamp	Message
TOC 0 - 20 ppmC	Sample	2	2017/03/21 11:23	
TOC 0 - 20 ppmC	Sample	3	2017/03/21 11:44	
TOC 0 - 20 ppmC	Sample	4	2017/03/21 12:09	
TOC 0 - 20 ppmC	...Sample	.. 5..	2017/03/21 12:36	..
TOC 0 - 20 ppmC	Sample	6	2017/03/21 13:11	Max Integration Time Reached
TOC 0 - 20 ppmC	Sample	7	2017/03/21 13:38	
TOC 0 - 20 ppmC	...Sample	.. 8..	2017/03/21 14:02	..
TOC 0 - 20 ppmC	Sample	9	2017/03/21 14:26	
TOC 0 - 20 ppmC	Sample	2	2017/03/21 14:40	
TOC 0 - 20 ppmC	...Sample	.. 3..	2017/03/21 14:53	..

Cal. Curve ID: TOC WATER 0-20PPM
Created: 2017/03/20 12:28
Calibration Factor (m): 1.027e+05
Y Intercept (b): 14049
r-squared: 0.99964

Standard ID	Y Raw Data	X Expected ug C	Measured ug C	Message	Date & Time
0.0PPM	8869	0.000	-0.050		2017/03/20 11:05
0.5PPM	32419	0.250	0.179		2017/03/20 11:19
1.0PPM	63052	0.500	0.477		2017/03/20 11:32
2.0PPM	120171	1.000	1.033		2017/03/20 11:45
5.0PPM	280664	2.500	2.596		2017/03/20 11:59
10.0PPM	535837	5.000	5.080		2017/03/20 12:13
20.0PPM	1034598	10.000	9.936		2017/03/20 12:27

3-22-17
AK

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
0.0PPM	8869			TOC	
0.5PPM	32419			TOC	
1.0PPM	63052			TOC	
2.0PPM.....	120171...	..	...	TOC	..
5.0PPM	280664			TOC	
10.0PPM	535837			TOC	
20.0PPM.....	1034598...	..	...	TOC	..
ICV	10.6158			TOC	
ICB	0.2279			TOC	

3-22-17
AK

Sample Results Report Print Date/Time: 2017/03/21 16:26:44

Method ID	Sample Type	Vial	Timestamp	Message
TOC 0 - 20 ppmC	TOC Standard	1	2017/03/20 11:05	
TOC 0 - 20 ppmC	TOC Standard	2	2017/03/20 11:19	
TOC 0 - 20 ppmC	TOC Standard	3	2017/03/20 11:32	
TOC 0 - 20 ppmC	...TOC Standard	.. 4..	2017/03/20 11:45	..
TOC 0 - 20 ppmC	TOC Standard	5	2017/03/20 11:59	
TOC 0 - 20 ppmC	TOC Standard	6	2017/03/20 12:13	
TOC 0 - 20 ppmC	...TOC Standard	.. 7..	2017/03/20 12:27	..
TOC 0 - 20 ppmC	Sample	8	2017/03/20 12:42	
TOC 0 - 20 ppmC	Sample	9	2017/03/20 12:54	