

LB86492



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908-789-8900 Fax: 908-789-892

Analysis Method: 9060 TOC
Parameter: TOC
Run Number: LB86492
Instrument: Appolo-9000

**Analytical Summary Report**

Analysis Method: 9060 TOC
Parameter: TOC
Run Number: LB86492
Instrument: Appolo-9000

Seq	Lab ID	Client ID	Sample Type	Inst Conc. ppm/ppb	Dil	Analytical Date
1	0.0PPM	0.0PPM	CAL	8869.000	1	3/20/17
2	0.5PPM	0.5PPM	CAL	32419.000	1	3/20/17
3	1.0PPM	1.0PPM	CAL	63052.000	1	3/20/17
4	2.0PPM	2.0PPM	CAL	120171.000	1	3/20/17
5	5.0PPM	5.0PPM	CAL	280664.000	1	3/20/17
6	10.0PPM	10.0PPM	CAL	535837.000	1	3/20/17
7	20.0PPM	20.0PPM	CAL	1034598.000	1	3/20/17
8	ICV	ICV	ICV	10.616	1	3/20/17
9	ICB	ICB	ICB	0.000	1	3/20/17
10	CCV	CCV	CCV	10.105	1	3/22/17
11	CCB	CCB	CCB	0.000	1	3/22/17
12	LB86492BLW	LB86492BLW	MB	0.000	1	3/22/17
13	LB86492BSW	LB86492BSW	LCS	9.750	1	3/22/17
14	I2354-01	EFF-170321-03	SAM	4.406	1	3/22/17
15	I2354-01S	EFF-170321-03S	MS	13.638	1	3/22/17
16	I2354-01SD	EFF-170321-03SD	MSD	14.341	1	3/22/17
17	CCV	CCV	CCV	9.845	1	3/22/17
18	CCB	CCB	CCB	0.000	1	3/22/17

Chemtech Consulting Group

Analytical Review Report

Date Printed :
Analyst :
Data File :

3/23/17
AK
LB86492.MDB

Approved By :
Approved Date :
Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units
											Line 2
TOC											
0.0PPM	0.0PPM				W	3/20/17					
TOC		PASS	8869.000		8870						mg/L
0.5PPM	0.5PPM				W	3/20/17					
TOC		PASS	32419.000		32400						mg/L
1.0PPM	1.0PPM				W	3/20/17					
TOC		PASS	63052.000		63100						mg/L
2.0PPM	2.0PPM				W	3/20/17					
TOC		PASS	120171.000		120000						mg/L
5.0PPM	5.0PPM				W	3/20/17					
TOC		PASS	280664.000		281000						mg/L
10.0PPM	10.0PPM				W	3/20/17					
TOC		PASS	535837.000		536000						mg/L
20.0PPM	20.0PPM				W	3/20/17					
TOC		PASS	034598.000		1030000						mg/L
ICV1	ICV1				W	3/20/17					
TOC		PASS	10.616		10.60	106.0	90	110			mg/L
ICB1	ICB1				W	3/20/17					
TOC		PASS	0.000		0.000			+/-0.0970			mg/L
CCV1	CCV1				W	3/22/17					
TOC		PASS	10.105		10.10	101.0	90	110			mg/L
CCB1	CCB1				W	3/22/17					
TOC		PASS	0.000		0.000			+/-0.0970			mg/L
LB86492BLW	LB86492BLW				W	3/22/17					
TOC		PASS	0.000		0.000			+/-0.0970			mg/L
LB86492BSW	LB86492BSW				W	3/22/17					
TOC		PASS	9.750		9.75	98.0	90	110			mg/L
I2354-01	EFF-170321-03			1	W	3/22/17					
TOC		PASS	4.406		4.41						mg/L
I2354-01S	EFF-170321-03S			1	W	3/22/17					
TOC		PASS	13.638		13.60	92.0	75	125			mg/L
I2354-01SD	EFF-170321-03SD			1	W	3/22/17					
TOC		PASS	14.341		14.30	99.0	75	125	5.0	20	mg/L
CCV2	CCV2				W	3/22/17					
TOC		PASS	9.845		9.84	98.0	90	110			mg/L
CCB2	CCB2				W	3/22/17					
TOC		PASS	0.000		0.000			+/-0.0970			mg/L