


Clear table

ident	concentrati on F-	concentrati on CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2017-03-16_	3/16/17 13:51
STD2	0.2400	0.4300	0.3940	1.3890	0.3510	0.6790	2.0740	_2017-03-16_	3/16/17 14:22
STD3	0.4810	0.8130	0.7500	2.6280	0.6560	1.3630	3.8490	_2017-03-16_	3/16/17 14:59
STD4	0.8410	1.3110	1.2540	4.3200	1.0720	2.3230	6.3000	_2017-03-16_	3/16/17 15:30
STD5	1.9260	2.8310	2.8670	9.4750	2.3640	4.9930	13.8440	_2017-03-16_	3/16/17 16:02
STD6	3.7740	5.6290	5.7140	18.8990	4.7210	8.8390	27.8000	_2017-03-16_	3/16/17 16:34
STD7	4.9030	7.1200	7.2290	23.8130	5.9720	11.9580	37.4430	_2017-03-16_	3/16/17 17:06
STD8	10.1340	15.3150	15.2430	50.9770	12.7390	25.5950	75.9400	_2017-03-16_	3/16/17 17:37
ICV	4.9030	7.1450	7.3950	23.9630	6.0120	12.4810	35.2460	_2017-03-16_	3/16/17 19:13
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2017-03-16_	3/16/17 19:44
CCV	5.0320	7.2280	7.4490	24.0940	6.0240	13.0800	35.7930	_2017-03-29_	3/29/17 9:18
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2017-03-29_	3/29/17 9:49
LB86592BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2017-03-29_	3/29/17 10:19
LB86592BSW	5.0370	7.2830	7.4960	24.2470	6.0670	13.2180	36.3280	_2017-03-29_	3/29/17 10:50
I2437-02	0.4480	115.2110	0.0000	1.5620	0.8240	0.0000	186.2070	_2017-03-29_	3/29/17 11:21
I2437-01	0.5980	337.8840	0.0000	4.2710	0.0000	0.0000	368.0620	_2017-03-29_	3/29/17 11:52
I2437-03	0.0000	330.9630	0.0000	0.0000	0.0000	0.0000	415.5220	_2017-03-29_	3/29/17 12:22
I2437-02DUP	0.5010	116.9490	0.0000	1.5360	0.8210	0.0000	187.6670	_2017-03-29_	3/29/17 12:53
I2437-02MS	5.7270	91.8240	7.8000	25.8650	6.8060	0.0000	175.5210	_2017-03-29_	3/29/17 13:23
I2437-01DLX50	0.0000	21.6320	0.0000	0.0000	0.0000	0.0000	20.1940	_2017-03-29_	3/29/17 13:54
I2437-03DLX50	0.0000	17.2300	0.0000	0.0000	0.0000	0.0000	27.3630	_2017-03-29_	3/29/17 14:24
I2437-02DLX10	0.0000	25.3010	0.0000	0.0000	0.0000	0.0000	23.7190	_2017-03-29_	3/29/17 14:55
CCV	5.0700	7.3310	7.5040	24.2740	6.0780	13.3670	36.2790	_2017-03-29_	3/29/17 15:25
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2017-03-29_	3/29/17 16:02

m300.0
JC-2
AK 3-30-17-

CALIBRATION OF COMPONENT F-

Method: AnionIC2-031617.mtw
 Equation: $Q = 0.0172477 \cdot A + 3.21771$
 RSD: 4.695 %
 Correlation coefficient: 0.999252



K3 = 0 K2 = 0 K1 = 0.0172477 K0 = 3.21771

Base: Area

Ref.channel: ch1

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	6.475	91.81	0.1	20	5.175		
3	20.58	371.3	0.4	20	5.175		
4	43.5	789	0.8	20	5.175		
5	113	2046	2	20	5.175		
6	221.5	4190	4	20	5.175		
7	287.8	5499	5	20	5.175		
8	632.3	1.157e+04	10	20	5.175		

CALIBRATION OF COMPONENT CL-

Method: AnionIC2-031617.mtw
 Equation: $Q = 0.0246174 \cdot A + 6.10341$
 RSD: 6.915 %
 Correlation coefficient: 0.998376



K3 = 0 K2 = 0 K1 = 0.0246174 K0 = 6.10341

Base: Area

Ref.channel: ch1

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	7.885	101.5	0.15	20	9.032		
3	30.88	412.5	0.6	20	9.032		
4	61.79	817.5	1.2	20	9.032		
5	157.4	2052	3	20	9.032		
6	338.3	4326	6	20	9.032		
7	430.1	5537	7.5	20	9.032		
8	962.4	1.219e+04	15	20	9.032		

CALIBRATION OF COMPONENT NO2

Method: AnionIC2-031617.mtw
 Equation: $Q = 0.0106527 \cdot A + 5.51398$
 RSD: 5.275 %
 Correlation coefficient: 0.999055



K3 = 0 K2 = 0 K1 = 0.0106527 K0 = 5.51398

Base: Area

Ref.channel: ch1

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	13.99	222.8	0.15	20	11.72		
3	55.93	890.3	0.6	20	11.72		
4	114.6	1837	1.2	20	11.72		
5	301.8	4864	3	20	11.72		
6	639.8	1.021e+04	6	20	11.72		
7	809.4	1.305e+04	7.5	20	11.72		
8	1646	2.81e+04	15	20	11.72		

CALIBRATION OF COMPONENT BR-

Method: AnionIC2-031617.mtw
 Equation: $Q = 0.0554211 \cdot A + 19.8449$
 RSD: 6.363 %
 Correlation coefficient: 0.998625



K3 = 0 K2 = 0 K1 = 0.0554211 K0 = 19.8449

Base: Area

Ref.channel: chl

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	6.891	143.1	0.5	20	15.11		
3	28.25	590.1	2	20	15.11		
4	57.3	1201	4	20	15.11		
5	146.7	3061	10	20	15.11		
6	310.5	6462	20	20	15.11		
7	393.8	8235	25	20	15.11		
8	867.6	1.804e+04	50	20	15.11		

CALIBRATION OF COMPONENT NO3

Method: AnionIC2-031617.mtw
 Equation: $Q = 0.00984921 \cdot A + 5.07064$
 RSD: 6.266 %
 Correlation coefficient: 0.998667




K3 = 0 K2 = 0 K1 = 0.00984921 K0 = 5.07064

Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	8.19	198.5	0.125	20	17.96		
3	33.37	818.1	0.5	20	17.96		
4	67.22	1661	1	20	17.96		
5	169.8	4286	2.5	20	17.96		
6	352.1	9072	5	20	17.96		
7	443.5	1.161e+04	6.25	20	17.96		
8	928.4	2.535e+04	12.5	20	17.96		

CALIBRATION OF COMPONENT HPO4

Method: AnionIC2-031617.mtw
 Equation: $Q = 0.0283193 \cdot A + 7.72924$
 RSD: 8.727 %
 Correlation coefficient: 0.997412




K3 = 0 K2 = 0 K1 = 0.0283193 K0 = 7.72924
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	3.863	206.4	0.25	20	23.35		
3	14.85	689.5	1	20	23.35		
4	33.4	1368	2	20	23.35		
5	83.92	3253	5	20	23.35		
6	146.2	5970	10	20	23.35		
7	208.9	8172	12.5	20	23.35		
8	454.9	1.78e+04	25	20	23.35		

CALIBRATION OF COMPONENT SO4

Method: AnionIC2-031617.mtw
 Equation: $Q = 0.0318159 \cdot A + 29.6515$
 RSD: 5.808 %
 Correlation coefficient: 0.998855



K3 = 0 K2 = 0 K1 = 0.0318159 K0 = 29.6515
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	10.01	371.7	0.75	20	25.64		
3	39.87	1488	3	20	25.64		
4	80.64	3029	6	20	25.64		
5	207.7	7771	15	20	25.64		
6	441.4	1.654e+04	30	20	25.64		
7	600.3	2.261e+04	37.5	20	25.64		
8	1212	4.681e+04	75	20	25.64		

Report date: 3/16/2017 6:13:07 PM
Printed by: wet


Ident: STD1
Analysis from: 3/16/2017 1:51:24 PM
File: _2017-03-16_

Last save: 3/16/2017 6:12:31 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 13959

Last save: 3/16/2017 1:17:1

SAMPLE:
: HM/AP
Vial number: 2
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 3/16/2017 6:13:17 PM
Printed by: wet


Ident: STD2
Analysis from: 3/16/2017 2:22:21 PM
File: _2017-03-16_

Last save: 3/16/2017 6:12:31 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 13960

Last save: 3/16/2017 1:17:1

SAMPLE:
: HM/AP
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.14	0.190	6.47	11.30	91.815	6.87
2	8.97	0.194	7.88	13.76	101.517	7.60
3	11.50	0.243	14.00	24.42	222.752	16.68
4	15.08	0.316	6.89	12.02	143.081	10.71
5	17.98	0.370	8.19	14.29	198.482	14.86
6	23.35	0.770	3.86	6.74	206.361	15.45
7	26.18	0.572	10.01	17.47	371.680	27.83
7	28.00	0.379	57.31	99.99	1335.688	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/16/2017 6:13:26 PM
Printed by: wet


Ident: STD3
Analysis from: 3/16/2017 2:59:11 PM
File: _2017-03-16_

Last save: 3/16/2017 6:12:31 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 13961

Last save: 3/16/2017 1:17:1

SAMPLE:
: HM/AP
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.17	0.233	20.58	9.20	371.251	7.06
2	8.96	0.196	30.87	13.80	412.543	7.84
3	11.49	0.240	55.93	25.00	890.303	16.93
4	15.05	0.313	28.25	12.63	590.141	11.22
5	17.93	0.370	33.37	14.92	818.147	15.56
6	23.32	0.673	14.85	6.64	689.451	13.11
7	26.19	0.573	39.87	17.82	1487.513	28.28
7	28.00	0.371	223.72	100.00	5259.349	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/16/2017 6:13:33 PM
Printed by: wet


Ident: STD4
Analysis from: 3/16/2017 3:30:10 PM
File: _2017-03-16_

Last save: 3/16/2017 6:12:31 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 13962

Last save: 3/16/2017 1:17:1

SAMPLE:
: HM/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.18	0.235	43.50	9.49	789.039	7.37
2	8.97	0.196	61.79	13.48	817.477	7.64
3	11.51	0.240	114.61	25.00	1836.726	17.16
4	15.04	0.313	57.30	12.50	1200.861	11.22
5	17.91	0.371	67.21	14.66	1661.060	15.52
6	23.30	0.601	33.40	7.29	1367.657	12.78
7	26.21	0.573	80.64	17.59	3028.529	28.30
7	29.00	0.361	458.45	100.00	10701.350	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/16/2017 6:13:41 PM
Printed by: wet


Ident: STD5
Analysis from: 3/16/2017 4:02:08 PM
File: _2017-03-16_

Last save: 3/16/2017 6:12:31 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 13963

Last save: 3/16/2017 1:17:1

SAMPLE:
: HM/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.16	0.221	112.99	9.57	2046.384	7.49
2	8.95	0.194	157.36	13.33	2051.686	7.51
3	11.50	0.240	301.77	25.57	4864.186	17.80
4	14.98	0.310	146.70	12.43	3061.347	11.20
5	17.81	0.374	169.82	14.39	4285.697	15.68
6	23.23	0.576	83.92	7.11	3253.177	11.90
7	26.15	0.570	207.67	17.60	7770.583	28.43
7	29.00	0.355	1180.24	100.00	27333.061	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/16/2017 6:13:53 PM
Printed by: wet


Ident: STD6
Analysis from: 3/16/2017 4:34:06 PM
File: _2017-03-16_

Last save: 3/16/2017 6:12:32 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 13964

Last save: 3/16/2017 1:17:1

SAMPLE:
: HM/AP
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.19	0.240	221.47	9.04	4189.835	7.38
2	8.95	0.190	338.27	13.81	4325.595	7.62
3	11.51	0.238	639.80	26.12	10209.251	17.98
4	14.93	0.307	310.54	12.68	6462.124	11.38
5	17.73	0.378	352.07	14.37	9071.578	15.98
6	23.21	0.604	146.25	5.97	5969.711	10.52
7	26.13	0.571	441.41	18.02	16543.403	29.14
7	29.00	0.361	2449.81	100.00	56771.497	100.00

Report date: 3/16/2017 6:14:01 PM
Printed by: wet


Ident: STD7
Analysis from: 3/16/2017 5:06:04 PM
File: _2017-03-16_

Last save: 3/16/2017 6:12:32 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 13965

Last save: 3/16/2017 1:17:1

SAMPLE:
: HM/AP
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.20	0.248	287.77	9.07	5499.278	7.36
2	8.96	0.191	430.14	13.55	5536.771	7.41
3	11.53	0.241	809.41	25.50	13053.916	17.47
4	14.93	0.308	393.83	12.41	8235.290	11.02
5	17.71	0.383	443.51	13.97	11612.347	15.54
6	23.20	0.586	208.93	6.58	8172.162	10.94
7	26.13	0.577	600.30	18.91	22605.193	30.26
7	28.50	0.362	3173.89	100.00	74714.956	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/16/2017 6:14:07 PM
Printed by: wet


Ident: STD8
Analysis from: 3/16/2017 5:37:32 PM
File: _2017-03-16_

Last save: 3/16/2017 6:12:36 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 13966

Last save: 3/16/2017 1:17:1

SAMPLE:
: HM/AP
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.20	0.236	632.30	9.43	11565.063	7.23
2	8.94	0.188	962.36	14.36	12194.823	7.63
3	11.56	0.259	1645.45	24.55	28100.452	17.58
4	14.85	0.306	867.60	12.94	18038.043	11.28
5	17.57	0.397	928.32	13.85	25352.392	15.86
6	23.06	0.594	454.93	6.79	17803.422	11.14
7	26.02	0.594	1212.19	18.08	46805.269	29.28
7	28.50	0.368	6703.16	100.00	159859.464	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/17/2017 9:51:19 AM
Printed by: wet


Ident: ICV
Analysis from: 3/16/2017 7:13:47 PM
File: _2017-03-16_

Last save: 3/16/2017 7:41:48 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 13969

Last save: 3/16/2017 6:39:2

SAMPLE:
: AK/AP
Vial number: 12
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.18	0.225	309.10	9.69	5482.274	7.40
2	8.94	0.191	432.78	13.57	5557.879	7.50
3	11.51	0.244	819.87	25.70	13366.261	18.03
4	14.90	0.309	396.64	12.43	8289.698	11.18
5	17.69	0.384	446.66	14.00	11693.863	15.78
6	23.16	0.582	220.22	6.90	8511.850	11.48
7	26.10	0.578	564.79	17.70	21224.268	28.63
7	28.00	0.359	3190.07	100.00	74126.094	100.00

Report date: 3/17/2017 9:51:28 AM
Printed by: wet


Ident: ICB
Analysis from: 3/16/2017 7:44:47 PM
File: _2017-03-16_

Last save: 3/16/2017 8:12:48 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 13970

Last save: 3/16/2017 6:39:2

SAMPLE:
: AK/AP
Vial number: 13
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 3/30/2017 2:26:20 PM
Printed by: wet


Ident: CCV
Analysis from: 3/29/2017 9:18:49 AM
File: _2017-03-29_

Last save: 3/29/2017 9:46:20 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14201

Last save: 3/28/2017 12:18:

SAMPLE: *AV*
: HM/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.14	0.267	273.12	8.38	5633.401	7.48
2	8.83	0.194	419.96	12.88	5625.264	7.47
3	11.34	0.247	817.36	25.07	13468.337	17.89
4	14.66	0.309	397.48	12.19	8336.870	11.07
5	17.37	0.379	453.25	13.90	11716.774	15.56
6	23.23	0.569	231.09	7.09	8938.574	11.87
7	25.24	0.495	667.51	20.48	21567.859	28.65
7	27.50	0.351	3259.76	100.00	75287.078	100.00

Report date: 3/30/2017 2:26:29 PM
Printed by: wet


Ident: CCB
Analysis from: 3/29/2017 9:49:33 AM
File: _2017-03-29_

Last save: 3/29/2017 10:17:03 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14202

Last save: 3/28/2017 12:18:

SAMPLE:
: AK/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 3/30/2017 2:26:34 PM
Printed by: wet


Ident: LB86592BLW
Analysis from: 3/29/2017 10:19:59 AM
File: _2017-03-29_

Last save: 3/29/2017 10:47:29 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14203

Last save: 3/28/2017 12:18:

SAMPLE:
: AK/AP
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	0.00	0.000	0.00	0.00	0.000	0.00
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	0.00	0.000	0.00	0.00	0.000	0.00
7	27.50	0.000	0.00	0.00	0.000	0.00

Report date: 3/30/2017 2:26:41 PM
Printed by: wet


Ident: LB86592BSW
Analysis from: 3/29/2017 10:50:25 AM
File: _2017-03-29_

Last save: 3/29/2017 11:17:55 AM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14204

Last save: 3/28/2017 12:18:

SAMPLE:
: AK/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.14	0.266	275.99	8.40	5638.673	7.42
2	8.84	0.195	423.78	12.91	5669.996	7.46
3	11.35	0.247	823.86	25.09	13555.326	17.84
4	14.67	0.309	400.41	12.19	8391.961	11.04
5	17.38	0.380	457.12	13.92	11805.701	15.53
6	23.24	0.572	232.92	7.09	9037.431	11.89
7	25.27	0.502	669.60	20.39	21904.466	28.82
7	27.50	0.353	3283.69	100.00	76003.555	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/30/2017 2:26:52 PM
Printed by: wet


Ident: I2437-02
Analysis from: 3/29/2017 11:21:54 AM
File: _2017-03-29_

Last save: 3/29/2017 12:09:29 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14205

Last save: 3/28/2017 12:18:

SAMPLE:
: AK/AP
Vial number: 14
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.11	0.367	10.23	0.26	277.060	0.13
2	8.78	0.742	1932.58	49.17	93319.521	44.21
3	0.00	0.000	0.00	0.00	0.000	0.00
4	14.90	0.317	10.04	0.26	205.460	0.10
5	17.66	0.369	48.23	1.23	1159.088	0.55
6	0.00	0.000	0.00	0.00	0.000	0.00
7	24.74	0.947	1932.22	49.16	116120.742	55.01
7	27.50	0.392	3933.30	100.07	211081.870	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/30/2017 2:27:04 PM
Printed by: wet


Ident: I2437-01
Analysis from: 3/29/2017 11:52:22 AM
File: _2017-03-29_

Last save: 3/29/2017 1:04:20 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14206

Last save: 3/28/2017 12:18:

SAMPLE:
: AK/AP
Vial number: 12
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	4.98	0.569	12.58	0.32	453.071	0.09
2	8.66	2.329	1933.13	49.58	274154.273	54.16
3	0.00	0.000	0.00	0.00	0.000	0.00
4	14.96	0.317	55.60	1.43	1183.365	0.23
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.82	1.984	1904.99	48.86	230437.267	45.52
7	27.50	0.743	3906.30	100.18	506227.976	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/30/2017 2:27:10 PM
Printed by: wet


Ident: I2437-03
Analysis from: 3/29/2017 12:22:50 PM
File: _2017-03-29_

Last save: 3/29/2017 1:04:44 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14207

Last save: 3/28/2017 12:18:

SAMPLE:
: AK/AP
Vial number: 17
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.60	2.291	1935.90	44.51	268534.000	47.57
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.55	2.266	1892.56	43.51	260271.837	46.10
7	27.50	0.651	3828.46	88.02	528805.837	93.67

This report has been created by IC Net
METROHM LTD

Report date: 3/30/2017 2:27:19 PM
Printed by: wet


Ident: I2437-02DUP
Analysis from: 3/29/2017 12:53:17 PM
File: _2017-03-29_

Last save: 3/29/2017 1:20:47 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14208

Last save: 3/29/2017 12:09:

SAMPLE:
: AK/AP
Vial number: 15
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.12	0.373	11.60	0.29	339.917	0.16
2	8.95	0.739	1939.43	49.16	94730.900	44.38
3	0.00	0.000	0.00	0.00	0.000	0.00
4	14.93	0.316	9.71	0.25	196.397	0.09
5	17.71	0.375	47.53	1.20	1153.300	0.54
6	0.00	0.000	0.00	0.00	0.000	0.00
7	24.77	0.951	1937.26	49.10	117038.526	54.83
7	27.50	0.394	3945.53	100.01	213459.041	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/30/2017 2:27:30 PM
Printed by: wet


Ident: I2437-02MS
Analysis from: 3/29/2017 1:23:49 PM
File: _2017-03-29_

Last save: 3/29/2017 3:06:32 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14209

Last save: 3/29/2017 12:09:

SAMPLE:
: AK/AP
Vial number: 16
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.17	0.340	252.10	4.27	6445.373	2.84
2	8.85	0.568	1929.41	32.64	74326.364	32.80
3	11.51	0.236	910.92	15.41	14127.349	6.23
4	14.80	0.311	430.75	7.29	8976.006	3.96
5	17.52	0.386	514.59	8.71	13306.234	5.87
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.41	0.919	1875.86	31.74	109403.301	48.28
7	27.50	0.394	5913.62	100.05	226584.627	100.00

Report date: 3/30/2017 2:27:38 PM
Printed by: wet


Ident: I2437-01DLX50
Analysis from: 3/29/2017 1:54:16 PM
File: _2017-03-29_

Last save: 3/29/2017 2:21:46 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14210

Last save: 3/29/2017 12:09:

SAMPLE:
: AK/AP
Vial number: 19
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.87	0.193	1304.48	78.57	17323.172	59.56
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.34	0.510	355.72	21.43	11762.094	40.44
7	27.50	0.100	1660.20	100.00	29085.266	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/30/2017 2:28:10 PM
Printed by: wet


Ident: I2437-03DLX50
Analysis from: 3/29/2017 2:24:42 PM
File: _2017-03-29_

Last save: 3/30/2017 2:28:09 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14211

Last save: 3/29/2017 12:09:

SAMPLE:
: AK/AP
Vial number: 20
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.87	0.195	1022.44	67.46	13748.560	45.80
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.32	0.509	493.25	32.54	16268.993	54.20
7	27.50	0.101	1515.70	100.00	30017.552	100.00

Report date: 3/30/2017 2:28:53 PM
Printed by: wet


Ident: I2437-02DLX10
Analysis from: 3/29/2017 2:55:08 PM
File: _2017-03-29_

Last save: 3/30/2017 2:28:53 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14212

Last save: 3/29/2017 12:09:

SAMPLE:
: AK/AP
Vial number: 21
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.88	0.194	1524.17	78.24	20302.901	59.22
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.34	0.509	423.91	21.76	13978.245	40.78
7	27.50	0.100	1948.08	100.00	34281.146	100.00

This report has been created by IC Net
METROHM LTD

Report date: 3/30/2017 2:29:01 PM
Printed by: wet


Ident: CCV
Analysis from: 3/29/2017 3:25:41 PM
File: _2017-03-29_

Last save: 3/29/2017 5:31:49 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14213

Last save: 3/29/2017 12:09:

SAMPLE:
: AK/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.15	0.267	277.31	8.51	5677.777	7.45
2	8.85	0.200	419.93	12.88	5709.155	7.49
3	11.37	0.251	818.47	25.10	13570.779	17.81
4	14.71	0.317	396.43	12.16	8401.762	11.03
5	17.44	0.390	451.94	13.86	11828.265	15.52
6	23.25	0.589	232.27	7.12	9143.159	12.00
7	25.26	0.509	664.15	20.37	21873.878	28.70
7	27.50	0.361	3260.49	100.00	76204.775	100.00

Report date: 3/30/2017 2:29:54 PM
Printed by: wet


Ident: CCB
Analysis from: 3/29/2017 4:02:57 PM
File: _2017-03-29_

Last save: 3/29/2017 4:30:28 PM

Method: AnionIC2-031617.mtw
Run operator: wet
Analysis number: 14214

Last save: 3/29/2017 12:09:

SAMPLE:
: AK/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD